

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112520\
 Data File : BF122484.D
 Acq On : 25 Nov 2020 14:51
 Operator : JU/CG
 Sample : L4882-04
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PAV-B48-112320-0-10

Quant Time: Nov 25 16:06:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 24 14:38:56 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	190833	20.00	ng	0.00
21) Naphthalene-d8	8.139	136	739938	20.00	ng	0.00
39) Acenaphthene-d10	9.898	164	392383	20.00	ng	0.00
64) Phenanthrene-d10	11.380	188	681222	20.00	ng	0.00
76) Chrysene-d12	14.021	240	465647	20.00	ng	-0.01
86) Perylene-d12	15.492	264	493020	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	1339999	107.82	ng	0.01
7) Phenol-d6	6.492	99	1565276	96.27	ng	0.00
23) Nitrobenzene-d5	7.422	82	1160030	95.98	ng	0.00
42) 2,4,6-Tribromophenol	10.686	330	553841	131.51	ng	0.00
45) 2-Fluorobiphenyl	9.222	172	2065076	82.23	ng	0.00
79) Terphenyl-d14	12.974	244	2013913	91.62	ng	0.00
Target Compounds						
10) Phenol	6.504	94	55471	3.46	ng	81
50) Dimethylphthalate	9.610	163	112933	4.05	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112520\
Data File : BF122484.D
Acq On : 25 Nov 2020 14:51
Operator : JU/CG
Sample : L4882-04
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PAV-B48-112320-0-10

Quant Time: Nov 25 16:06:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 24 14:38:56 2020
Response via : Initial Calibration

