


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF112615\
 Data File : BF083322.D
 Acq On : 26 Nov 2015 16:28
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\Methods\8270-BF112115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sun Nov 22 09:22:46 2015



AutoFind: Scans 821, 822, 823; Background Corrected with Scan 818

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.0	27265	PASS
68	69	0.00	2	1.2	309	PASS
69	198	0.00	100	44.9	25468	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.2	29080	PASS
197	198	0.00	2	0.2	108	PASS
198	198	100	100	100.0	56753	PASS
199	198	5	9	6.4	3654	PASS
275	198	10	60	22.9	13012	PASS
365	198	1	100	2.5	1402	PASS
441	442	0.01	24	16.4	6672	PASS
442	198	50	100	71.8	40764	PASS
443	442	15	24	19.1	7780	PASS