

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112624\
 Data File : BF140629.D
 Acq On : 26 Nov 2024 09:35
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 10:51:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.493	0.492	0.2	109	0.00
3	Pyridine	1.084	1.189	-9.7	108	-0.01
4	n-Nitrosodimethylamine	0.637	0.584	8.3	97	-0.01
5 S	2-Fluorophenol	1.172	1.147	2.1	106	0.00
6	Aniline	1.091	1.316	-20.6	114	0.00
7 S	Phenol-d6	1.550	1.512	2.5	102	0.00
8	2-Chlorophenol	1.261	1.226	2.8	103	0.00
9	Benzaldehyde	0.820	0.758	7.6	111	0.00
10 C	Phenol	1.583	1.584	-0.1	103	0.00
11	bis(2-Chloroethyl)ether	1.211	1.200	0.9	104	0.00
12	1,3-Dichlorobenzene	1.417	1.400	1.2	108	0.00
13 C	1,4-Dichlorobenzene	1.435	1.397	2.6	106	0.00
14	1,2-Dichlorobenzene	1.344	1.311	2.5	105	0.00
15	Benzyl Alcohol	1.149	1.117	2.8	99	0.00
16	2,2'-oxybis(1-Chloropropane	1.431	1.458	-1.9	108	0.00
17	2-Methylphenol	1.009	1.003	0.6	104	0.00
18	Hexachloroethane	0.536	0.516	3.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	0.916	0.884	3.5	101	0.00
20	3+4-Methylphenols	1.298	1.257	3.2	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
22	Acetophenone	0.488	0.460	5.7	100	0.00
23 S	Nitrobenzene-d5	0.391	0.378	3.3	101	0.00
24	Nitrobenzene	0.404	0.393	2.7	102	0.00
25	Isophorone	0.652	0.635	2.6	100	0.00
26 C	2-Nitrophenol	0.179	0.183	-2.2	106	0.00
27	2,4-Dimethylphenol	0.214	0.228	-6.5	106	0.00
28	bis(2-Chloroethoxy)methane	0.397	0.396	0.3	104	0.00
29 C	2,4-Dichlorophenol	0.284	0.283	0.4	105	0.00
30	1,2,4-Trichlorobenzene	0.324	0.313	3.4	103	0.00
31	Naphthalene	1.030	1.008	2.1	104	0.00
32	Benzoic acid	0.164	0.173	-5.5	102	0.00
33	4-Chloroaniline	0.308	0.341	-10.7	110	0.00
34 C	Hexachlorobutadiene	0.215	0.206	4.2	102	0.00
35	Caprolactam	0.088	0.090	-2.3	104	0.00
36 C	4-Chloro-3-methylphenol	0.318	0.314	1.3	100	0.00
37	2-Methylnaphthalene	0.654	0.644	1.5	103	0.00
38	1-Methylnaphthalene	0.641	0.628	2.0	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.586	0.557	4.9	103	0.00
41 P	Hexachlorocyclopentadiene	0.107	0.095	11.2	86	0.00
42 S	2,4,6-Tribromophenol	0.214	0.207	3.3	100	0.00
43 C	2,4,6-Trichlorophenol	0.367	0.356	3.0	101	0.00
44	2,4,5-Trichlorophenol	0.399	0.404	-1.3	104	0.00
45 S	2-Fluorobiphenyl	1.342	1.257	6.3	101	0.00
46	1,1'-Biphenyl	1.493	1.436	3.8	103	0.00
47	2-Chloronaphthalene	1.131	1.097	3.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112624\
 Data File : BF140629.D
 Acq On : 26 Nov 2024 09:35
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 10:51:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.363	0.346	4.7	98	0.00
49	Acenaphthylene	1.710	1.659	3.0	104	0.00
50	Dimethylphthalate	1.317	1.270	3.6	101	0.00
51	2,6-Dinitrotoluene	0.299	0.292	2.3	101	0.00
52 C	Acenaphthene	1.086	1.044	3.9	102	0.00
53	3-Nitroaniline	0.292	0.303	-3.8	105	0.00
54 P	2,4-Dinitrophenol	0.122	0.136	-11.5	101	0.00
55	Dibenzofuran	1.657	1.584	4.4	101	0.00
56 P	4-Nitrophenol	0.199	0.189	5.0	95	0.00
57	2,4-Dinitrotoluene	0.397	0.395	0.5	102	0.00
58	Fluorene	1.330	1.280	3.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.306	0.310	-1.3	103	0.00
60	Diethylphthalate	1.338	1.307	2.3	104	0.00
61	4-Chlorophenyl-phenylether	0.653	0.621	4.9	101	0.00
62	4-Nitroaniline	0.306	0.306	0.0	102	0.00
63	Azobenzene	1.267	1.239	2.2	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.102	0.112	-9.8	105	0.00
66 c	n-Nitrosodiphenylamine	0.591	0.583	1.4	104	0.00
67	4-Bromophenyl-phenylether	0.206	0.205	0.5	105	0.00
68	Hexachlorobenzene	0.238	0.232	2.5	102	0.00
69	Atrazine	0.166	0.163	1.8	119	0.00
70 C	Pentachlorophenol	0.105	0.112	-6.7	99	0.00
71	Phenanthrene	0.961	0.928	3.4	101	0.00
72	Anthracene	0.940	0.926	1.5	103	0.00
73	Carbazole	0.905	0.893	1.3	103	0.00
74	Di-n-butylphthalate	1.044	1.063	-1.8	107	0.00
75 C	Fluoranthene	1.044	1.033	1.1	105	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
77	Benزيدine	0.581	0.610	-5.0	111	0.00
78	Pyrene	1.849	1.845	0.2	104	0.00
79 S	Terphenyl-d14	1.284	1.263	1.6	103	0.00
80	Butylbenzylphthalate	0.666	0.715	-7.4	110	0.00
81	Benzo(a)anthracene	1.324	1.340	-1.2	109	0.00
82	3,3'-Dichlorobenzidine	0.397	0.400	-0.8	101	0.00
83	Chrysene	1.211	1.144	5.5	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.841	0.933	-10.9	117	0.00
85 c	Di-n-octyl phthalate	1.150	1.287	-11.9	119	0.01
86 I	Perylene-d12	1.000	1.000	0.0	102	0.01
87	Indeno(1,2,3-cd)pyrene	1.303	1.258	3.5	98	0.02
88	Benzo(b)fluoranthene	1.256	1.211	3.6	104	0.01
89	Benzo(k)fluoranthene	1.099	1.084	1.4	100	0.01
90 C	Benzo(a)pyrene	1.021	1.002	1.9	101	0.01
91	Dibenzo(a,h)anthracene	1.067	1.035	3.0	98	0.02
92	Benzo(g,h,i)perylene	1.089	1.065	2.2	100	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112624\
Data File : BF140629.D
Acq On : 26 Nov 2024 09:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 10:51:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0