

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112624\
 Data File : BF140629.D
 Acq On : 26 Nov 2024 09:35
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 10:51:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	39.895	0.3	109	0.00
3	Pyridine	40.000	43.855	-9.6	108	-0.01
4	n-Nitrosodimethylamine	40.000	36.633	8.4	97	-0.01
5 S	2-Fluorophenol	80.000	78.312	2.1	106	0.00
6	Aniline	40.000	48.243	-20.6	114	0.00
7 S	Phenol-d6	80.000	78.046	2.4	102	0.00
8	2-Chlorophenol	40.000	38.892	2.8	103	0.00
9	Benzaldehyde	40.000	36.974	7.6	111	0.00
10 C	Phenol	40.000	40.036	-0.1	103	0.00
11	bis(2-Chloroethyl)ether	40.000	39.643	0.9	104	0.00
12	1,3-Dichlorobenzene	40.000	39.507	1.2	108	0.00
13 C	1,4-Dichlorobenzene	40.000	38.941	2.6	106	0.00
14	1,2-Dichlorobenzene	40.000	39.016	2.5	105	0.00
15	Benzyl Alcohol	40.000	38.862	2.8	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.754	-1.9	108	0.00
17	2-Methylphenol	40.000	39.732	0.7	104	0.00
18	Hexachloroethane	40.000	38.534	3.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.629	3.4	101	0.00
20	3+4-Methylphenols	40.000	38.763	3.1	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	37.744	5.6	100	0.00
23 S	Nitrobenzene-d5	80.000	77.298	3.4	101	0.00
24	Nitrobenzene	40.000	38.893	2.8	102	0.00
25	Isophorone	40.000	38.957	2.6	100	0.00
26 C	2-Nitrophenol	40.000	40.812	-2.0	106	0.00
27	2,4-Dimethylphenol	40.000	42.482	-6.2	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.830	0.4	104	0.00
29 C	2,4-Dichlorophenol	40.000	39.861	0.3	105	0.00
30	1,2,4-Trichlorobenzene	40.000	38.613	3.5	103	0.00
31	Naphthalene	40.000	39.153	2.1	104	0.00
32	Benzoic acid	40.000	38.676	3.3	102	0.00
33	4-Chloroaniline	40.000	44.206	-10.5	110	0.00
34 C	Hexachlorobutadiene	40.000	38.417	4.0	102	0.00
35	Caprolactam	40.000	41.079	-2.7	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.508	1.2	100	0.00
37	2-Methylnaphthalene	40.000	39.360	1.6	103	0.00
38	1-Methylnaphthalene	40.000	39.149	2.1	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.051	4.9	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.464	16.3	86	0.00
42 S	2,4,6-Tribromophenol	80.000	77.498	3.1	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.758	3.1	101	0.00
44	2,4,5-Trichlorophenol	40.000	40.575	-1.4	104	0.00
45 S	2-Fluorobiphenyl	80.000	74.932	6.3	101	0.00
46	1,1'-Biphenyl	40.000	38.477	3.8	103	0.00
47	2-Chloronaphthalene	40.000	38.765	3.1	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112624\
 Data File : BF140629.D
 Acq On : 26 Nov 2024 09:35
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 10:51:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.058	4.9	98	0.00
49	Acenaphthylene	40.000	38.811	3.0	104	0.00
50	Dimethylphthalate	40.000	38.556	3.6	101	0.00
51	2,6-Dinitrotoluene	40.000	39.077	2.3	101	0.00
52 C	Acenaphthene	40.000	38.434	3.9	102	0.00
53	3-Nitroaniline	40.000	41.489	-3.7	105	0.00
54 P	2,4-Dinitrophenol	40.000	39.575	1.1	101	0.00
55	Dibenzofuran	40.000	38.234	4.4	101	0.00
56 P	4-Nitrophenol	40.000	38.011	5.0	95	0.00
57	2,4-Dinitrotoluene	40.000	39.809	0.5	102	0.00
58	Fluorene	40.000	38.490	3.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.445	-1.1	103	0.00
60	Diethylphthalate	40.000	39.079	2.3	104	0.00
61	4-Chlorophenyl-phenylether	40.000	38.055	4.9	101	0.00
62	4-Nitroaniline	40.000	40.002	-0.0	102	0.00
63	Azobenzene	40.000	39.108	2.2	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.867	-9.7	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.427	1.4	104	0.00
67	4-Bromophenyl-phenylether	40.000	39.770	0.6	105	0.00
68	Hexachlorobenzene	40.000	38.990	2.5	102	0.00
69	Atrazine	40.000	39.093	2.3	119	0.00
70 C	Pentachlorophenol	40.000	42.588	-6.5	99	0.00
71	Phenanthrene	40.000	38.625	3.4	101	0.00
72	Anthracene	40.000	39.395	1.5	103	0.00
73	Carbazole	40.000	39.485	1.3	103	0.00
74	Di-n-butylphthalate	40.000	40.700	-1.8	107	0.00
75 C	Fluoranthene	40.000	39.604	1.0	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	41.984	-5.0	111	0.00
78	Pyrene	40.000	39.898	0.3	104	0.00
79 S	Terphenyl-d14	80.000	78.692	1.6	103	0.00
80	Butylbenzylphthalate	40.000	42.923	-7.3	110	0.00
81	Benzo(a)anthracene	40.000	40.491	-1.2	109	0.00
82	3,3'-Dichlorobenzidine	40.000	40.205	-0.5	101	0.00
83	Chrysene	40.000	37.796	5.5	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.382	-11.0	117	0.00
85 c	Di-n-octyl phthalate	40.000	44.768	-11.9	119	0.01
86 I	Perylene-d12	20.000	20.000	0.0	102	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	38.598	3.5	98	0.02
88	Benzo(b)fluoranthene	40.000	38.547	3.6	104	0.01
89	Benzo(k)fluoranthene	40.000	39.437	1.4	100	0.01
90 C	Benzo(a)pyrene	40.000	39.225	1.9	101	0.01
91	Dibenzo(a,h)anthracene	40.000	38.798	3.0	98	0.02
92	Benzo(g,h,i)perylene	40.000	39.105	2.2	100	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112624\
Data File : BF140629.D
Acq On : 26 Nov 2024 09:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 10:51:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0