

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF112724\  
 Data File : BF140674.D  
 Acq On : 27 Nov 2024 17:06  
 Operator : RC/JU  
 Sample : P5000-08  
 Misc :  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 MH-733

Quant Time: Nov 28 00:11:28 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF112124.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 21 15:23:48 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.869  | 152  | 65314    | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.151  | 136  | 246753   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.904  | 164  | 139477   | 20.000  | ng    | -0.01    |
| 64) Phenanthrene-d10        | 11.392 | 188  | 275850   | 20.000  | ng    | -0.01    |
| 76) Chrysene-d12            | 14.045 | 240  | 156653   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 15.539 | 264  | 98285    | 20.000  | ng    | 0.00     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.510  | 112  | 561668   | 146.726 | ng    | 0.01     |
| 7) Phenol-d6                | 6.510  | 99   | 699729   | 138.267 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.434  | 82   | 500488   | 103.748 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.698 | 330  | 256848   | 172.183 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.222  | 172  | 998798   | 106.695 | ng    | -0.01    |
| 79) Terphenyl-d14           | 12.986 | 244  | 1127838  | 112.107 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF112724\  
Data File : BF140674.D  
Acq On : 27 Nov 2024 17:06  
Operator : RC/JU  
Sample : P5000-08  
Misc :  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
MH-733

Quant Time: Nov 28 00:11:28 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF112124.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Nov 21 15:23:48 2024  
Response via : Initial Calibration

