


Data Path : Z:\HPCHEM1\BNA F\Data\BF112817\
 Data File : BF100948.D
 Acq On : 28 Nov 2017 22:34
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110817.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 28 15:21:02 2017



AutoFind: Scans 1621, 1622, 1623; Background Corrected with Scan 1617

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.7	8093	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	30.5	6729	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.7	8970	PASS
197	198	0.00	2	0.6	131	PASS
198	198	100	100	100.0	22035	PASS
199	198	5	9	7.2	1578	PASS
275	198	10	60	25.1	5533	PASS
365	198	1	100	3.7	816	PASS
441	443	0.01	100	85.6	3456	PASS
442	198	50	100	95.8	21107	PASS
443	442	15	24	19.1	4037	PASS