


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF112818\
 Data File : BF111040.D
 Acq On : 28 Nov 2018 12:44
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF112618.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 27 15:32:10 2018



AutoFind: Scans 1633, 1634, 1635; Background Corrected with Scan 1628

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.3	24735	PASS
68	69	0.00	2	1.5	364	PASS
69	198	0.00	100	36.7	23683	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.3	31832	PASS
197	198	0.00	2	0.3	170	PASS
198	198	100	100	100.0	64586	PASS
199	198	5	9	6.7	4359	PASS
275	198	10	60	26.6	17153	PASS
365	198	1	100	3.2	2036	PASS
441	198	0.01	100	15.6	10067	PASS
442	198	50	100	96.8	62514	PASS
443	442	15	24	19.6	12248	PASS