

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112916\
 Data File : BF091324.D
 Acq On : 29 Nov 2016 13:51
 Operator : UM/SJ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Nov 29 14:41:56 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF112916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 13:33:46 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	160013	20.00	ng	0.00
21) Naphthalene-d8	8.15	136	597126	20.00	ng	-0.02
38) Acenaphthene-d10	9.91	164	343077	20.00	ng	0.00
63) Phenanthrene-d10	11.38	188	541091	20.00	ng	-0.02
75) Chrysene-d12	14.01	240	339528	20.00	ng	-0.03
86) Perylene-d12	15.44	264	319503	20.00	ng	-0.03

System Monitoring Compounds

5) 2-Fluorophenol	5.45	112	713162	72.81	ng	-0.02
7) Phenol-d6	6.49	99	894122	74.16	ng	0.00
23) Nitrobenzene-d5	7.43	82	850016	79.77	ng	-0.02
41) 2,4,6-Tribromophenol	10.69	330	233869	72.01	ng	-0.02
44) 2-Fluorobiphenyl	9.24	172	1507541	79.56	ng	0.00
78) Terphenyl-d14	12.97	244	1378368	88.24	ng	-0.02

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.46	88	167652	37.84	ng	# 86
3) Pyridine	3.18	79	474480	37.84	ng	87
4) n-Nitrosodimethylamine	3.12	42	254419	38.67	ng	96
6) Aniline	6.53	93	618920	38.64	ng	# 70
8) 2-Chlorophenol	6.65	128	412317	38.39	ng	98
9) Benzaldehyde	6.41	77	301980	38.95	ng	79
10) Phenol	6.50	94	548126	38.63	ng	93
11) bis(2-Chloroethyl)ether	6.60	93	382314	37.71	ng	91
12) 1,3-Dichlorobenzene	7.04	146	430100	36.00	ng	97
13) 1,4-Dichlorobenzene	6.88	146	440824	37.10	ng	97
14) 1,2-Dichlorobenzene	7.04	146	430025	38.85	ng	94
15) Benzyl Alcohol	7.01	79	371422	38.49	ng	97
16) 2,2'-oxybis(1-Chloropropan	7.14	45	838264	38.46	ng	71
17) 2-Methylphenol	7.12	107	337712	39.78	ng	# 75
18) Hexachloroethane	7.38	117	169202	40.10	ng	95
19) n-Nitroso-di-n-propylamine	7.28	70	271302	35.71	ng	# 96
20) 3+4-Methylphenols	7.27	107	370030	35.71	ng	# 75
22) Acetophenone	7.28	105	594677	42.02	ng	97
24) Nitrobenzene	7.45	77	466811	42.79	ng	# 77
25) Isophorone	7.69	82	755792	40.04	ng	99
26) 2-Nitrophenol	7.76	139	196000	39.42	ng	98
27) 2,4-Dimethylphenol	7.81	122	388855	41.81	ng	93
28) bis(2-Chloroethoxy)methane	7.90	93	408749	35.97	ng	99
29) 2,4-Dichlorophenol	8.00	162	307021	37.53	ng	97
30) 1,2,4-Trichlorobenzene	8.09	180	365648	39.87	ng	99
31) Naphthalene	8.17	128	1120538	39.48	ng	99
32) Benzoic acid	7.92	122	303794	44.31	ng	89
33) 4-Chloroaniline	8.22	127	467161	38.51	ng	94
34) Hexachlorobutadiene	8.30	225	231516	41.05	ng	98
35) Caprolactam	8.58	113	88559	37.66	ng	99
36) 4-Chloro-3-methylphenol	8.70	107	346213	39.20	ng	96
37) 2-Methylnaphthalene	8.87	142	750177	38.91	ng	95
39) 1,2,4,5-Tetrachlorobenzene	9.03	216	347259	35.90	ng	97
40) Hexachlorocyclopentadiene	9.02	237	172410	38.46	ng	99

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112916\
 Data File : BF091324.D
 Acq On : 29 Nov 2016 13:51
 Operator : UM/SJ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Nov 29 14:41:56 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF112916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 13:33:46 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.14	196	244985	38.97	nq	96
43) 2,4,5-Trichlorophenol	9.18	196	253416	40.23	nq	95
45) 1,1'-Biphenyl	9.33	154	880912	35.71	nq	97
46) 2-Chloronaphthalene	9.35	162	655102	35.66	nq	99
47) 2-Nitroaniline	9.44	65	232636	35.26	nq	# 76
48) Acenaphthylene	9.77	152	1080730	37.37	nq	99
49) Dimethylphthalate	9.64	163	797376	38.38	nq	99
50) 2,6-Dinitrotoluene	9.69	165	181717	39.14	nq	# 58
51) Acenaphthene	9.94	154	769299	39.44	nq	97
52) 3-Nitroaniline	9.85	138	188065	34.99	nq	92
53) 2,4-Dinitrophenol	9.96	184	87429	43.06	nq	92
54) Dibenzofuran	10.12	168	992574	38.00	nq	# 89
55) 4-Nitrophenol	10.00	139	141239	36.39	nq	94
56) 2,4-Dinitrotoluene	10.09	165	237175	39.38	nq	# 85
57) Fluorene	10.46	166	733693	36.59	nq	99
58) 2,3,4,6-Tetrachlorophenol	10.23	232	209667	38.98	nq	94
59) Diethylphthalate	10.33	149	778975	37.99	nq	95
60) 4-Chlorophenyl-phenylether	10.45	204	358636	35.66	nq	# 90
61) 4-Nitroaniline	10.46	138	188923	37.44	nq	90
62) Azobenzene	10.61	77	848056	40.53	nq	93
64) 4,6-Dinitro-2-methylphenol	10.49	198	124484	41.76	nq	# 68
65) n-Nitrosodiphenylamine	10.56	169	614495	37.46	nq	99
66) 4-Bromophenyl-phenylether	10.94	248	244851	42.10	nq	# 79
67) Hexachlorobenzene	11.00	284	229040	36.45	nq	# 76
68) Atrazine	11.09	200	178339	33.57	nq	97
69) Pentachlorophenol	11.19	266	158619	42.87	nq	97
70) Phenanthrene	11.41	178	1026828	36.52	nq	99
71) Anthracene	11.46	178	1179937	42.29	nq	99
72) Carbazole	11.61	167	954410	37.69	nq	98
73) Di-n-butylphthalate	11.96	149	1129408	39.36	nq	99
74) Fluoranthene	12.60	202	1106646	41.55	nq	99
76) Benzidine	12.71	184	414279	41.56	nq	97
77) Pyrene	12.82	202	1131650	43.92	nq	99
79) Butylbenzylphthalate	13.44	149	369564	38.30	nq	# 84
80) Benzo(a)anthracene	14.00	228	773615	38.96	nq	98
81) 3,3'-Dichlorobenzidine	13.97	252	266871	38.15	nq	# 97
82) Chrysene	14.05	228	812427	41.35	nq	98
83) Bis(2-ethylhexyl)phthalate	14.01	149	475477	38.88	nq	# 92
84) Di-n-octyl phthalate	14.62	149	731260	39.51	nq	97
85) Indeno(1,2,3-cd)pyrene	16.85	276	866099	48.14	nq	95
87) Benzo(b)fluoranthene	15.07	252	1429133	71.96	nq	# 97
88) Benzo(k)fluoranthene	15.07	252	1429133	85.58	nq	# 96
89) Benzo(a)pyrene	15.39	252	690901	38.74	nq	97
90) Dibenzo(a,h)anthracene	16.87	278	731015	44.32	nq	# 95
91) Benzo(g,h,i)perylene	17.27	276	763295	44.88	nq	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112916\
 Data File : BF091324.D
 Acq On : 29 Nov 2016 13:51
 Operator : UM/SJ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampled :

Quant Time: Nov 29 14:41:56 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF112916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 13:33:46 2016
 Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.87 min Scan# 401
Delta R.T. -0.01 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 160013
Ion Ratio Lower Upper
152 100
150 153.2 122.5 183.7
115 53.8 51.8 77.8




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 37.84 ng
RT: 2.46 min Scan# 15
Delta R.T. -0.04 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion: 88 Resp: 167652
Ion Ratio Lower Upper
88 100
58 105.5 70.3 105.5#
43 37.9 31.3 46.9



Abundance Ion 88.00 (87.70 to 88.70): BF0
Ion 58.00 (57.70 to 58.70): BF0
Ion 43.00 (42.70 to 43.70): BF0





#3

Pvridine

Concen: 37.84 ng

RT: 3.18 min Scan# 78

Delta R.T. -0.04 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 79 Resp: 474480

Ion Ratio Lower Upper

79 100

52 106.0 71.0 106.4

51 48.3 37.0 55.4



#4

n-Nitrosodimethylamine

Concen: 38.67 ng

RT: 3.12 min Scan# 73

Delta R.T. -0.05 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51




Tgt Ion: 42 Resp: 254419


Ion Ratio Lower Upper

42 100

74 102.7 78.9 118.3

44 7.5 5.8 8.6





#5

2-Fluorophenol

Concen: 72.81 ng

RT: 5.45 min Scan# 277

Delta R.T. -0.02 min

Lab File: BF091324.D


Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :




Tot Ion:	112	Resp:	713162
Ion	Ratio	Lower	Upper
112	100		
64	78.8	61.5	92.3
63	42.2	33.3	49.9



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 38.64 ng


RT: 6.53 min Scan# 371

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51



Tgt Ion:	93	Resp:	618920
Ion	Ratio	Lower	Upper
93	100		
66	42.4	55.8	83.6#
65	22.5	29.9	44.9#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 74.16 ng

RT: 6.49 min Scan# 368

Delta R.T. -0.01 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 99 Resp: 894122

Ion Ratio Lower Upper




99 100

42 23.6 20.6 31.0

71 35.7 29.9 44.9



Abundance Ion 99.00 (98.70 to 99.70): BF0
Ion 42.00 (41.70 to 42.70): BF0
Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 38.39 ng

RT: 6.65 min Scan# 382

Delta R.T. -0.01 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion: 128 Resp: 412317

Ion Ratio Lower Upper



128 100


130 32.7 13.1 53.1

64 42.1 24.1 64.1



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 38.95 ng

RT: 6.41 min Scan# 361

Delta R.T. -0.01 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 77 Resp: 301980

Ion Ratio Lower Upper

77 100

105 104.6 66.4 106.4




106 100.4 60.9 100.9



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 38.63 ng

RT: 6.50 min Scan# 369

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion: 94 Resp: 548126

Ion Ratio Lower Upper

94 100

65 34.0 16.9 56.9



66 44.5 30.6 70.6




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0





#11

bis(2-Chloroethyl)ether

Concen: 37.71 ng

RT: 6.60 min Scan# 377

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 93 Resp: 382314

Ion Ratio Lower Upper

93 100

63 106.1 74.8 114.8

95 32.5 12.4 52.4



#12

1,3-Dichlorobenzene

Concen: 36.00 ng

RT: 7.04 min Scan# 416

Delta R.T. 0.22 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51




Tgt Ion: 146 Resp: 430100


Ion Ratio Lower Upper

146 100

148 64.3 51.4 77.0

75 24.9 23.6 35.4









#13
 1,4-Dichlorobenzene
 Concen: 37.10 ng
 RT: 6.88 min Scan# 402
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:146 Resp: 440824
 Ion Ratio Lower Upper
 146 100
 148 63.5 51.4 77.0
 111 44.6 32.2 48.2



#14
 1,2-Dichlorobenzene
 Concen: 38.85 ng
 RT: 7.04 min Scan# 416
 Delta R.T. -0.01 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion:146 Resp: 430025
 Ion Ratio Lower Upper
 146 100
 148 64.3 51.0 76.6
 111 39.3 38.9 58.3





#15

Benzyl Alcohol

Concen: 38.49 ng

RT: 7.01 min Scan# 413

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampled :





Tot Ion: 79 Resp: 371422

Ion Ratio Lower Upper

79 100

108 75.0 62.7 94.1

77 62.8 51.7 77.5



#16

2,2'-oxybis(1-chloropropane)

Concen: 38.46 ng

RT: 7.14 min Scan# 425

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51




Tgt Ion: 45 Resp: 838264


Ion Ratio Lower Upper

45 100

77 9.6 3.8 43.8

79 7.2 0.2 40.2





#17

2-Methylphenol

Concen: 39.78 ng

RT: 7.12 min Scan# 423

Delta R.T. -0.01 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampled :

Tot Ion:107 Resp: 337712



Ion Ratio Lower Upper

107 100

108 115.7 67.7 101.5#

77 44.8 40.7 61.1

79 44.4 23.4 35.2#





Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0



#18

Hexachloroethane

Concen: 40.10 ng

RT: 7.38 min Scan# 446

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51



Tgt Ion:117 Resp: 169202

Ion Ratio Lower Upper

117 100

119 97.0 78.6 118.0


201 100.4 86.7 130.1






Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 35.71 ng
RT: 7.28 min Scan# 437
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51


Instrument :
BNA_F
ClientSampleId :


Tot Ion: 70 Resp: 271302
Ion Ratio Lower Upper
70 100
42 77.8 64.8 97.2
101 9.2 6.2 9.4
130 18.3 11.7 17.5#



#20
3+4-Methylphenols
Concen: 35.71 ng
RT: 7.27 min Scan# 436
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion: 107 Resp: 370030
Ion Ratio Lower Upper
107 100
108 88.8 64.6 104.6
77 100.1 30.9 70.9#
79 30.3 9.3 49.3





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.15 min Scan# 513

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampled :

Tot Ion:136 Resp: 597126


Ion Ratio Lower Upper

136 100

137 10.9 9.1 13.7

54 12.3 9.1 13.7

68 5.7 5.9 8.9#




Abundance Ion 136.00 (135.70 to 136.70): E



Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



Time-->



#22

Acetophenone

Concen: 42.02 ng

RT: 7.28 min Scan# 437

Delta R.T. -0.01 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Tgt Ion:105 Resp: 594677


Ion Ratio Lower Upper

105 100

71 8.6 7.9 11.9

51 32.0 24.5 36.7

120 22.4 16.6 24.8




Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0


Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E



Time-->









#23
 Nitrobenzene-d5
 Concen: 79.77 ng
 RT: 7.43 min Scan# 450
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	82	Resp	850016
Ion	Ratio	Lower	Upper
82	100		
128	41.7	31.8	47.6
54	65.9	49.1	73.7



#24
 Nitrobenzene
 Concen: 42.79 ng
 RT: 7.45 min Scan# 452
 Delta R.T. -0.01 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion	77	Resp	466811
Ion	Ratio	Lower	Upper
77	100		
123	54.1	28.5	42.7#
65	14.8	12.5	18.7









#25
Isophorone
Concen: 40.04 ng
RT: 7.69 min Scan# 473
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51




Instrument :
BNA_F
ClientSampleId :


Tot Ion: 82 Resp: 755792
Ion Ratio Lower Upper
82 100
95 7.4 5.7 8.5
138 17.3 13.4 20.0



#26
2-Nitrophenol
Concen: 39.42 ng
RT: 7.76 min Scan# 479
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion: 139 Resp: 196000
Ion Ratio Lower Upper
139 100
109 31.3 25.3 37.9
65 67.5 55.4 83.2









#27
 2,4-Dimethylphenol
 Concen: 41.81 ng
 RT: 7.81 min Scan# 483
 Delta R.T. -0.01 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:122 Resp: 388855
 Ion Ratio Lower Upper
 122 100
 107 110.8 96.8 145.2
 121 57.2 47.7 71.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 35.97 ng
 RT: 7.90 min Scan# 491
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion: 93 Resp: 408749
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.0 39.0
 123 10.9 8.6 12.8









#29
 2,4-Dichlorophenol
 Concen: 37.53 ng
 RT: 8.00 min Scan# 500
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:162 Resp: 307021
 Ion Ratio Lower Upper
 162 100
 164 65.7 45.0 85.0
 98 45.8 21.3 61.3




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 39.87 ng
 RT: 8.09 min Scan# 508
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion:180 Resp: 365648
 Ion Ratio Lower Upper
 180 100
 182 94.8 75.8 113.8
 145 31.3 27.0 40.4



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 39.48 ng

RT: 8.17 min Scan# 515

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion:128 Resp: 1120538

Ion Ratio Lower Upper

128 100

129 11.4 9.1 13.7




127 13.2 11.0 16.4



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 44.31 ng

RT: 7.92 min Scan# 493

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:122 Resp: 303794

Ion Ratio Lower Upper

122 100

105 120.4 113.0 153.0



77 89.7 82.0 122.0




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 38.51 ng
RT: 8.22 min Scan# 519
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51




Instrument :
BNA_F
ClientSampleId :


Tot Ion:127	Resp:	467161
Ion Ratio	Lower	Upper
127	100	
129	31.9	26.2 39.4
65	39.2	26.6 39.8
92	20.8	15.4 23.0



#34
Hexachlorobutadiene
Concen: 41.05 ng
RT: 8.30 min Scan# 526
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion:225	Resp:	231516
Ion Ratio	Lower	Upper
225	100	
223	62.9	49.8 74.6
227	63.7	52.6 79.0





#35

Caprolactam

Concen: 37.66 ng

RT: 8.58 min Scan# 551

Delta R.T. -0.03 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion:113 Resp: 88559

Ion Ratio Lower Upper

113 100

55 260.2 239.8 279.8




56 187.3 165.6 205.6



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 39.20 ng

RT: 8.70 min Scan# 561

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:107 Resp: 346213

Ion Ratio Lower Upper

107 100

144 25.5 18.9 28.3



142 76.7 58.5 87.7




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 38.91 ng
RT: 8.87 min Scan# 576
Delta R.T. -0.01 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
142	100		
141	83.8	71.7	107.5
115	29.5	25.0	37.6




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E






#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.91 min Scan# 667
Delta R.T. -0.01 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion	Ratio	Lower	Upper
164	100		
162	96.7	82.6	124.0
160	42.9	36.7	55.1



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E





#39

1,2,4,5-Tetrachlorobenzene

Concen: 35.90 ng

RT: 9.03 min Scan# 590

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :

Tot Ion:216 Resp: 347259


Ion Ratio Lower Upper

216 100

214 77.9 63.8 95.8

179 21.8 18.2 27.2

108 18.6 17.2 25.8






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 38.46 ng

RT: 9.02 min Scan# 589

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:237 Resp: 172410

Ion Ratio Lower Upper

237 100

235 63.8 43.9 83.9



272 11.7 0.0 33.1




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 72.01 ng
 RT: 10.69 min Scan# 735
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:330 Resp: 233869
 Ion Ratio Lower Upper
 330 100
 332 97.6 76.2 114.4
 141 32.2 33.6 50.4#



#42
 2,4,6-Trichlorophenol
 Concen: 38.97 ng
 RT: 9.14 min Scan# 600
 Delta R.T. -0.01 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion:196 Resp: 244985
 Ion Ratio Lower Upper
 196 100
 198 101.7 77.4 116.0
 200 32.5 25.2 37.8









#43
 2,4,5-Trichlorophenol
 Concen: 40.23 ng
 RT: 9.18 min Scan# 603
 Delta R.T. -0.01 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:196 Resp: 253416
 Ion Ratio Lower Upper
 196 100
 198 101.0 77.0 115.6
 97 46.7 33.5 50.3
 132 27.0 20.5 30.7




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 79.56 ng
 RT: 9.24 min Scan# 608
 Delta R.T. -0.00 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion:172 Resp: 1507541
 Ion Ratio Lower Upper
 172 100
 171 35.2 29.4 44.0
 170 23.6 19.7 29.5



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 35.71 ng

RT: 9.33 min Scan# 616

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampled :


Tot Ion:154 Resp: 880912

Ion Ratio Lower Upper

154 100

153 40.0 20.6 60.6




76 12.9 0.0 36.6



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 35.66 ng

RT: 9.35 min Scan# 618

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:162 Resp: 655102

Ion Ratio Lower Upper

162 100

127 43.2 34.3 51.5



164 34.4 26.6 40.0




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E





#47

2-Nitroaniline

Concen: 35.26 ng

RT: 9.44 min Scan# 626

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 65 Resp: 232636

Ion Ratio Lower Upper




65 100

92 58.4 53.6 80.4

138 78.7 91.0 136.6#



Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#48

Acenaphthylene

Concen: 37.37 ng

RT: 9.77 min Scan# 655

Delta R.T. -0.01 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:152 Resp: 1080730

Ion Ratio Lower Upper



152 100


151 19.9 16.6 24.8

153 13.2 10.7 16.1



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 38.38 ng
RT: 9.64 min Scan# 643
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Instrument :
BNA_F
ClientSampled :

Tot Ion:163 Resp: 797376
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.4
164 10.1 7.8 11.8




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 39.14 ng
RT: 9.69 min Scan# 648
Delta R.T. -0.01 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion:165 Resp: 181717
Ion Ratio Lower Upper
165 100
63 33.9 59.4 89.0#
89 36.6 50.8 76.2#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 39.44 ng

RT: 9.94 min Scan# 670

Delta R.T. -0.01 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion:154 Resp: 769299

Ion Ratio Lower Upper

154 100

153 111.4 91.0 136.6




152 51.1 44.2 66.2



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 34.99 ng

RT: 9.85 min Scan# 662

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:138 Resp: 188065

Ion Ratio Lower Upper

138 100

108 11.3 8.1 12.1



92 125.3 92.5 138.7




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 43.06 ng
RT: 9.96 min Scan# 671
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Instrument :
BNA_F
ClientSampled :

Tot Ion:184 Resp: 87429
Ion Ratio Lower Upper
184 100
63 79.2 58.3 87.5
154 67.6 48.6 73.0




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 38.00 ng
RT: 10.12 min Scan# 685
Delta R.T. -0.00 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion:168 Resp: 992574
Ion Ratio Lower Upper
168 100
139 32.6 33.4 50.2#
169 12.9 11.1 16.7



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 36.39 ng
RT: 10.00 min Scan# 675
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51




Instrument :
BNA_F
ClientSampled :


Tot Ion	Ratio	Lower	Upper
139	100		
109	74.2	54.9	94.9
65	131.8	102.2	142.2



#56
2,4-Dinitrotoluene
Concen: 39.38 ng
RT: 10.09 min Scan# 683
Delta R.T. -0.01 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion	Ratio	Lower	Upper
165	100		
63	26.4	29.0	43.4#
89	43.9	43.1	64.7
182	2.5	1.9	2.9









#57
Fluorene
 Concen: 36.59 ng
 RT: 10.46 min Scan# 715
 Delta R.T. -0.00 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:166 Resp: 733693
 Ion Ratio Lower Upper
 166 100
 165 98.7 80.0 120.0
 167 12.7 10.8 16.2



#58
2,3,4,6-Tetrachlorophenol
 Concen: 38.98 ng
 RT: 10.23 min Scan# 695
 Delta R.T. -0.01 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion:232 Resp: 209667
 Ion Ratio Lower Upper
 232 100
 131 40.8 38.4 57.6
 130 2.1 1.8 2.8
 166 29.8 23.9 35.9





#59

Diethylphthalate

Concen: 37.99 ng

RT: 10.33 min Scan# 704

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 778975

Ion Ratio Lower Upper

149 100

177 23.1 15.7 23.5




150 12.0 10.2 15.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 35.66 ng

RT: 10.45 min Scan# 714

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:204 Resp: 358636

Ion Ratio Lower Upper

204 100

206 32.6 26.1 39.1



141 62.7 40.6 61.0#




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 37.44 ng

RT: 10.46 min Scan# 715

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion:138 Resp: 188923

Ion Ratio Lower Upper

138 100

92 56.7 26.9 66.9




108 74.9 49.1 89.1



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 40.53 ng

RT: 10.61 min Scan# 728

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Tgt Ion: 77 Resp: 848056


Ion Ratio Lower Upper

77 100

182 26.0 12.5 52.5

105 23.0 3.7 43.7

51 35.4 11.9 51.9






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 11.38 min Scan# 796
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:188 Resp: 541091
 Ion Ratio Lower Upper
 188 100
 94 9.1 9.2 13.8#
 80 9.3 10.5 15.7#



#64
 4,6-Dinitro-2-methylphenol
 Concen: 41.76 ng
 RT: 10.49 min Scan# 718
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion:198 Resp: 124484
 Ion Ratio Lower Upper
 198 100
 51 59.9 77.1 117.1#
 105 45.8 44.8 84.8





#65

n-Nitrosodiphenylamine

Concen: 37.46 ng

RT: 10.56 min Scan# 724

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampled :


Tot Ion:169 Resp: 614495

Ion Ratio Lower Upper

169 100

168 66.7 53.8 80.8




167 36.9 29.8 44.8



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 42.10 ng

RT: 10.94 min Scan# 757

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:248 Resp: 244851

Ion Ratio Lower Upper

248 100

250 97.6 78.2 117.4



141 48.9 71.9 107.9#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 36.45 ng
RT: 11.00 min Scan# 762
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Instrument :
BNA_F
ClientSampled :

Tot Ion	Ratio	Lower	Upper
284	100		
142	43.6	17.9	26.9
249	34.1	23.6	35.4




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 33.57 ng
RT: 11.09 min Scan# 770
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion	Ratio	Lower	Upper
200	100		
173	28.5	7.8	47.8
215	47.0	24.3	64.3



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 42.87 ng
 RT: 11.19 min Scan# 779
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampled :


Tot Ion:266 Resp: 158619
 Ion Ratio Lower Upper
 266 100
 268 66.5 50.3 75.5
 264 61.6 48.2 72.2



#70
 Phenanthrene
 Concen: 36.52 ng
 RT: 11.41 min Scan# 798
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion:178 Resp: 1026828
 Ion Ratio Lower Upper
 178 100
 176 19.6 15.9 23.9
 179 15.9 12.6 18.8





#71

Anthracene

Concen: 42.29 ng

RT: 11.46 min Scan# 803

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Instrument :

BNA_F

ClientSampleId :

Tot Ion:178 Resp: 1179937



Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.8	23.6
179	15.7	12.7	19.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 37.69 ng

RT: 11.61 min Scan# 816


Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Tgt Ion:167 Resp: 954410


Ion	Ratio	Lower	Upper
167	100		
166	21.9	18.1	27.1
139	13.2	11.5	17.3




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 39.36 ng

RT: 11.96 min Scan# 846

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :





Tot Ion:149 Resp: 1129408

Ion Ratio Lower Upper

149 100

150 9.3 7.5 11.3

104 4.5 4.0 6.0



#74

Fluoranthene

Concen: 41.55 ng

RT: 12.60 min Scan# 902

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51




Tgt Ion:202 Resp: 1106646


Ion Ratio Lower Upper

202 100

101 8.9 0.0 29.5

203 17.9 0.0 37.8





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.01 min Scan# 1026

Delta R.T. -0.03 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion:240 Resp: 339528

Ion Ratio Lower Upper




240 100

120 13.0 12.1 18.1

236 25.9 21.0 31.6



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 41.56 ng

RT: 12.71 min Scan# 912

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:184 Resp: 414279

Ion Ratio Lower Upper



184 100


185 18.0 13.1 19.7

183 11.4 8.9 13.3



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 185.00 (184.70 to 185.70): E
Ion 183.00 (182.70 to 183.70): E









#77
Pvrene
Concen: 43.92 ng
RT: 12.82 min Scan# 922
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51




Instrument :
BNA_F
ClientSampleId :


Tot Ion:202 Resp: 1131650
Ion Ratio Lower Upper
202 100
200 21.2 17.6 26.4
203 17.7 14.0 21.0



#78
Terphenyl-d14
Concen: 88.24 ng
RT: 12.97 min Scan# 935
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion:244 Resp: 1378368
Ion Ratio Lower Upper
244 100
212 6.9 6.5 9.7
122 7.2 9.5 14.3#









#79
Butylbenzylphthalate
Concen: 38.30 ng
RT: 13.44 min Scan# 976
Delta R.T. -0.03 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
149	100		
91	86.9	57.4	86.2#
206	15.5	15.4	23.2



#80
Benzo(a)anthracene
Concen: 38.96 ng
RT: 14.00 min Scan# 1025
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.0	22.5	33.7
229	20.1	16.4	24.6





#81

3,3'-Dichlorobenzidine

Concen: 38.15 ng

RT: 13.97 min Scan# 1022

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51

Instrument :

BNA_F

ClientSampleId :


Tot Ion:252 Resp: 266871

Ion Ratio Lower Upper

252 100

254 63.2 52.1 78.1




126 12.9 7.5 11.3#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E



#82

Chrysene

Concen: 41.35 ng

RT: 14.05 min Scan# 1029

Delta R.T. -0.02 min

Lab File: BF091324.D

Acq: 29 Nov 2016 13:51


Tgt Ion:228 Resp: 812427

Ion Ratio Lower Upper

228 100

226 29.8 25.0 37.4



229 19.2 16.0 24.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 38.88 ng
 RT: 14.01 min Scan# 1026
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:149 Resp: 475477
 Ion Ratio Lower Upper
 149 100
 167 28.8 19.6 29.4
 279 3.9 1.9 2.9#



#84
 Di-n-octyl phthalate
 Concen: 39.51 ng
 RT: 14.62 min Scan# 1079
 Delta R.T. -0.02 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion:149 Resp: 731260
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.3 1.9
 43 13.7 11.9 17.9









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 48.14 ng
 RT: 16.85 min Scan# 1274
 Delta R.T. -0.04 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
276	100		
138	21.7	21.0	31.4
277	24.9	20.2	30.2



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.44 min Scan# 1151
 Delta R.T. -0.03 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.3	18.8	28.2
265	22.2	17.0	25.6









#87
Benzo(b)fluoranthene
Concen: 71.96 ng
RT: 15.07 min Scan# 1118
Delta R.T. 0.01 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51




Instrument :
BNA_F
ClientSampled :


Tot Ion:252 Resp: 1429133
Ion Ratio Lower Upper
252 100
253 22.5 18.2 27.2
125 7.9 8.6 13.0#



#88
Benzo(k)fluoranthene
Concen: 85.58 ng
RT: 15.07 min Scan# 1118
Delta R.T. -0.02 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Tgt Ion:252 Resp: 1429133
Ion Ratio Lower Upper
252 100
253 22.5 17.8 26.8
125 7.9 9.6 14.4#







#89
Benzo(a)pyrene
Concen: 38.74 ng
RT: 15.39 min Scan# 1146
Delta R.T. -0.03 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51

Instrument :
BNA_F
ClientSampleId :



Tot Ion:252 Resp: 690901
Ion Ratio Lower Upper
252 100
253 21.9 18.2 27.4
125 10.6 10.4 15.6



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E




Time-->




#90
Dibenzo(a,h)anthracene
Concen: 44.32 ng
RT: 16.87 min Scan# 1276
Delta R.T. -0.04 min
Lab File: BF091324.D
Acq: 29 Nov 2016 13:51


Tgt Ion:278 Resp: 731015
Ion Ratio Lower Upper
278 100
139 14.2 15.1 22.7#
279 24.0 19.0 28.6




Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



Time-->





#91
 Benzo(a,h,i)perylene
 Concen: 44.88 ng
 RT: 17.27 min Scan# 1311
 Delta R.T. -0.04 min
 Lab File: BF091324.D
 Acq: 29 Nov 2016 13:51

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
276	100		
277	23.2	19.0	28.4
138	20.0	15.4	23.0

