

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112917\
 Data File : BF100992.D
 Acq On : 29 Nov 2017 21:38
 Operator : SJ/JU
 Sample : I6610-16
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-10(8-10)-20171127

Manual Integrations
 APPROVED

Sohil
 11/30/2017 4:30:11 PM

Quant Time: Nov 30 07:13:16 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 28 15:21:02 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	56333	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	208013	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	85346	20.00	ng	0.00
63) Phenanthrene-d10	11.38	188	143527	20.00	ng	0.00
75) Chrysene-d12	14.03	240	116057	20.00	ng	0.00
86) Perylene-d12	15.50	264	91138	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	370894	105.54	ng	0.00
7) Phenol-d6	6.49	99	439897	101.51	ng	0.00
23) Nitrobenzene-d5	7.42	82	256063	81.39	ng	-0.01
41) 2,4,6-Tribromophenol	10.68	330	93163	107.12	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	491990	87.78	ng	0.00
78) Terphenyl-d14	12.96	244	396001	78.40	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.50	94	13999	3.077	ng	78
31) Naphthalene	8.16	128	75006	7.486	ng	99
37) 2-Methylnaphthalene	8.85	142	50126	7.536	ng	98
48) Acenaphthylene	9.75	152	36918	4.160	ng	96
49) Dimethylphthalate	9.60	163	33696	5.171	ng	# 98
51) Acenaphthene	9.92	154	65057	12.054	ng	99
54) Dibenzofuran	10.10	168	57750	7.634	ng	97
57) Fluorene	10.44	166	87956	15.872	ng	100
70) Phenanthrene	11.41	178	681263	90.151	ng	98
71) Anthracene	11.46	178	202923	26.467	ng	98
72) Carbazole	11.61	167	51653	7.302	ng	99
74) Fluoranthene	12.60	202	787859	98.373	ng	96
77) Pvrene	12.83	202	701771	84.827	ng	99
79) Butylbenzylphthalate	13.43	149	8580	2.543	ng	# 95
80) Benzo(a)anthracene	14.02	228	359728	51.471	ng	99
82) Chrysene	14.05	228	292859	43.272	ng	97
83) Bis(2-ethylhexyl)phthalate	13.98	149	33462	7.541	ng	96
85) Indeno(1,2,3-cd)pyrene	16.98	276	122309	22.343	ng	# 90
87) Benzo(b)fluoranthene	15.07	252	306944m	56.847	ng	
88) Benzo(k)fluoranthene	15.10	252	107398m	21.051	ng	
89) Benzo(a)pyrene	15.44	252	217023	45.338	ng	97
90) Dibenzo(a,h)anthracene	16.99	278	32870m	8.397	ng	
91) Benzo(g,h,i)perylene	17.43	276	115765	30.210	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA F\DATA\BF112917\
Data File : BF100992.D
Acq On : 29 Nov 2017 21:38
Operator : SJ/JU
Sample : I6610-16
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Manual Integrations
APPROVED

Sohil
11/30/2017 4:30:11 PM

Quant Time: Nov 30 07:13:16 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 28 15:21:02 2017
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.85 min Scan# 810
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
152	100		
150	155.8	124.6	186.8
115	58.8	50.1	75.1

Manual Integrations
APPROVED




Sohil
11/30/2017 4:30:11 PM




#5

2-Fluorophenol
Concen: 105.540 ng
RT: 5.47 min Scan# 576
Delta R.T. -0.00 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
112	100		
64	60.1	47.9	71.9
63	31.4	23.7	35.5





#7

Phenol-d6

Concen: 101.510 ng

RT: 6.49 min Scan# 749

Delta R.T. -0.00 min

Lab File: BF100992.D

Acq: 29 Nov 2017 21:38

Instrument :

BNA_F





ClientSampleId :

SB-10(8-10)-20171127

Tot Ion:	99	Resp:	439897
Ion Ratio	Lower	Upper	
99	100		
42	19.6	14.5	21.7
71	34.1	25.4	38.0

Manual Integrations
APPROVED

Sohil
11/30/2017 4:30:11 PM



#10

Phenol

Concen: 3.077 ng




RT: 6.50 min Scan# 750

Delta R.T. -0.01 min

Lab File: BF100992.D

Acq: 29 Nov 2017 21:38

Tgt Ion:	94	Resp:	13999
Ion Ratio	Lower	Upper	
94	100		
65	34.8	7.9	47.9
66	53.2	16.2	56.2









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.13 min Scan# 1028
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	11.4	8.9	13.3	
54	8.4	6.6	9.8	
68	5.1	5.0	7.4	




Manual Integrations
APPROVED


Sohil
11/30/2017 4:30:11 PM



#23
Nitrobenzene-d5
Concen: 81.385 ng
RT: 7.42 min Scan# 906
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	45.5	36.1	54.1	
54	55.0	41.4	62.2	






#31
Naphthalene
Concen: 7.486 ng
RT: 8.16 min Scan# 1032
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127




Tot Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.8	13.2
127	13.3	10.9	16.3

Manual Integrations
APPROVED

Sohil
11/30/2017 4:30:11 PM




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#37
2-Methylnaphthalene
Concen: 7.536 ng
RT: 8.85 min Scan# 1149
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
142	100		
141	86.7	70.8	106.2
115	30.8	26.1	39.1



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E









#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.89 min Scan# 1327
 Delta R.T. -0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Instrument :
 BNA_F
 ClientSampleId :
 SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
164	100		
162	106.5	86.1	129.1
160	44.6	38.3	57.5




Manual Integrations
 APPROVED


Sohil
 11/30/2017 4:30:11 PM



#41
 2,4,6-Tribromophenol
 Concen: 107.124 ng
 RT: 10.68 min Scan# 1461
 Delta R.T. -0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
330	100		
332	97.0	77.0	115.6
141	30.3	29.9	44.9









#44
 2-Fluorobiphenyl
 Concen: 87.779 ng
 RT: 9.21 min Scan# 1211
 Delta R.T. -0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Instrument :
 BNA_F
 ClientSampleId :
 SB-10(8-10)-20171127

Tot Ion	Ratio	Resp	Lower	Upper
172	100	491990		
171	35.1		29.7	44.5
170	23.0		19.6	29.4




Manual Integrations
 APPROVED


Sohil
 11/30/2017 4:30:11 PM



#48
 Acenaphthylene
 Concen: 4.160 ng
 RT: 9.75 min Scan# 1303
 Delta R.T. -0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	36918		
151	18.7		16.3	24.5
153	14.8		10.7	16.1









#49
Dimethylphthalate
Concen: 5.171 ng
RT: 9.60 min Scan# 1278
Delta R.T. -0.02 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Tot Ion	Ratio	Resp	Lower	Upper
163	100			
194	4.7		2.7	4.1#
164	9.7		8.2	12.2




Manual Integrations
APPROVED


Sohil
11/30/2017 4:30:11 PM



#51
Acenaphthene
Concen: 12.054 ng
RT: 9.92 min Scan# 1332
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Resp	Lower	Upper
154	100			
153	112.6		91.2	136.8
152	53.8		43.8	65.8









#54
Dibenzenofuran
Concen: 7.634 ng
RT: 10.10 min Scan# 1362
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
168	100		
139	35.1	30.1	45.1
169	14.3	10.9	16.3




Manual Integrations
APPROVED


Sohil
11/30/2017 4:30:11 PM



#57
Fluorene
Concen: 15.872 ng
RT: 10.44 min Scan# 1420
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.6	79.8	119.8
167	13.3	10.6	16.0









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.38 min Scan# 1580
Delta R.T. -0.00 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Tot Ion	Ratio	Resp Lower	Upper
188	100		
94	7.7	8.5	12.7#
80	9.0	9.2	13.8#




Manual Integrations
APPROVED


Sohil
11/30/2017 4:30:11 PM



#70
Phenanthrene
Concen: 90.151 ng
RT: 11.41 min Scan# 1585
Delta R.T. -0.00 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Resp Lower	Upper
178	100		
176	18.8	15.8	23.8
179	15.3	13.0	19.6









#71
 Anthracene
 Concen: 26.467 ng
 RT: 11.46 min Scan# 1593
 Delta R.T. -0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Instrument :
 BNA_F
 ClientSampleId :
 SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
178	100		
176	17.9	15.4	23.2
179	15.8	12.6	19.0




Manual Integrations
 APPROVED


Sohil
 11/30/2017 4:30:11 PM



#72
 Carbazole
 Concen: 7.302 ng
 RT: 11.61 min Scan# 1619
 Delta R.T. -0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.6	26.4
139	13.2	10.9	16.3





#74

Fluoranthene

Concen: 98.373 ng

RT: 12.60 min Scan# 1788

Delta R.T. 0.01 min

Lab File: BF100992.D

Acq: 29 Nov 2017 21:38

Instrument :

BNA_F





ClientSampleId :

SB-10(8-10)-20171127

Tot Ion	Ratio	Resp Lower	Upper
202	100		
101	11.3	0.0	34.1
203	17.4	0.0	38.1

Manual Integrations
APPROVED

Sohil
11/30/2017 4:30:11 PM



#75

Chrysene-d12

Concen: 20.000 ng




RT: 14.03 min Scan# 2030


Delta R.T. -0.00 min

Lab File: BF100992.D

Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Resp Lower	Upper
240	100		
120	10.4	9.5	14.3
236	26.4	20.7	31.1









#77
Pvrene
Concen: 84.827 ng
RT: 12.83 min Scan# 1827
Delta R.T. 0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	21.1	17.4	26.2
203	17.5	14.3	21.5




Manual Integrations
APPROVED


Sohil
11/30/2017 4:30:11 PM



#78
Terphenyl-d14
Concen: 78.401 ng
RT: 12.96 min Scan# 1849
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	6.9	5.4	8.0
122	11.5	11.0	16.4









#79
Butylbenzylphthalate
Concen: 2.543 ng
RT: 13.43 min Scan# 1929
Delta R.T. -0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
149	100		
91	68.4	56.8	85.2
206	22.6	14.1	21.1




Manual Integrations
APPROVED


Sohil
11/30/2017 4:30:11 PM



#80
Benzo(a)anthracene
Concen: 51.471 ng
RT: 14.02 min Scan# 2028
Delta R.T. -0.00 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.3	22.6	33.8
229	20.3	15.8	23.6









#82
 Chrysene
 Concen: 43.272 ng
 RT: 14.05 min Scan# 2034
 Delta R.T. -0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Instrument :
 BNA_F
 ClientSampleId :
 SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
228	100		
226	29.8	25.0	37.6
229	21.6	16.2	24.4




Manual Integrations
 APPROVED


Sohil
 11/30/2017 4:30:11 PM



#83
 Bis(2-ethylhexyl)phthalate
 Concen: 7.541 ng
 RT: 13.98 min Scan# 2022
 Delta R.T. -0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
149	100		
167	24.6	21.3	31.9
279	3.1	3.0	4.4









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 22.343 ng
 RT: 16.98 min Scan# 2532
 Delta R.T. 0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Instrument :
 BNA_F
 ClientSampleId :
 SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
276	100		
138	21.1	25.0	37.6#
277	25.0	20.1	30.1




Manual Integrations
 APPROVED


Sohil
 11/30/2017 4:30:11 PM



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.50 min Scan# 2280
 Delta R.T. 0.01 min
 Lab File: BF100992.D
 Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
264	100		
260	21.7	19.2	28.8
265	21.8	17.5	26.3









#87
Benzo(b)fluoranthene
Concen: 56.847 ng m
RT: 15.07 min Scan# 2207
Delta R.T. 0.01 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Instrument :
BNA_F
ClientSampleId :
SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.6	17.0	25.6
125	11.5	9.0	13.6




Manual Integrations
APPROVED


Sohil
11/30/2017 4:30:11 PM



#88
Benzo(k)fluoranthene
Concen: 21.051 ng m
RT: 15.10 min Scan# 2212
Delta R.T. -0.00 min
Lab File: BF100992.D
Acq: 29 Nov 2017 21:38

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.7	17.9	26.9
125	8.6	10.2	15.2#





#89

Benzo(a)pyrene

Concen: 45.338 ng

RT: 15.44 min Scan# 2270

Delta R.T. 0.01 min

Lab File: BF100992.D

Acq: 29 Nov 2017 21:38

Instrument :

BNA_F





ClientSampleId :

SB-10(8-10)-20171127

Tot Ion:	252	Resp:	217023
Ion Ratio	Lower	Upper	
252	100		
253	23.3	17.1	25.7
125	11.3	9.8	14.6

Manual Integrations
APPROVED

Sohil
11/30/2017 4:30:11 PM



#90

Dibenzo(a,h)anthracene

Concen: 8.397 ng m




RT: 16.99 min Scan# 2533




Delta R.T. -0.00 min

Lab File: BF100992.D

Acq: 29 Nov 2017 21:38

Tgt Ion:	278	Resp:	32870
Ion Ratio	Lower	Upper	
278	100		
139	19.3	15.9	23.9
279	28.5	19.0	28.4#





#91

Benzo(a,h,i)perylene

Concen: 30.210 ng

RT: 17.43 min Scan# 2609

Delta R.T. 0.01 min

Lab File: BF100992.D

Acq: 29 Nov 2017 21:38

Instrument :

BNA_F

ClientSampleId :

SB-10(8-10)-20171127

Tot Ion	Ratio	Lower	Upper
276	100		
277	23.6	17.9	26.9
138	19.8	21.8	32.8

Manual Integrations
APPROVED

Sohil
11/30/2017 4:30:11 PM

