

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112922\
 Data File : BF131464.D
 Acq On : 29 Nov 2022 19:14
 Operator : CG\JU
 Sample : N5796-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MH-15

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 11/30/2022
 Supervised By : mohammad ahmed 12/01/2022

Quant Time: Nov 30 01:08:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 00:18:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.945	152	132076	20.000	ng	0.00
21) Naphthalene-d8	8.228	136	499683	20.000	ng	-0.02
39) Acenaphthene-d10	9.998	164	287897	20.000	ng	-0.01
64) Phenanthrene-d10	11.492	188	530694	20.000	ng	-0.01
76) Chrysene-d12	14.151	240	402872	20.000	ng	-0.02
86) Perylene-d12	15.680	264	311347	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.575	112	964697	139.013	ng	0.02
7) Phenol-d6	6.569	99	1143716	129.173	ng	0.00
23) Nitrobenzene-d5	7.510	82	836021	84.339	ng	-0.02
42) 2,4,6-Tribromophenol	10.792	330	413862	170.869	ng	-0.01
45) 2-Fluorobiphenyl	9.316	172	1547795	78.214	ng	-0.01
79) Terphenyl-d14	13.092	244	1931130	78.385	ng	-0.01
Target Compounds						Qvalue
32) Benzoic acid	7.922	122	8235m	7.371	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112922\
Data File : BF131464.D
Acq On : 29 Nov 2022 19:14
Operator : CG\JU
Sample : N5796-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MH-15

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 11/30/2022
Supervised By :mohammad ahmed 12/01/2022

Quant Time: Nov 30 01:08:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 00:18:36 2022
Response via : Initial Calibration

