

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112923\
 Data File : BF136299.D
 Acq On : 29 Nov 2023 23:30
 Operator : CG\JU
 Sample : 05253-04DL 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 L-3(195FT)(0-5)DL

Quant Time: Nov 30 00:31:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 28 03:04:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.804	152	123526	20.000	ng	0.00
21) Naphthalene-d8	8.081	136	467194	20.000	ng	0.00
39) Acenaphthene-d10	9.839	164	193812	20.000	ng	0.00
64) Phenanthrene-d10	11.327	188	304790	20.000	ng	0.00
76) Chrysene-d12	13.986	240	279844	20.000	ng	0.00
86) Perylene-d12	15.468	264	237308	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	143535	16.173	ng	-0.01
7) Phenol-d6	6.428	99	174777	15.750	ng	-0.02
23) Nitrobenzene-d5	7.357	82	101822	11.517	ng	-0.02
42) 2,4,6-Tribromophenol	10.628	330	31303	13.211	ng	-0.01
45) 2-Fluorobiphenyl	9.157	172	199097	13.121	ng	-0.01
79) Terphenyl-d14	12.927	244	192406	7.931	ng	0.00
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.357	70	13492	2.033	ng	# 76
25) Isophorone	7.357	82	101774	6.384	ng	# 68
31) Naphthalene	8.110	128	2886350	109.915	ng	99
33) 4-Chloroaniline	8.110	127	417720	46.797	ng	# 47
35) Caprolactam	8.798	113	20392	10.763	ng	# 1
37) 2-Methylnaphthalene	8.798	142	917297	55.686	ng	100
38) 1-Methylnaphthalene	8.892	142	595024	37.029	ng	94
46) 1,1'-Biphenyl	9.257	154	172191	10.075	ng	96
49) Acenaphthylene	9.698	152	130017	6.864	ng	# 93
52) Acenaphthene	9.869	154	307420	25.683	ng	95
55) Dibenzofuran	10.039	168	37788	2.122	ng	# 79
58) Fluorene	10.386	166	233771	16.732	ng	97
71) Phenanthrene	11.357	178	1164693	67.989	ng	100
72) Anthracene	11.404	178	288570	16.798	ng	97
75) Fluoranthene	12.551	202	492234	28.305	ng	96
78) Pyrene	12.780	202	883313	28.821	ng	98
81) Benzo(a)anthracene	14.010	228	301156	14.488	ng	94
83) Chrysene	14.010	228	301156	15.006	ng	96
85) Di-n-octyl phthalate	14.586	149	229	5.173	ng	# 1
87) Indeno(1,2,3-cd)pyrene	16.962	276	66476	6.218	ng	98
88) Benzo(b)fluoranthene	15.033	252	277996	16.697	ng	# 95
89) Benzo(k)fluoranthene	15.033	252	277996	17.580	ng	# 94
90) Benzo(a)pyrene	15.404	252	218090	16.024	ng	98
91) Dibenzo(a,h)anthracene	16.974	278	22583	4.210	ng	93
92) Benzo(g,h,i)perylene	17.415	276	71413	6.702	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112923\
Data File : BF136299.D
Acq On : 29 Nov 2023 23:30
Operator : CG\JU
Sample : 05253-04DL 5X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL

Quant Time: Nov 30 00:31:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 28 03:04:20 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.804 min Scan# 803
 Delta R.T. -0.006 min
 Lab File: BF136299.D
 Acq: 29 Nov 2023 23:30




Instrument :
 BNA_F
 ClientSampleId :
 L-3(195FT)(0-5)DL


Tgt Ion:152 Resp: 123526
 Ion Ratio Lower Upper
 152 100
 150 168.5 137.8 206.8
 115 66.7 52.7 79.1



#5
 2-Fluorophenol
 Concen: 16.173 ng
 RT: 5.428 min Scan# 568
 Delta R.T. -0.012 min
 Lab File: BF136299.D
 Acq: 29 Nov 2023 23:30

Tgt Ion:112 Resp: 143535
 Ion Ratio Lower Upper
 112 100
 64 58.3 48.2 72.4
 63 30.8 27.0 40.6









#7
Phenol-d6
Concen: 15.750 ng
RT: 6.428 min Scan# 71
Delta R.T. -0.018 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion: 99 Resp: 174777
Ion Ratio Lower Upper
99 100
42 17.0 15.0 22.6
71 33.7 28.9 43.3



#19
n-Nitroso-di-n-propylamine
Concen: 2.033 ng
RT: 7.357 min Scan# 896
Delta R.T. 0.129 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30

Tgt Ion: 70 Resp: 13492
Ion Ratio Lower Upper
70 100
42 44.9 47.0 70.4#
101 0.0 7.8 11.8#
130 3.1 16.8 25.2#





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.081 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30

Instrument :

BNA_F

ClientSampleId :

L-3(195FT)(0-5)DL

Tgt Ion:136 Resp: 467194





Ion Ratio Lower Upper

136 100

137 10.2 7.4 11.2

54 7.6 5.5 8.3

68 5.3 3.7 5.5



#23

Nitrobenzene-d5

Concen: 11.517 ng

RT: 7.357 min Scan# 896

Delta R.T. -0.018 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30




Tgt Ion: 82 Resp: 101822


Ion Ratio Lower Upper

82 100

128 43.0 33.4 50.2

54 51.8 41.7 62.5









#25
Isophorone
Concen: 6.384 ng
RT: 7.357 min Scan# 896
Delta R.T. -0.271 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion: 82 Resp: 101774
Ion Ratio Lower Upper
82 100
95 0.0 5.5 8.3#
138 0.0 12.8 19.2#



#31
Naphthalene
Concen: 109.915 ng
RT: 8.110 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30

Tgt Ion: 128 Resp: 2886350
Ion Ratio Lower Upper
128 100
129 9.2 7.3 10.9
127 15.0 11.8 17.6









#33
4-Chloroaniline
Concen: 46.797 ng
RT: 8.110 min Scan# 1141
Delta R.T. -0.047 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion:127 Resp: 417720
Ion Ratio Lower Upper
127 100
129 61.0 23.4 35.2#
65 0.0 23.8 35.6#
92 0.0 14.6 21.8#



#35
Caprolactam
Concen: 10.763 ng
RT: 8.798 min Scan# 1141
Delta R.T. 0.259 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30

Tgt Ion:113 Resp: 20392
Ion Ratio Lower Upper
113 100
55 8.3 143.1 183.1#
56 0.0 103.6 143.6#









#37
2-Methylnaphthalene
Concen: 55.686 ng
RT: 8.798 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion:142 Resp: 917297
Ion Ratio Lower Upper
142 100
141 97.5 77.8 116.8



#38
1-Methylnaphthalene
Concen: 37.029 ng
RT: 8.892 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30

Tgt Ion:142 Resp: 595024
Ion Ratio Lower Upper
142 100
141 98.5 83.8 125.6









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.839 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion:164 Resp: 193812
Ion Ratio Lower Upper
164 100
162 105.9 89.5 134.3
160 46.3 39.6 59.4



#42
2,4,6-Tribromophenol
Concen: 13.211 ng
RT: 10.628 min Scan# 1452
Delta R.T. -0.012 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30

Tgt Ion:330 Resp: 31303
Ion Ratio Lower Upper
330 100
332 93.6 73.0 109.6
141 34.5 24.2 36.2









#45
2-Fluorobiphenyl
Concen: 13.121 ng
RT: 9.157 min Scan# 1119
Delta R.T. -0.012 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion:172 Resp: 199097
Ion Ratio Lower Upper
172 100
171 36.7 31.5 47.3
170 24.5 20.8 31.2



#46
1,1'-Biphenyl
Concen: 10.075 ng
RT: 9.257 min Scan# 1219
Delta R.T. -0.012 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30

Tgt Ion:154 Resp: 172191
Ion Ratio Lower Upper
154 100
153 42.9 24.6 64.6
76 12.4 0.0 29.4









#49
Acenaphthylene
Concen: 6.864 ng
RT: 9.698 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion:152 Resp: 130017
Ion Ratio Lower Upper
152 100
151 19.8 17.5 26.3
153 16.3 9.6 14.4#



#52
Acenaphthene
Concen: 25.683 ng
RT: 9.869 min Scan# 1323
Delta R.T. -0.012 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30

Tgt Ion:154 Resp: 307420
Ion Ratio Lower Upper
154 100
153 119.7 100.8 151.2
152 56.5 47.0 70.6









#55
Dibenzenofuran
Concen: 2.122 ng
RT: 10.039 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion:168 Resp: 37788
Ion Ratio Lower Upper
168 100
139 44.0 30.9 46.3
169 34.9 9.1 13.7#



#58
Fluorene
Concen: 16.732 ng
RT: 10.386 min Scan# 1411
Delta R.T. -0.012 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30

Tgt Ion:166 Resp: 233771
Ion Ratio Lower Upper
166 100
165 104.5 86.2 129.4
167 13.2 9.1 13.7





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.327 min Scan# 1571

Delta R.T. -0.006 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30

Instrument :

BNA_F

ClientSampleId :

L-3(195FT)(0-5)DL





Tgt Ion:188 Resp: 304790

Ion Ratio Lower Upper

188 100

94 8.8 6.7 10.1

80 9.8 7.2 10.8



#71

Phenanthrene

Concen: 67.989 ng

RT: 11.357 min Scan# 1576

Delta R.T. -0.006 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30




Tgt Ion:178 Resp: 1164693


Ion Ratio Lower Upper

178 100

176 20.2 16.2 24.2

179 13.3 10.3 15.5









#72
 Anthracene
 Concen: 16.798 ng
 RT: 11.404 min Scan# 11
 Delta R.T. -0.012 min
 Lab File: BF136299.D
 Acq: 29 Nov 2023 23:30




Instrument :
 BNA_F
 ClientSampleId :
 L-3(195FT)(0-5)DL


Tgt Ion:178 Resp: 288570
 Ion Ratio Lower Upper
 178 100
 176 18.7 15.5 23.3
 179 15.4 10.5 15.7



#75
 Fluoranthene
 Concen: 28.305 ng
 RT: 12.551 min Scan# 1779
 Delta R.T. -0.006 min
 Lab File: BF136299.D
 Acq: 29 Nov 2023 23:30

Tgt Ion:202 Resp: 492234
 Ion Ratio Lower Upper
 202 100
 101 8.8 0.0 27.8
 203 16.4 0.0 34.5





#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.986 min Scan# 20

Delta R.T. -0.006 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30

Instrument :

BNA_F

ClientSampleId :

L-3(195FT)(0-5)DL





Tgt Ion:240 Resp: 279844

Ion Ratio Lower Upper

240 100

120 8.8 7.4 11.2

236 26.1 21.1 31.7



#78

Pyrene

Concen: 28.821 ng

RT: 12.780 min Scan# 1818

Delta R.T. -0.006 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30




Tgt Ion:202 Resp: 883313


Ion Ratio Lower Upper

202 100

200 22.5 18.4 27.6

203 15.9 11.8 17.6









#79
 Terphenyl-d14
 Concen: 7.931 ng
 RT: 12.927 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF136299.D
 Acq: 29 Nov 2023 23:30




Instrument :
 BNA_F
 ClientSampleId :
 L-3(195FT)(0-5)DL


Tgt Ion:244 Resp: 192406
 Ion Ratio Lower Upper
 244 100
 212 5.7 4.6 7.0
 122 8.4 7.3 10.9



#81
 Benzo(a)anthracene
 Concen: 14.488 ng
 RT: 14.010 min Scan# 2027
 Delta R.T. 0.029 min
 Lab File: BF136299.D
 Acq: 29 Nov 2023 23:30

Tgt Ion:228 Resp: 301156
 Ion Ratio Lower Upper
 228 100
 226 30.9 22.6 34.0
 229 20.5 14.0 21.0





#83

Chrysene

Concen: 15.006 ng

RT: 14.010 min Scan# 20

Delta R.T. -0.006 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30

Instrument :

BNA_F

ClientSampleId :

L-3(195FT)(0-5)DL





Tgt Ion:228 Resp: 301156

Ion Ratio Lower Upper

228 100

226 30.9 25.5 38.3

229 20.5 13.9 20.9



#85

Di-n-octyl phthalate

Concen: 5.173 ng

RT: 14.586 min Scan# 2125

Delta R.T. -0.012 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30




Tgt Ion:149 Resp: 229


Ion Ratio Lower Upper

149 100

167 0.0 1.0 1.6#

43 354.6 7.2 10.8#





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.468 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30

Instrument :

BNA_F

ClientSampleId :

L-3(195FT)(0-5)DL





Tgt Ion:264 Resp: 237308

Ion Ratio Lower Upper

264 100

260 25.0 18.6 27.8

265 22.6 15.7 23.5



#87

Indeno(1,2,3-cd)pyrene

Concen: 6.218 ng

RT: 16.962 min Scan# 2529

Delta R.T. -0.018 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30




Tgt Ion:276 Resp: 66476


Ion Ratio Lower Upper

276 100

138 21.6 17.7 26.5

277 25.4 19.6 29.4





#88

Benzo(b)fluoranthene

Concen: 16.697 ng

RT: 15.033 min Scan# 2201

Delta R.T. -0.006 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30

Instrument :

BNA_F

ClientSampleId :

L-3(195FT)(0-5)DL





Tgt Ion:252 Resp: 277996

Ion Ratio Lower Upper

252 100

253 22.3 16.2 24.2

125 11.0 6.9 10.3#



#89

Benzo(k)fluoranthene

Concen: 17.580 ng

RT: 15.033 min Scan# 2201

Delta R.T. -0.035 min

Lab File: BF136299.D

Acq: 29 Nov 2023 23:30




Tgt Ion:252 Resp: 277996


Ion Ratio Lower Upper

252 100

253 22.3 15.6 23.4

125 11.0 7.0 10.4#









#90
Benzo(a)pyrene
Concen: 16.024 ng
RT: 15.404 min Scan# 218090
Delta R.T. -0.005 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30




Instrument :
BNA_F
ClientSampleId :
L-3(195FT)(0-5)DL


Tgt Ion:252 Resp: 218090
Ion Ratio Lower Upper
252 100
253 20.6 15.5 23.3
125 9.9 7.9 11.9



#91
Dibenzo(a,h)anthracene
Concen: 4.210 ng
RT: 16.974 min Scan# 2531
Delta R.T. -0.029 min
Lab File: BF136299.D
Acq: 29 Nov 2023 23:30


Tgt Ion:278 Resp: 22583
Ion Ratio Lower Upper
278 100
139 12.1 11.7 17.5
279 26.2 17.8 26.8





#92
 Benzo(g,h,i)perylene
 Concen: 6.702 ng
 RT: 17.415 min Scan# 20
 Delta R.T. -0.024 min
 Lab File: BF136299.D
 Acq: 29 Nov 2023 23:30

Instrument :
 BNA_F
 ClientSampleId :
 L-3(195FT)(0-5)DL



Tgt Ion:	276	Resp:	71413
Ion Ratio	Lower	Upper	
276	100		
277	22.0	18.6	27.8
138	22.9	15.8	23.8

