

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112923\
 Data File : BF136274.D
 Acq On : 29 Nov 2023 10:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 29 11:22:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 28 03:04:20 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	40.196	-0.5	85	-0.01
3	Pyridine	40.000	40.449	-1.1	86	-0.01
4	n-Nitrosodimethylamine	40.000	36.732	8.2	78	-0.01
5 S	2-Fluorophenol	80.000	81.793	-2.2	87	0.00
6	Aniline	40.000	43.248	-8.1	88	0.00
7 S	Phenol-d6	80.000	80.682	-0.9	86	0.00
8	2-Chlorophenol	40.000	40.915	-2.3	87	0.00
9	Benzaldehyde	40.000	38.902	2.7	94	0.00
10 C	Phenol	40.000	40.661	-1.7	85	0.00
11	bis(2-Chloroethyl)ether	40.000	40.428	-1.1	86	0.00
12	1,3-Dichlorobenzene	40.000	41.297	-3.2	87	0.00
13 C	1,4-Dichlorobenzene	40.000	40.517	-1.3	86	0.00
14	1,2-Dichlorobenzene	40.000	40.614	-1.5	86	0.00
15	Benzyl Alcohol	40.000	40.966	-2.4	86	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.596	1.0	85	0.00
17	2-Methylphenol	40.000	40.672	-1.7	87	0.00
18	Hexachloroethane	40.000	40.607	-1.5	86	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.413	-1.0	85	0.00
20	3+4-Methylphenols	40.000	40.273	-0.7	85	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	40.800	-2.0	84	0.00
23 S	Nitrobenzene-d5	80.000	83.189	-4.0	86	0.00
24	Nitrobenzene	40.000	41.126	-2.8	85	0.00
25	Isophorone	40.000	41.269	-3.2	85	0.00
26 C	2-Nitrophenol	40.000	45.159	-12.9	91	0.00
27	2,4-Dimethylphenol	40.000	42.736	-6.8	88	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.856	-4.6	86	0.00
29 C	2,4-Dichlorophenol	40.000	42.860	-7.1	88	0.00
30	1,2,4-Trichlorobenzene	40.000	40.930	-2.3	85	0.00
31	Naphthalene	40.000	41.672	-4.2	86	0.00
32	Benzoic acid	40.000	48.422	-21.1	96	0.00
33	4-Chloroaniline	40.000	42.762	-6.9	87	0.00
34 C	Hexachlorobutadiene	40.000	40.654	-1.6	83	0.00
35	Caprolactam	40.000	43.490	-8.7	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.443	-6.1	87	0.00
37	2-Methylnaphthalene	40.000	41.534	-3.8	87	0.00
38	1-Methylnaphthalene	40.000	41.507	-3.8	86	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.935	-2.3	85	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.591	-4.0	80	0.00
42 S	2,4,6-Tribromophenol	80.000	85.363	-6.7	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.987	-5.0	86	0.00
44	2,4,5-Trichlorophenol	40.000	42.573	-6.4	89	0.00
45 S	2-Fluorobiphenyl	80.000	81.606	-2.0	86	0.00
46	1,1'-Biphenyl	40.000	41.649	-4.1	86	0.00
47	2-Chloronaphthalene	40.000	41.741	-4.4	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112923\
 Data File : BF136274.D
 Acq On : 29 Nov 2023 10:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 29 11:22:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 28 03:04:20 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.072	-5.2	85	0.00
49	Acenaphthylene	40.000	42.444	-6.1	87	0.00
50	Dimethylphthalate	40.000	41.626	-4.1	86	0.00
51	2,6-Dinitrotoluene	40.000	44.039	-10.1	91	0.00
52 C	Acenaphthene	40.000	41.705	-4.3	87	0.00
53	3-Nitroaniline	40.000	43.401	-8.5	91	0.00
54 P	2,4-Dinitrophenol	40.000	44.712	-11.8	91	0.00
55	Dibenzofuran	40.000	41.256	-3.1	85	0.00
56 P	4-Nitrophenol	40.000	44.659	-11.6	90	0.00
57	2,4-Dinitrotoluene	40.000	43.820	-9.6	89	0.00
58	Fluorene	40.000	41.144	-2.9	87	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.519	-6.3	88	0.00
60	Diethylphthalate	40.000	41.319	-3.3	87	0.00
61	4-Chlorophenyl-phenylether	40.000	41.129	-2.8	86	0.00
62	4-Nitroaniline	40.000	44.683	-11.7	91	0.00
63	Azobenzene	40.000	41.013	-2.5	86	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.256	-13.1	95	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.349	-5.9	89	0.00
67	4-Bromophenyl-phenylether	40.000	41.966	-4.9	87	0.00
68	Hexachlorobenzene	40.000	41.775	-4.4	87	0.00
69	Atrazine	40.000	40.924	-2.3	86	0.00
70 C	Pentachlorophenol	40.000	43.321	-8.3	88	0.00
71	Phenanthrene	40.000	41.513	-3.8	87	0.00
72	Anthracene	40.000	42.202	-5.5	87	0.00
73	Carbazole	40.000	42.543	-6.4	88	0.00
74	Di-n-butylphthalate	40.000	42.978	-7.4	88	0.00
75 C	Fluoranthene	40.000	41.410	-3.5	87	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	83	0.00
77	Benzydine	40.000	75.185	-88.0#	135	0.00
78	Pyrene	40.000	46.852	-17.1	88	0.00
79 S	Terphenyl-d14	80.000	92.131	-15.2	85	0.00
80	Butylbenzylphthalate	40.000	46.049	-15.1	84	0.00
81	Benzo(a)anthracene	40.000	41.052	-2.6	80	0.00
82	3,3'-Dichlorobenzidine	40.000	45.266	-13.2	87	0.00
83	Chrysene	40.000	41.468	-3.7	82	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.076	-2.7	79	0.00
85 c	Di-n-octyl phthalate	40.000	39.730	0.7	85	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.175	-2.9	91	0.00
88	Benzo(b)fluoranthene	40.000	38.784	3.0	89	0.00
89	Benzo(k)fluoranthene	40.000	38.421	3.9	83	0.00
90 C	Benzo(a)pyrene	40.000	40.949	-2.4	90	0.00
91	Dibenzo(a,h)anthracene	40.000	40.855	-2.1	91	0.00
92	Benzo(g,h,i)perylene	40.000	41.707	-4.3	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112923\
Data File : BF136274.D
Acq On : 29 Nov 2023 10:10
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 29 11:22:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 28 03:04:20 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0