

Data Path : Z:\HPCHEM1\BNA F\DATA\BF113016\
 Data File : BF091359.D
 Acq On : 30 Nov 2016 17:28
 Operator : UM/SJ
 Sample : H5767-03
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 112216-DSDC-E-U-B

Quant Time: Dec 01 00:39:59 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF112916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 13:33:46 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	183581	20.00	ng	-0.02
21) Naphthalene-d8	8.15	136	744993	20.00	ng	-0.02
38) Acenaphthene-d10	9.90	164	370111	20.00	ng	-0.02
63) Phenanthrene-d10	11.38	188	547740	20.00	ng	-0.02
75) Chrysene-d12	14.01	240	411670	20.00	ng	-0.03
86) Perylene-d12	15.44	264	375008	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	404654	36.01	ng	-0.02
7) Phenol-d6	6.48	99	350606	25.34	ng	-0.02
23) Nitrobenzene-d5	7.43	82	1285666	96.71	ng	-0.02
41) 2,4,6-Tribromophenol	10.69	330	252864	72.17	ng	-0.02
44) 2-Fluorobiphenyl	9.22	172	2084216	101.96	ng	-0.02
78) Terphenyl-d14	12.97	244	1584429	83.65	ng	-0.02
Target Compounds						
10) Phenol	6.49	94	46303	2.84	ng	92
49) Dimethylphthalate	9.61	163	120191	5.36	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF113016\
Data File : BF091359.D
Acq On : 30 Nov 2016 17:28
Operator : UM/SJ
Sample : H5767-03
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
112216-DSDC-E-U-B

Quant Time: Dec 01 00:39:59 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF112916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 13:33:46 2016
Response via : Initial Calibration

