


Data Path : Z:\HPCHEM1\BNA F\DATA\BF113017\
 Data File : BF101047.D
 Acq On : 1 Dec 2017 5:35
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF113017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Dec 01 09:47:02 2017



AutoFind: Scans 1620, 1621, 1622; Background Corrected with Scan 1616

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.6	9884	PASS
68	69	0.00	2	1.1	101	PASS
69	198	0.00	100	34.3	8791	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.3	11109	PASS
197	198	0.00	2	0.9	239	PASS
198	198	100	100	100.0	25629	PASS
199	198	5	9	7.7	1964	PASS
275	198	10	60	25.1	6426	PASS
365	198	1	100	2.5	635	PASS
441	443	0.01	100	86.3	3812	PASS
442	198	50	100	90.0	23064	PASS
443	442	15	24	19.2	4419	PASS