

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120115\
 Data File : BF083386.D
 Acq On : 1 Dec 2015 21:58
 Operator : UM/IZ
 Sample : G4593-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S14-15

Manual Integrations
 APPROVED

Mmdadoda
 12/2/2015 6:12:55 PM

Quant Time: Dec 02 04:03:00 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 01 01:17:16 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.54	152	160946	20.00	ng	0.00
21) Naphthalene-d8	7.84	136	613550	20.00	ng	0.00
38) Acenaphthene-d10	9.58	164	299448	20.00	ng	0.00
63) Phenanthrene-d10	11.13	188	540023	20.00	ng	0.00
75) Chrysene-d12	14.75	240	426903	20.00	ng	-0.01
86) Perylene-d12	16.81	264	324636	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.14	112	1382935	128.94	ng	0.01
7) Phenol-d6	6.22	99	2134958	145.43	ng	0.00
23) Nitrobenzene-d5	7.13	82	1326962	95.07	ng	0.00
41) 2,4,6-Tribromophenol	10.37	330	377731	125.66	ng	-0.01
44) 2-Fluorobiphenyl	8.92	172	2053640	97.73	ng	0.00
78) Terphenyl-d14	13.22	244	1793512	100.19	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.32	163	500430	20.56	ng	98
70) Phenanthrene	11.15	178	92227	2.89	ng	97
74) Fluoranthene	12.66	202	121081	3.66	ng	93
77) Pyrene	12.97	202	108205	3.63	ng	97
87) Benzo(b)fluoranthene	16.31	252	47398m	2.28	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120115\
Data File : BF083386.D
Acq On : 1 Dec 2015 21:58
Operator : UM/IZ
Sample : G4593-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S14-15

Manual Integrations
APPROVED

Mmdadoda
12/2/2015 6:12:55 PM

Quant Time: Dec 02 04:03:00 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 01 01:17:16 2015
Response via : Initial Calibration

