

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120115\
 Data File : BF083411.D
 Acq On : 2 Dec 2015 10:49
 Operator : UM/IZ
 Sample : G4567-05RE 2X
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 STOCKPILE-GRAB1

Manual Integrations
 APPROVED

Mmdadoda
 12/2/2015 6:13:09 PM

Quant Time: Dec 02 12:19:04 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 01 01:17:16 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.54	152	143848	20.00	ng	0.00
21) Naphthalene-d8	7.84	136	525310	20.00	ng	0.00
38) Acenaphthene-d10	9.58	164	224240	20.00	ng	0.00
63) Phenanthrene-d10	11.17	188	336680	20.00	ng	0.05
75) Chrysene-d12	14.95	240	261932	20.00	ng	0.18
86) Perylene-d12	16.95	264	190398	20.00	ng	0.13
System Monitoring Compounds						
5) 2-Fluorophenol	5.14	112	84739	8.84	ng	0.01
7) Phenol-d6	6.22	99	130258	9.93	ng	0.00
23) Nitrobenzene-d5	7.12	82	78027	6.53	ng	-0.01
41) 2,4,6-Tribromophenol	10.40	330	17949	7.97	ng	0.01
44) 2-Fluorobiphenyl	8.92	172	107968	6.86	ng	0.00
78) Terphenyl-d14	13.35	244	73440	6.69	ng	0.13
Target Compounds						
31) Naphthalene	7.86	128	101746	3.57	ng	Qvalue 98
37) 2-Methylnaphthalene	8.54	142	50062m	2.63	ng	
48) Acenaphthylene	9.45	152	165114	6.91	ng	# 95
51) Acenaphthene	9.62	154	53756	3.82	ng	94
70) Phenanthrene	11.20	178	692329	34.75	ng	# 74
77) Pyrene	13.11	202	275793	15.08	ng	# 56
80) Benzo(a)anthracene	14.93	228	111297m	6.94	ng	
82) Chrysene	14.98	228	139367m	8.99	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120115\
Data File : BF083411.D
Acq On : 2 Dec 2015 10:49
Operator : UM/IZ
Sample : G4567-05RE 2X
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
STOCKPILE-GRAB1

Manual Integrations
APPROVED

Mmdadoda
12/2/2015 6:13:09 PM

Quant Time: Dec 02 12:19:04 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 01 01:17:16 2015
Response via : Initial Calibration

