




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120116\
Data File : BF091370.D
Acq On : 1 Dec 2016 11:09
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 02 00:52:39 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 13:33:46 2016
Response via : Initial Calibration



TIC: BF091370.D

(69) Pentachlorophenol (C)




11.179min (-0.029) 21414.74ng

response 36753

Ion	Exp%	Act%
265.70	100	100
268.00	62.90	67.50
264.00	60.20	65.21
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120116\
Data File : BF091370.D
Acq On : 1 Dec 2016 11:09
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 02 00:52:39 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 13:33:46 2016
Response via : Initial Calibration



TIC: BF091370.D

(76) Benzidine

12.711min (-0.016) 0.00ng

response 274037

Ion	Exp%	Act%
184.00	100	100
185.00	16.40	14.24
183.00	11.10	12.02
0.00	0.00	0.00