

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF120117\
 Data File : BF101071.D
 Acq On : 1 Dec 2017 18:51
 Operator : SJ/JU
 Sample : I6622-07
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DBEW1A-2-4

Manual Integrations
 APPROVED

Sohil
 12/4/2017 11:56:46 AM

Quant Time: Dec 01 23:35:36 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 01 15:52:19 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	52304	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	199854	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	86314	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	139051	20.00	ng	0.00
75) Chrysene-d12	14.02	240	121802	20.00	ng	0.00
86) Perylene-d12	15.49	264	106846	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	247519	78.20	ng	0.00
7) Phenol-d6	6.48	99	308055	80.16	ng	0.00
23) Nitrobenzene-d5	7.41	82	186730	55.74	ng	0.00
41) 2,4,6-Tribromophenol	10.67	330	53568	51.66	ng	0.00
44) 2-Fluorobiphenyl	9.20	172	296520	47.00	ng	0.00
78) Terphenyl-d14	12.96	244	203192	36.49	ng	0.00
Target Compounds						
10) Phenol	6.49	94	9095	2.128	ng	95
49) Dimethylphthalate	9.59	163	25781	3.704	ng	# 99
70) Phenanthrene	11.40	178	15421	2.014	ng	98
82) Chrysene	14.04	228	18799m	2.632	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF120117\
Data File : BF101071.D
Acq On : 1 Dec 2017 18:51
Operator : SJ/JU
Sample : I6622-07
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DBEW1A-2-4

Manual Integrations
APPROVED

Sohil
12/4/2017 11:56:46 AM

Quant Time: Dec 01 23:35:36 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 15:52:19 2017
Response via : Initial Calibration

