



Data Path : Z:\HPCHEM1\BNA F\DATA\BF120117\
Data File : BF101055.D
Acq On : 1 Dec 2017 11:28
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 02 04:01:40 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 15:52:19 2017
Response via : Initial Calibration



TIC: BF101055.D

(76) Benzidine



12.710min (+0.000) 0.00ng

response 234503

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	13.28
183.00	11.60	12.17
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120117\
Data File : BF101055.D
Acq On : 1 Dec 2017 11:28
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 02 04:01:40 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 01 15:52:19 2017
Response via : Initial Calibration



TIC: BF101055.D

(69) Pentachlorophenol (C)

11.175min (-0.006) 0.00ng

response 28135

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	64.53
264.00	61.90	64.61
0.00	0.00	0.00