


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF120117\
 Data File : BF101082.D
 Acq On : 2 Dec 2017 00:40
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF113017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Dec 01 15:52:19 2017



AutoFind: Scans 1619, 1620, 1621; Background Corrected with Scan 1615

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.5	12795	PASS
68	69	0.00	2	1.1	118	PASS
69	198	0.00	100	32.8	11188	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.2	14748	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	34106	PASS
199	198	5	9	7.4	2540	PASS
275	198	10	60	25.2	8611	PASS
365	198	1	100	2.6	884	PASS
441	443	0.01	100	89.2	5014	PASS
442	198	50	100	88.2	30087	PASS
443	442	15	24	18.7	5620	PASS