


Data Path : Z:\HPCHEM1\BNA F\Data\BF120215\  
 Data File : BF083412.D  
 Acq On : 2 Dec 2015 12:02  
 Operator : UM/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF113015.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Dec 01 01:17:16 2015



AutoFind: Scans 855, 856, 857; Background Corrected with Scan 851

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 46.1      | 49038   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.2       | 540     | PASS             |
| 69          | 198          | 0.00         | 100          | 42.8      | 45512   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.3       | 114     | PASS             |
| 127         | 198          | 10           | 80           | 48.0      | 51071   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.5       | 555     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 106448  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 7290    | PASS             |
| 275         | 198          | 10           | 60           | 23.2      | 24746   | PASS             |
| 365         | 198          | 1            | 100          | 2.2       | 2394    | PASS             |
| 441         | 442          | 0.01         | 24           | 16.3      | 12662   | PASS             |
| 442         | 198          | 50           | 100          | 72.9      | 77555   | PASS             |
| 443         | 442          | 15           | 24           | 19.4      | 15072   | PASS             |