




Data Path : Z:\HPCHEM1\BNA\_F\Data\BF120216\  
Data File : BF091384.D  
Acq On : 2 Dec 2016 8:19  
Operator : UM/SJ  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 03 05:56:16 2016  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF112916.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Nov 29 13:33:46 2016  
Response via : Initial Calibration



TIC: BF091384.D

(69) Pentachlorophenol (C)




11.179min (-0.029) 13092.30ng

response 59173

| Ion    | Exp%  | Act%  |
|--------|-------|-------|
| 265.70 | 100   | 100   |
| 268.00 | 62.90 | 65.73 |
| 264.00 | 60.20 | 62.39 |
| 0.00   | 0.00  | 0.00  |

Data Path : Z:\HPCHEM1\BNA\_F\Data\BF120216\  
Data File : BF091384.D  
Acq On : 2 Dec 2016 8:19  
Operator : UM/SJ  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 03 05:56:16 2016  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF112916.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Nov 29 13:33:46 2016  
Response via : Initial Calibration



TIC: BF091384.D

(76) Benzidine

12.711min (-0.016) 2386.10ng

response 303445

| Ion    | Exp%  | Act%  |
|--------|-------|-------|
| 184.00 | 100   | 100   |
| 185.00 | 16.40 | 14.60 |
| 183.00 | 11.10 | 11.42 |
| 0.00   | 0.00  | 0.00  |