

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120222\
 Data File : BF131497.D
 Acq On : 02 Dec 2022 15:53
 Operator : CG\JU
 Sample : N5848-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PE-2

Quant Time: Dec 05 00:02:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 05 00:00:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.928	152	62003	20.000	ng	0.00
21) Naphthalene-d8	8.216	136	232144	20.000	ng	0.00
39) Acenaphthene-d10	9.981	164	188545	20.000	ng	0.00
64) Phenanthrene-d10	11.475	188	335615	20.000	ng	0.00
76) Chrysene-d12	14.127	240	223391	20.000	ng	#-0.01
86) Perylene-d12	15.657	264	186148	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.546	112	384286	117.959	ng	0.01
7) Phenol-d6	6.546	99	485480	116.798	ng	-0.01
23) Nitrobenzene-d5	7.493	82	326395	70.875	ng	0.00
42) 2,4,6-Tribromophenol	10.775	330	209077	131.806	ng	0.00
45) 2-Fluorobiphenyl	9.298	172	852856	65.807	ng	0.00
79) Terphenyl-d14	13.075	244	917002	67.126	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120222\
Data File : BF131497.D
Acq On : 02 Dec 2022 15:53
Operator : CG\JU
Sample : N5848-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PE-2

Quant Time: Dec 05 00:02:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 05 00:00:18 2022
Response via : Initial Calibration

