

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120222\
 Data File : BF131498.D
 Acq On : 02 Dec 2022 16:25
 Operator : CG\JU
 Sample : N5848-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PE-3

Quant Time: Dec 05 00:03:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 05 00:00:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.928	152	69618	20.000	ng	0.00
21) Naphthalene-d8	8.216	136	296662	20.000	ng	# 0.00
39) Acenaphthene-d10	9.980	164	193333	20.000	ng	0.00
64) Phenanthrene-d10	11.474	188	350230	20.000	ng	0.00
76) Chrysene-d12	14.127	240	229813	20.000	ng	#-0.01
86) Perylene-d12	15.657	264	193908	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.545	112	431136	117.864	ng	0.01
7) Phenol-d6	6.545	99	558045	119.571	ng	-0.01
23) Nitrobenzene-d5	7.492	82	358697	60.950	ng	0.00
42) 2,4,6-Tribromophenol	10.775	330	199278	122.518	ng	0.00
45) 2-Fluorobiphenyl	9.298	172	808298	60.824	ng	0.00
79) Terphenyl-d14	13.074	244	857734	61.033	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120222\
Data File : BF131498.D
Acq On : 02 Dec 2022 16:25
Operator : CG\JU
Sample : N5848-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PE-3

Quant Time: Dec 05 00:03:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 05 00:00:18 2022
Response via : Initial Calibration

