

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120222\
 Data File : BF131501.D
 Acq On : 02 Dec 2022 18:02
 Operator : CG\JU
 Sample : N5847-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-40-GW-01

Quant Time: Dec 05 00:03:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 05 00:00:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.928	152	71866	20.000	ng	0.00
21) Naphthalene-d8	8.216	136	270029	20.000	ng	# 0.00
39) Acenaphthene-d10	9.981	164	157991	20.000	ng	0.00
64) Phenanthrene-d10	11.475	188	251583	20.000	ng	0.00
76) Chrysene-d12	14.127	240	209658	20.000	ng	#-0.01
86) Perylene-d12	15.657	264	191440	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.546	112	333951	88.440	ng	0.01
7) Phenol-d6	6.546	99	328906	68.269	ng	-0.01
23) Nitrobenzene-d5	7.493	82	493331	92.094	ng	0.00
42) 2,4,6-Tribromophenol	10.775	330	224973	169.256	ng	0.00
45) 2-Fluorobiphenyl	9.298	172	971205	89.431	ng	0.00
79) Terphenyl-d14	13.075	244	927128	72.313	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120222\
Data File : BF131501.D
Acq On : 02 Dec 2022 18:02
Operator : CG\JU
Sample : N5847-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-40-GW-01

Quant Time: Dec 05 00:03:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 05 00:00:18 2022
Response via : Initial Calibration

