


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120315\
 Data File : BF083458.D
 Acq On : 3 Dec 2015 13:56
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF113015.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 01 01:17:16 2015



AutoFind: Scans 853, 854, 855; Background Corrected with Scan 849

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.2	31939	PASS
68	69	0.00	2	1.6	464	PASS
69	198	0.00	100	43.9	29717	PASS
70	69	0.00	2	0.8	251	PASS
127	198	10	80	49.4	33411	PASS
197	198	0.00	2	0.4	247	PASS
198	198	100	100	100.0	67688	PASS
199	198	5	9	6.6	4469	PASS
275	198	10	60	22.9	15506	PASS
365	198	1	100	2.4	1622	PASS
441	442	0.01	24	15.4	8007	PASS
442	198	50	100	76.6	51856	PASS
443	442	15	24	18.3	9486	PASS