





Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF120415.M
Calibration Table Last Updated: Sat Dec 05 04:12:46 2015

Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $2.96 \times 10^{-1} \times \text{Amt} - 1.54 \times 10^{-1}$
Coef of Det (r^2) = 0.992 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF120415.M
Calibration Table Last Updated: Sat Dec 05 04:12:46 2015

2,4-Dinitrophenol


Response Ratio



Resp Ratio = $1.89 \times 10^{-1} \times \text{Amt} - 7.15 \times 10^{-2}$
Coef of Det (r^2) = 0.992 Curve Fit: Linear


Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF120415.M
Calibration Table Last Updated: Sat Dec 05 04:12:46 2015

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.46 \times 10^{-1} \times \text{Amt} - 2.53 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF120415.M
Calibration Table Last Updated: Sat Dec 05 04:12:46 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF120415.M
Calibration Table Last Updated: Sat Dec 05 04:12:46 2015