

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120417\
 Data File : BF101106.D
 Acq On : 4 Dec 2017 18:56
 Operator : SJ/JU
 Sample : I6697-04
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-04A

Manual Integrations
 APPROVED

Sohil
 12/5/2017 3:24:43 PM

Quant Time: Dec 05 01:39:21 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 04 13:07:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	59907	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	226197	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	100419	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	165527	20.00	ng	0.00
75) Chrysene-d12	14.02	240	130516	20.00	ng	0.00
86) Perylene-d12	15.50	264	114369	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	274229	75.64	ng	0.01
7) Phenol-d6	6.49	99	326506	74.18	ng	0.00
23) Nitrobenzene-d5	7.42	82	198208	52.27	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	72803	60.35	ng	0.00
44) 2-Fluorobiphenyl	9.20	172	421086	57.37	ng	0.00
78) Terphenyl-d14	12.96	244	285835	47.90	ng	0.00
Target Compounds						
10) Phenol	6.49	94	11426	2.33	ng	Qvalue 83
49) Dimethylphthalate	9.60	163	45081	5.57	ng	# 98
70) Phenanthrene	11.40	178	19255	2.11	ng	97
77) Pyrene	12.82	202	18558	2.06	ng	99
82) Chrysene	14.04	228	47541	6.21	ng	# 85
87) Benzo(b)fluoranthene	15.06	252	15573m	2.27	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120417\
Data File : BF101106.D
Acq On : 4 Dec 2017 18:56
Operator : SJ/JU
Sample : I6697-04
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-04A

Manual Integrations
APPROVED

Sohil
12/5/2017 3:24:43 PM

Quant Time: Dec 05 01:39:21 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 04 13:07:49 2017
Response via : Initial Calibration

