

Data Path : Z:\HPCHEM1\BNA F\Data\BF120417\
 Data File : BF101086.D
 Acq On : 4 Dec 2017 9:57
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 04 15:16:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 04 13:07:49 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.500	0.491	1.8	91	0.01
3	Pyridine	1.297	1.214	6.4	80	0.00
4	n-Nitrosodimethylamine	0.634	0.687	-8.4	99	0.00
5 S	2-Fluorophenol	1.210	1.211	-0.1	93	0.00
6	Aniline	1.911	1.949	-2.0	95	0.00
7 S	Phenol-d6	1.469	1.505	-2.5	96	0.00
8	2-Chlorophenol	1.320	1.344	-1.8	96	0.00
9	Benzaldehyde	0.899	0.860	4.3	93	0.00
10 C	Phenol	1.634	1.669	-2.1	96	0.00
11	bis(2-Chloroethyl)ether	1.226	1.238	-1.0	94	0.00
12	1,3-Dichlorobenzene	1.537	1.568	-2.0	95	0.00
13 C	1,4-Dichlorobenzene	1.591	1.613	-1.4	95	0.00
14	1,2-Dichlorobenzene	1.484	1.521	-2.5	98	0.00
15	Benzyl Alcohol	1.135	1.199	-5.6	97	0.00
16	2,2'-oxybis(1-Chloropropane	1.666	1.713	-2.8	94	0.00
17	2-Methylphenol	1.066	1.077	-1.0	95	0.00
18	Hexachloroethane	0.511	0.520	-1.8	96	0.00
19 P	n-Nitroso-di-n-propylamine	0.990	1.024	-3.4	99	0.00
20	3+4-Methylphenols	1.390	1.440	-3.6	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.483	0.486	-0.6	97	0.00
23 S	Nitrobenzene-d5	0.335	0.334	0.3	95	0.00
24	Nitrobenzene	0.329	0.329	0.0	97	0.00
25	Isophorone	0.573	0.571	0.3	96	0.00
26 C	2-Nitrophenol	0.180	0.183	-1.7	97	0.00
27	2,4-Dimethylphenol	0.261	0.262	-0.4	96	0.00
28	bis(2-Chloroethoxy)methane	0.385	0.383	0.5	96	0.00
29 C	2,4-Dichlorophenol	0.284	0.285	-0.4	94	0.00
30	1,2,4-Trichlorobenzene	0.312	0.316	-1.3	98	0.00
31	Naphthalene	0.986	0.979	0.7	96	0.00
32	Benzoic acid	0.203	0.204	-0.5	101	0.00
33	4-Chloroaniline	0.396	0.399	-0.8	95	0.00
34 C	Hexachlorobutadiene	0.192	0.191	0.5	96	0.00
35	Caprolactam	0.089	0.085	4.5	89	0.00
36 C	4-Chloro-3-methylphenol	0.294	0.296	-0.7	97	0.00
37	2-Methylnaphthalene	0.670	0.666	0.6	95	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
39	1,2,4,5-Tetrachlorobenzene	0.727	0.751	-3.3	98	0.00
40 P	Hexachlorocyclopentadiene	0.215	0.204	5.1	87	0.00
41 S	2,4,6-Tribromophenol	0.240	0.238	0.8	92	0.00
42 C	2,4,6-Trichlorophenol	0.426	0.437	-2.6	94	0.00
43	2,4,5-Trichlorophenol	0.455	0.473	-4.0	95	0.00
44 S	2-Fluorobiphenyl	1.462	1.491	-2.0	96	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF120417\
 Data File : BF101086.D
 Acq On : 4 Dec 2017 9:57
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 04 15:16:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 04 13:07:49 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.832	1.850	-1.0	96	0.00
46	2-Chloronaphthalene	1.301	1.308	-0.5	94	0.00
47	2-Nitroaniline	0.385	0.382	0.8	92	0.00
48	Acenaphthylene	2.139	2.129	0.5	93	0.00
49	Dimethylphthalate	1.613	1.588	1.5	93	0.00
50	2,6-Dinitrotoluene	0.340	0.339	0.3	94	0.00
51 C	Acenaphthene	1.281	1.269	0.9	95	0.00
52	3-Nitroaniline	0.382	0.375	1.8	91	0.00
53 P	2,4-Dinitrophenol	0.155	0.149	3.9	86	0.00
54	Dibenzofuran	1.845	1.845	0.0	94	0.00
55 P	4-Nitrophenol	0.258	0.246	4.7	86	0.00
56	2,4-Dinitrotoluene	0.455	0.444	2.4	90	0.00
57	Fluorene	1.500	1.480	1.3	93	0.00
58	2,3,4,6-Tetrachlorophenol	0.365	0.372	-1.9	94	0.00
59	Diethylphthalate	1.591	1.576	0.9	92	0.00
60	4-Chlorophenyl-phenylether	0.741	0.747	-0.8	94	0.00
61	4-Nitroaniline	0.397	0.373	6.0	88	0.00
62	Azobenzene	1.350	1.311	2.9	91	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
64	4,6-Dinitro-2-methylphenol	0.134	0.131	2.2	86	0.00
65 c	n-Nitrosodiphenylamine	0.706	0.701	0.7	89	0.00
66	4-Bromophenyl-phenylether	0.224	0.229	-2.2	91	0.00
67	Hexachlorobenzene	0.240	0.244	-1.7	92	0.00
68	Atrazine	0.216	0.215	0.5	90	0.00
69 C	Pentachlorophenol	0.132	0.132	0.0	85	0.00
70	Phenanthrene	1.101	1.088	1.2	89	0.00
71	Anthracene	1.115	1.097	1.6	88	0.00
72	Carbazole	1.018	0.988	2.9	88	0.00
73	Di-n-butylphthalate	1.277	1.248	2.3	87	0.00
74 C	Fluoranthene	1.221	1.153	5.6	86	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
76	Benzidine	0.650	0.683	-5.1	92	0.00
77	Pyrene	1.380	1.370	0.7	87	0.00
78 S	Terphenyl-d14	0.914	0.914	0.0	89	0.00
79	Butylbenzylphthalate	0.608	0.605	0.5	85	0.00
80	Benzo(a)anthracene	1.263	1.245	1.4	87	0.00
81	3,3'-Dichlorobenzidine	0.437	0.426	2.5	84	0.00
82	Chrysene	1.173	1.156	1.4	87	0.00
83	Bis(2-ethylhexyl)phthalate	0.868	0.869	-0.1	86	0.00
84 c	Di-n-octyl phthalate	1.444	1.437	0.5	85	0.00
85	Indeno(1,2,3-cd)pyrene	1.043	1.045	-0.2	89	0.00
86 I	Perylene-d12	1.000	1.000	0.0	88	0.00
87	Benzo(b)fluoranthene	1.198	1.205	-0.6	89	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF120417\
Data File : BF101086.D
Acq On : 4 Dec 2017 9:57
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 04 15:16:57 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 04 13:07:49 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.180	1.132	4.1	82	0.00
89 C	Benzo(a)pyrene	1.089	1.070	1.7	86	0.00
90	Dibenzo(a,h)anthracene	0.960	0.948	1.3	88	0.00
91	Benzo(a,h,i)perylene	0.905	0.913	-0.9	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0