

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120516\
 Data File : BF091424.D
 Acq On : 5 Dec 2016 18:09
 Operator : UM/SJ
 Sample : H5838-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-9

Manual Integrations
 APPROVED

sohil
 12/6/2016 3:02:03 PM

Quant Time: Dec 06 00:30:34 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF112916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 06 00:16:00 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	147224	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	610782	20.00	ng	-0.01
38) Acenaphthene-d10	9.89	164	301879	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	538702	20.00	ng	0.00
75) Chrysene-d12	14.00	240	479183	20.00	ng	0.00
86) Perylene-d12	15.42	264	356716	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	615893	68.34	ng	0.00
7) Phenol-d6	6.47	99	457961	41.28	ng	-0.01
23) Nitrobenzene-d5	7.41	82	1169431	107.30	ng	-0.01
41) 2,4,6-Tribromophenol	10.68	330	433870	151.81	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	1781545	106.86	ng	0.00
78) Terphenyl-d14	12.96	244	2080038	94.35	ng	0.00
Target Compounds						
48) Acenaphthylene	9.75	152	93534	3.68	ng	# 88
51) Acenaphthene	9.92	154	559861	32.62	ng	98
54) Dibenzofuran	10.09	168	168442	7.33	ng	# 87
57) Fluorene	10.44	166	542649m	30.75	ng	
71) Anthracene	11.44	178	70862	2.55	ng	97
74) Fluoranthene	12.57	202	81991	3.09	ng	98
77) Pyrene	12.80	202	107561	2.96	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120516\
Data File : BF091424.D
Acq On : 5 Dec 2016 18:09
Operator : UM/SJ
Sample : H5838-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-9

Manual Integrations
APPROVED

sohil
12/6/2016 3:02:03 PM

Quant Time: Dec 06 00:30:34 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF112916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 06 00:16:00 2016
Response via : Initial Calibration

