

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120518\
 Data File : BF111117.D
 Acq On : 5 Dec 2018 19:54
 Operator : JU/SJ
 Sample : J5764-25 2X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-02-113018-C

Quant Time: Dec 06 01:24:28 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 05 17:30:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	570507	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	2195489	20.00	ng	0.00
39) Acenaphthene-d10	9.84	164	964030	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	1397909	20.00	ng	0.00
76) Chrysene-d12	13.96	240	1063847	20.00	ng	0.00
87) Perylene-d12	15.39	264	753181	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	1703562	46.60	ng	0.00
7) Phenol-d6	6.43	99	2183633	46.37	ng	0.00
23) Nitrobenzene-d5	7.36	82	1284801	34.65	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	319981	41.76	ng	0.00
45) 2-Fluorobiphenyl	9.16	172	2214463	33.87	ng	0.00
79) Terphenyl-d14	12.90	244	1526656	29.39	ng	0.00
Target Compounds						
10) Phenol	6.44	94	115663	2.252	ng	93
50) Dimethylphthalate	9.55	163	164778	2.456	ng	98
71) Phenanthrene	11.34	178	136234	2.026	ng	95
75) Fluoranthene	12.53	202	169930	2.583	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120518\
Data File : BF111117.D
Acq On : 5 Dec 2018 19:54
Operator : JU/SJ
Sample : J5764-25 2X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-02-113018-C

Quant Time: Dec 06 01:24:28 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 05 17:30:55 2018
Response via : Initial Calibration

