

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120520\
 Data File : BF122564.D
 Acq On : 5 Dec 2020 18:18
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF120520

Quant Time: Dec 07 12:12:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	40.439	-1.1	90	-0.02
3	Pyridine	40.000	41.093	-2.7	89	-0.02
4	n-Nitrosodimethylamine	40.000	40.719	-1.8	88	-0.02
5 S	2-Fluorophenol	80.000	80.977	-1.2	90	0.00
6	Aniline	40.000	40.774	-1.9	89	0.00
7 S	Phenol-d6	80.000	80.357	-0.4	89	0.00
8	2-Chlorophenol	40.000	40.739	-1.8	90	0.00
9	Benzaldehyde	40.000	36.168	9.6	91	0.00
10 C	Phenol	40.000	40.572	-1.4	91	0.00
11	bis(2-Chloroethyl)ether	40.000	41.191	-3.0	91	0.00
12	1,3-Dichlorobenzene	40.000	40.666	-1.7	90	0.00
13 C	1,4-Dichlorobenzene	40.000	40.726	-1.8	91	0.00
14	1,2-Dichlorobenzene	40.000	39.064	2.3	91	0.00
15	Benzyl Alcohol	40.000	41.406	-3.5	89	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.861	-2.2	90	0.00
17	2-Methylphenol	40.000	40.584	-1.5	87	0.00
18	Hexachloroethane	40.000	41.347	-3.4	91	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.348	-0.9	89	0.00
20	3+4-Methylphenols	40.000	39.405	1.5	89	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	89	0.00
22	Acetophenone	40.000	39.881	0.3	89	0.00
23 S	Nitrobenzene-d5	80.000	81.319	-1.6	89	0.00
24	Nitrobenzene	40.000	40.731	-1.8	90	0.00
25	Isophorone	40.000	40.697	-1.7	89	0.00
26 C	2-Nitrophenol	40.000	42.282	-5.7	90	0.00
27	2,4-Dimethylphenol	40.000	41.230	-3.1	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.667	-1.7	89	0.00
29 C	2,4-Dichlorophenol	40.000	41.042	-2.6	89	0.00
30	1,2,4-Trichlorobenzene	40.000	40.590	-1.5	90	0.00
31	Naphthalene	40.000	40.291	-0.7	89	0.00
32	Benzoic acid	40.000	42.409	-6.0	86	0.00
33	4-Chloroaniline	40.000	40.962	-2.4	89	0.00
34 C	Hexachlorobutadiene	40.000	40.859	-2.1	90	0.00
35	Caprolactam	40.000	41.552	-3.9	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.938	-2.3	89	0.00
37	2-Methylnaphthalene	40.000	40.375	-0.9	89	0.00
38	1-Methylnaphthalene	40.000	39.967	0.1	89	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	89	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.537	-1.3	89	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.939	-4.8	86	0.00
42 S	2,4,6-Tribromophenol	80.000	81.558	-1.9	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.011	-2.5	88	0.00
44	2,4,5-Trichlorophenol	40.000	41.444	-3.6	90	0.00
45 S	2-Fluorobiphenyl	80.000	75.445	5.7	89	0.00
46	1,1'-Biphenyl	40.000	40.514	-1.3	90	0.00
47	2-Chloronaphthalene	40.000	40.162	-0.4	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120520\
 Data File : BF122564.D
 Acq On : 5 Dec 2020 18:18
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF120520

Quant Time: Dec 07 12:12:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.804	-4.5	89	0.00
49	Acenaphthylene	40.000	40.529	-1.3	90	0.00
50	Dimethylphthalate	40.000	40.495	-1.2	89	0.00
51	2,6-Dinitrotoluene	40.000	42.425	-6.1	90	0.00
52 C	Acenaphthene	40.000	41.103	-2.8	90	0.00
53	3-Nitroaniline	40.000	41.459	-3.6	88	0.00
54 P	2,4-Dinitrophenol	40.000	43.958	-9.9	90	0.00
55	Dibenzofuran	40.000	40.467	-1.2	90	0.00
56 P	4-Nitrophenol	40.000	42.487	-6.2	88	0.00
57	2,4-Dinitrotoluene	40.000	42.257	-5.6	89	0.00
58	Fluorene	40.000	38.289	4.3	89	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.256	-3.1	88	0.00
60	Diethylphthalate	40.000	40.581	-1.5	89	0.00
61	4-Chlorophenyl-phenylether	40.000	39.916	0.2	90	0.00
62	4-Nitroaniline	40.000	41.361	-3.4	89	0.00
63	Azobenzene	40.000	40.747	-1.9	89	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.262	-8.2	91	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.011	-0.0	89	0.00
67	4-Bromophenyl-phenylether	40.000	40.337	-0.8	89	0.00
68	Hexachlorobenzene	40.000	40.020	-0.1	89	0.00
69	Atrazine	40.000	39.921	0.2	89	0.00
70 C	Pentachlorophenol	40.000	42.723	-6.8	87	0.00
71	Phenanthrene	40.000	40.043	-0.1	91	0.00
72	Anthracene	40.000	40.033	-0.1	90	0.00
73	Carbazole	40.000	39.662	0.8	89	0.00
74	Di-n-butylphthalate	40.000	40.566	-1.4	92	0.00
75 C	Fluoranthene	40.000	39.935	0.2	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
77	Benzydine	40.000	43.795	-9.5	86	0.00
78	Pyrene	40.000	40.635	-1.6	89	0.00
79 S	Terphenyl-d14	80.000	78.981	1.3	91	0.00
80	Butylbenzylphthalate	40.000	40.826	-2.1	89	0.00
81	Benzo(a)anthracene	40.000	39.978	0.1	90	0.00
82	3,3'-Dichlorobenzidine	40.000	40.386	-1.0	89	0.00
83	Chrysene	40.000	39.841	0.4	89	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.389	-1.0	91	0.00
85 c	Di-n-octyl phthalate	40.000	40.747	-1.9	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.121	-0.3	90	0.00
88	Benzo(b)fluoranthene	40.000	39.930	0.2	84	0.00
89	Benzo(k)fluoranthene	40.000	42.805	-7.0	95	0.00
90 C	Benzo(a)pyrene	40.000	40.804	-2.0	89	0.00
91	Dibenzo(a,h)anthracene	40.000	39.959	0.1	90	0.00
92	Benzo(g,h,i)perylene	40.000	40.251	-0.6	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120520\
Data File : BF122564.D
Acq On : 5 Dec 2020 18:18
Operator : JU/CG
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF120520

Quant Time: Dec 07 12:12:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 12:08:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0