

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120615\
 Data File : BF083553.D
 Acq On : 7 Dec 2015 10:18
 Operator : UM/IZ
 Sample : G4617-06RE
 Misc :
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-11RE

Manual Integrations
 APPROVED

Mmdadoda
 12/7/2015 6:33:45 PM

Quant Time: Dec 07 11:40:08 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 05 04:12:46 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)	
1) 1,4-Dichlorobenzene-d4	5.73	152	159671	20.00	ng	-0.03	
21) Naphthalene-d8	7.63	136	603627	20.00	ng	-0.05	
38) Acenaphthene-d10	10.43	164	269019	20.00	ng	-0.03	
63) Phenanthrene-d10	12.82	188	412527	20.00	ng	-0.03	
75) Chrysene-d12	17.01	240	343074	20.00	ng	-0.03	
86) Perylene-d12	18.65	264	267903	20.00	ng	-0.02	
System Monitoring Compounds							
5) 2-Fluorophenol	4.02	112	1103419	104.84	ng	-0.01	
7) Phenol-d6	5.30	99	1440847	99.36	ng	-0.02	
23) Nitrobenzene-d5	6.56	82	924064	71.56	ng	-0.05	
41) 2,4,6-Tribromophenol	11.71	330	211797	85.35	ng	-0.05	
44) 2-Fluorobiphenyl	9.38	172	1385184	69.84	ng	-0.05	
78) Terphenyl-d14	15.46	244	975422	63.27	ng	-0.02	
Target Compounds							
49) Dimethylphthalate	10.06	163	271547	12.82	ng		98
70) Phenanthrene	12.85	178	72341	3.01	ng		98
74) Fluoranthene	14.79	202	83742	3.52	ng		99
77) Pyrene	15.14	202	73762	2.87	ng		98
87) Benzo(b)fluoranthene	18.28	252	36700m	2.44	ng		

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120615\
 Data File : BF083553.D
 Acq On : 7 Dec 2015 10:18
 Operator : UM/IZ
 Sample : G4617-06RE
 Misc :
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampled :
 SB-11RE

Manual Integrations
 APPROVED
 Mmdadoda
 12/7/2015 6:33:45 PM

Quant Time: Dec 07 11:40:08 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 05 04:12:46 2015
 Response via : Initial Calibration

