

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120615\
 Data File : BF083537.D
 Acq On : 7 Dec 2015 1:47
 Operator : UM/IZ
 Sample : G4614-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S34-15.5

Manual Integrations
 APPROVED

Mmdadoda
 12/7/2015 6:33:23 PM

Quant Time: Dec 07 06:31:35 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 05 04:12:46 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.73	152	190569	20.00	ng	-0.03
21) Naphthalene-d8	7.63	136	728763	20.00	ng	-0.05
38) Acenaphthene-d10	10.42	164	346712	20.00	ng	-0.05
63) Phenanthrene-d10	12.82	188	581830	20.00	ng	-0.03
75) Chrysene-d12	17.03	240	372052	20.00	ng	-0.02
86) Perylene-d12	18.66	264	380648	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	4.02	112	1691256	134.64	ng	-0.01
7) Phenol-d6	5.30	99	2213213	127.88	ng	-0.02
23) Nitrobenzene-d5	6.57	82	1490528	95.61	ng	-0.03
41) 2,4,6-Tribromophenol	11.72	330	412814	129.07	ng	-0.03
44) 2-Fluorobiphenyl	9.38	172	2007350	78.53	ng	-0.05
78) Terphenyl-d14	15.47	244	1358305	81.24	ng	-0.01
Target Compounds						Qvalue
20) 3+4-Methylphenols	6.41	107	262671	17.54	ng	94
31) Naphthalene	7.66	128	363766	9.38	ng	99
37) 2-Methylnaphthalene	8.77	142	108776	4.53	ng	98
49) Dimethylphthalate	10.06	163	512418	18.78	ng	98
51) Acenaphthene	10.48	154	363344	16.61	ng	99
54) Dibenzofuran	10.76	168	308635	10.18	ng	97
57) Fluorene	11.31	166	416562	15.80	ng	97
70) Phenanthrene	12.88	178	3561632	105.17	ng	96
71) Anthracene	12.95	178	984227	28.50	ng	98
72) Carbazole	13.24	167	376555	12.70	ng	98
74) Fluoranthene	14.83	202	3571034	106.39	ng	95
77) Pyrene	15.17	202	3084397	110.82	ng	# 89
80) Benzo(a)anthracene	17.01	228	1425552	65.49	ng	# 96
82) Chrysene	17.07	228	1337701	61.69	ng	96
85) Indeno(1,2,3-cd)pyrene	19.89	276	637388	34.29	ng	# 100
87) Benzo(b)fluoranthene	18.28	252	1288641m	60.28	ng	
88) Benzo(k)fluoranthene	18.29	252	545730m	24.92	ng	
89) Benzo(a)pyrene	18.60	252	1015287	48.83	ng	98
90) Dibenzo(a,h)anthracene	19.91	278	175394	8.77	ng	98
91) Benzo(a,h,i)perylene	20.26	276	613461	30.43	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA F\DATA\BF120615\
Data File : BF083537.D
Acq On : 7 Dec 2015 1:47
Operator : UM/IZ
Sample : G4614-01
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
S34-15.5

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM

Quant Time: Dec 07 06:31:35 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 05 04:12:46 2015
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 5.73 min Scan# 354
Delta R.T. -0.03 min
Lab File: BF083537.D
Acq: 7 Dec 2015 1:47

Instrument :
BNA_F
ClientSampled :
S34-15.5

Tot Ion	Ratio	Lower	Upper
152	100		
150	155.9	126.5	189.7
115	64.5	52.3	78.5

Manual Integrations
APPROVED




Mmdadoda
12/7/2015 6:33:23 PM






#5

2-Fluorophenol
Concen: 134.64 ng
RT: 4.02 min Scan# 204
Delta R.T. -0.01 min
Lab File: BF083537.D
Acq: 7 Dec 2015 1:47

Tgt Ion	Ratio	Lower	Upper
112	100		
64	61.2	50.5	75.7
63	31.2	25.8	38.6





#7

Phenol-d6

Concen: 127.88 ng

RT: 5.30 min Scan# 316

Delta R.T. -0.02 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tot Ion: 99 Resp: 2213213

Ion Ratio Lower Upper

99 100

42 21.8 17.3 25.9

71 34.5 26.6 40.0

Instrument :





BNA_F

ClientSampleId :

S34-15.5

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



#20

3+4-Methylphenols

Concen: 17.54 ng

RT: 6.41 min Scan# 413

Delta R.T. -0.02 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tgt Ion: 107 Resp: 262671


Ion Ratio Lower Upper


107 100

108 81.6 65.2 105.2

77 29.7 15.2 55.2

79 22.1 6.0 46.0





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 7.63 min Scan# 520

Delta R.T. -0.05 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F


ClientSampleId :

S34-15.5

Tot Ion	Ion	Ratio	Resp	Lower	Upper
136	100		728763		
137	11.0	8.7	13.1		
54	11.3	7.8	11.6		
68	8.0	5.7	8.5		

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#23

Nitrobenzene-d5

Concen: 95.61 ng


RT: 6.57 min Scan# 427

Delta R.T. -0.03 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47



Tgt Ion	Ion	Ratio	Resp	Lower	Upper
82	100		1490528		
128	40.1	30.7	46.1		
54	53.2	42.5	63.7		




Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF0





#31

Naphthalene

Concen: 9.38 ng

RT: 7.66 min Scan# 523

Delta R.T. -0.05 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F


ClientSampleId :

S34-15.5




Tot Ion:	128	Resp:	363766
Ion Ratio	Lower	Upper	
128	100		
129	11.3	8.6	13.0
127	13.2	10.6	16.0

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E



#37

2-Methylnaphthalene

Concen: 4.53 ng


RT: 8.77 min Scan# 620

Delta R.T. -0.03 min



Lab File: BF083537.D


Acq: 7 Dec 2015 1:47

Tgt Ion:	142	Resp:	108776
Ion Ratio	Lower	Upper	
142	100		
141	90.2	70.3	105.5
115	35.8	29.2	43.8



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 10.42 min Scan# 764

Delta R.T. -0.05 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F





ClientSampleId :

S34-15.5

Tot Ion:	164	Resp:	346712
Ion Ratio	Lower	Upper	
164	100		
162	101.0	78.8	118.2
160	43.2	33.4	50.2

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



#41

2,4,6-Tribromophenol

Concen: 129.07 ng




RT: 11.72 min Scan# 878


Delta R.T. -0.03 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tgt Ion:	330	Resp:	412814
Ion Ratio	Lower	Upper	
330	100		
332	95.1	0.0	0.0#
141	38.7	0.0	0.0#









#44
2-Fluorobiphenyl
Concen: 78.53 ng
RT: 9.38 min Scan# 673
Delta R.T. -0.05 min
Lab File: BF083537.D
Acq: 7 Dec 2015 1:47

Instrument :
BNA_F
ClientSampleId :
S34-15.5

Tot Ion	Ratio	Lower	Upper
172	100		
171	36.6	28.6	42.8
170	24.6	19.4	29.0




Manual Integrations
APPROVED


Mmdadoda
12/7/2015 6:33:23 PM



#49
Dimethylphthalate
Concen: 18.78 ng
RT: 10.06 min Scan# 733
Delta R.T. -0.05 min
Lab File: BF083537.D
Acq: 7 Dec 2015 1:47

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.2	2.8	4.2
164	10.1	8.6	13.0





#51

Acenaphthene

Concen: 16.61 ng

RT: 10.48 min Scan# 769

Delta R.T. -0.05 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F


ClientSampleId :

S34-15.5




Tot Ion:	154	Resp:	363344
Ion Ratio	Lower	Upper	
154	100		
153	114.1	91.4	137.0
152	53.2	43.4	65.2

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#54

Dibenzofuran

Concen: 10.18 ng


RT: 10.76 min Scan# 794

Delta R.T. -0.05 min



Lab File: BF083537.D


Acq: 7 Dec 2015 1:47

Tgt Ion:	168	Resp:	308635
Ion Ratio	Lower	Upper	
168	100		
139	37.7	28.0	42.0
169	13.3	10.8	16.2



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E





#57

Fluorene

Concen: 15.80 ng

RT: 11.31 min Scan# 842

Delta R.T. -0.05 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F





ClientSampleId :

S34-15.5

Tot Ion:	166	Resp:	416562
Ion Ratio	Lower	Upper	
166	100		
165	96.3	80.0	120.0
167	13.4	10.6	16.0

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



#63

Phenanthrene-d10

Concen: 20.00 ng




RT: 12.82 min Scan# 974


Delta R.T. -0.03 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tgt Ion:	188	Resp:	581830
Ion Ratio	Lower	Upper	
188	100		
94	11.6	9.4	14.2
80	13.7	10.6	16.0









#70
Phenanthrene
Concen: 105.17 ng
RT: 12.88 min Scan# 979
Delta R.T. -0.02 min
Lab File: BF083537.D
Acq: 7 Dec 2015 1:47

Instrument :
BNA_F
ClientSampleId :
S34-15.5

Tot Ion:178 Resp: 3561632
Ion Ratio Lower Upper
178 100
176 21.4 15.7 23.5
179 16.7 12.2 18.4




Manual Integrations
APPROVED


Mmdadoda
12/7/2015 6:33:23 PM



#71
Anthracene
Concen: 28.50 ng
RT: 12.95 min Scan# 985
Delta R.T. -0.03 min
Lab File: BF083537.D
Acq: 7 Dec 2015 1:47

Tgt Ion:178 Resp: 984227
Ion Ratio Lower Upper
178 100
176 19.1 15.2 22.8
179 17.2 12.6 18.8









#72
 Carbazole
 Concen: 12.70 ng
 RT: 13.24 min Scan# 1011
 Delta R.T. -0.03 min
 Lab File: BF083537.D
 Acq: 7 Dec 2015 1:47

Instrument :
 BNA_F
 ClientSampleId :
 S34-15.5

Tot Ion	Ratio	Resp	Lower	Upper
167	100			
166	22.2	17.2	25.8	
139	12.4	9.3	13.9	




Manual Integrations
 APPROVED


Mmdadoda
 12/7/2015 6:33:23 PM



#74
 Fluoranthene
 Concen: 106.39 ng
 RT: 14.83 min Scan# 1150
 Delta R.T. 0.01 min
 Lab File: BF083537.D
 Acq: 7 Dec 2015 1:47

Tgt Ion	Ratio	Resp	Lower	Upper
202	100			
101	14.6	0.0	31.5	
203	18.6	0.0	37.6	





#75

Chrysene-d12

Concen: 20.00 ng

RT: 17.03 min Scan# 1342

Delta R.T. -0.02 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F





ClientSampleId :

S34-15.5

Tot Ion:	240	Resp:	372052
Ion Ratio	Lower	Upper	
240	100		
120	14.2	8.1	12.1#
236	26.3	20.8	31.2

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



#77

Pyrene

Concen: 110.82 ng




RT: 15.17 min Scan# 1180


Delta R.T. 0.00 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tgt Ion:	202	Resp:	3084397
Ion Ratio	Lower	Upper	
202	100		
200	22.6	16.7	25.1
203	26.3	14.3	21.5#





#78

Terphenyl-d14

Concen: 81.24 ng

RT: 15.47 min Scan# 1206

Delta R.T. -0.01 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F

ClientSampled :

S34-15.5

Tot Ion:244 Resp: 1358305

Ion Ratio Lower Upper


244 100

212 7.2 6.1 9.1

122 9.6 11.0 16.6#

Manual Integrations
APPROVED




Mmdadoda
12/7/2015 6:33:23 PM



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 65.49 ng

RT: 17.01 min Scan# 1341

Delta R.T. -0.02 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47


Tgt Ion:228 Resp: 1425552

Ion Ratio Lower Upper

228 100

226 28.4 22.3 33.5



229 23.7 15.7 23.5#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#82

Chrysene

Concen: 61.69 ng

RT: 17.07 min Scan# 1346

Delta R.T. -0.01 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F





ClientSampleId :

S34-15.5

Tot Ion:	228	Resp:	1337701
Ion	Ratio	Lower	Upper
228	100		
226	31.2	24.7	37.1
229	24.3	16.3	24.5

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



#85

Indeno(1,2,3-cd)pyrene

Concen: 34.29 ng




RT: 19.89 min Scan# 1593


Delta R.T. -0.02 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tgt Ion:	276	Resp:	637388
Ion	Ratio	Lower	Upper
276	100		
138	22.0	0.0	0.0#
277	27.0	0.0	0.0#





#86

Pervlene-d12

Concen: 20.00 ng

RT: 18.66 min Scan# 1485

Delta R.T. -0.01 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F


ClientSampleId :

S34-15.5




Tot Ion:	264	Resp:	380648
Ion Ratio		Lower	Upper
264	100		
260	24.5	19.4	29.0
265	22.3	17.8	26.6

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E



#87

Benzo(b)fluoranthene

Concen: 60.28 ng m

RT: 18.28 min Scan# 1452




Delta R.T. -0.01 min


Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tgt Ion:	252	Resp:	1288641
Ion Ratio		Lower	Upper
252	100		
253	22.7	18.2	27.4
125	12.9	8.2	12.4#

Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#88

Benzo(k)fluoranthene

Concen: 24.92 ng m

RT: 18.29 min Scan# 1453

Delta R.T. -0.02 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F





ClientSampleId :

S34-15.5

Tot Ion:	252	Resp:	545730
Ion Ratio	Lower	Upper	
252	100		
253	23.9	18.6	27.8
125	11.5	11.5	17.3#

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



#89

Benzo(a)pyrene

Concen: 48.83 ng




RT: 18.60 min Scan# 1480


Delta R.T. -0.01 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tgt Ion:	252	Resp:	1015287
Ion Ratio	Lower	Upper	
252	100		
253	23.2	17.5	26.3
125	13.8	11.0	16.6





#90

Dibenzo(a,h)anthracene

Concen: 8.77 ng

RT: 19.91 min Scan# 1594

Delta R.T. -0.02 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Instrument :

BNA_F


ClientSampleId :

S34-15.5




Tot Ion:	278	Resp:	175394
Ion Ratio	Lower	Upper	
278	100		
139	15.7	12.8	19.2
279	24.9	18.6	28.0

Manual Integrations
APPROVED

Mmdadoda
12/7/2015 6:33:23 PM



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 30.43 ng


RT: 20.26 min Scan# 1625

Delta R.T. -0.01 min

Lab File: BF083537.D

Acq: 7 Dec 2015 1:47

Tgt Ion:	276	Resp:	613461
Ion Ratio	Lower	Upper	
276	100		
277	24.8	19.2	28.8
138	20.9	19.5	29.3



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

