

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120616\
 Data File : BF091463.D
 Acq On : 6 Dec 2016 17:21
 Operator : UM/SJ
 Sample : PB95074BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB95074BL

Quant Time: Dec 07 00:44:17 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 07 00:12:14 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	130218	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	512815	20.00	ng	0.00
38) Acenaphthene-d10	9.88	164	312197	20.00	ng	0.00
63) Phenanthrene-d10	11.35	188	484932	20.00	ng	-0.01
75) Chrysene-d12	13.99	240	339109	20.00	ng	0.00
86) Perylene-d12	15.41	264	263628	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	895817	112.38	ng	0.01
7) Phenol-d6	6.47	99	1031342	105.11	ng	0.00
23) Nitrobenzene-d5	7.40	82	654692	71.55	ng	-0.01
41) 2,4,6-Tribromophenol	10.66	330	296776	100.41	ng	0.00
44) 2-Fluorobiphenyl	9.20	172	1346283	78.08	ng	0.00
78) Terphenyl-d14	12.94	244	1189901	76.27	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120616\
Data File : BF091463.D
Acq On : 6 Dec 2016 17:21
Operator : UM/SJ
Sample : PB95074BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB95074BL

Quant Time: Dec 07 00:44:17 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 07 00:12:14 2016
Response via : Initial Calibration

