

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120622\
 Data File : BF131534.D
 Acq On : 06 Dec 2022 18:55
 Operator : CG\JU
 Sample : N5793-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-40-SB01

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 12/07/2022
 Supervised By : Jagrut Upadhyay 12/07/2022

Quant Time: Dec 07 03:59:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 06 01:30:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.922	152	94058	20.000	ng	0.00
21) Naphthalene-d8	8.210	136	346204	20.000	ng	# 0.00
39) Acenaphthene-d10	9.975	164	200776	20.000	ng	0.00
64) Phenanthrene-d10	11.469	188	391082	20.000	ng	0.00
76) Chrysene-d12	14.121	240	230229	20.000	ng	0.00
86) Perylene-d12	15.639	264	184466	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.551	112	655685	119.044	ng	0.02
7) Phenol-d6	6.545	99	706297	102.689	ng	0.00
23) Nitrobenzene-d5	7.492	82	541293	71.759	ng	0.00
42) 2,4,6-Tribromophenol	10.769	330	412487	180.432	ng	0.00
45) 2-Fluorobiphenyl	9.292	172	1335309	87.248	ng	0.00
79) Terphenyl-d14	13.069	244	1722692	112.999	ng	0.00
Target Compounds						Qvalue
32) Benzoic acid	7.934	122	41140m	14.910	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120622\
Data File : BF131534.D
Acq On : 06 Dec 2022 18:55
Operator : CG\JU
Sample : N5793-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-40-SB01

Quant Time: Dec 07 03:59:35 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 06 01:30:30 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 12/07/2022
Supervised By :Jagrut Upadhyay 12/07/2022

