

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120718\
 Data File : BF111219.D
 Acq On : 8 Dec 2018 8:17
 Operator : JU/SJ
 Sample : J6260-03DL 2X
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-01-COMPDL

Manual Integrations
 APPROVED

Sohil
 12/11/2018 8:44:58 AM

Quant Time: Dec 10 01:47:13 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 05 17:30:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	526055	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	2046746	20.00	ng	0.00
39) Acenaphthene-d10	9.84	164	960657	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	1514477	20.00	ng	0.00
76) Chrysene-d12	13.96	240	1087105	20.00	ng	0.01
87) Perylene-d12	15.40	264	1038385	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	1817396	53.91	ng	0.00
7) Phenol-d6	6.43	99	2344684	53.99	ng	0.00
23) Nitrobenzene-d5	7.36	82	1429241	41.34	ng	0.00
42) 2,4,6-Tribromophenol	10.63	330	429476	56.24	ng	0.00
45) 2-Fluorobiphenyl	9.16	172	2239990	34.65	ng	0.00
79) Terphenyl-d14	12.91	244	1657422	31.22	ng	0.00
Target Compounds						
10) Phenol	6.44	94	161168	3.404	ng	91
31) Naphthalene	8.10	128	444271	4.985	ng	# 93
37) 2-Methylnaphthalene	8.80	142	140273	2.282	ng	95
49) Acenaphthylene	9.70	152	218740	2.561	ng	96
50) Dimethylphthalate	9.55	163	159690	2.388	ng	97
52) Acenaphthene	9.87	154	347604	6.160	ng	99
55) Dibenzofuran	10.05	168	380268	4.892	ng	# 89
58) Fluorene	10.39	166	391350	7.140	ng	98
71) Phenanthrene	11.35	178	3750688	51.486	ng	99
72) Anthracene	11.40	178	1145575	15.601	ng	95
73) Carbazole	11.55	167	514934	6.739	ng	93
75) Fluoranthene	12.54	202	4181104	58.654	ng	96
78) Pylene	12.77	202	3915066	46.226	ng	99
81) Benzo(a)anthracene	13.95	228	2193665	34.441	ng	96
83) Chrysene	13.99	228	1956737	30.946	ng	97
86) Indeno(1,2,3-cd)pyrene	16.81	276	744984	14.674	ng	# 90
88) Benzo(b)fluoranthene	14.99	252	2318961m	39.412	ng	
89) Benzo(k)fluoranthene	15.01	252	858525m	15.093	ng	
90) Benzo(a)pyrene	15.34	252	1631929	30.238	ng	97
91) Dibenzo(a,h)anthracene	16.82	278	199860m	4.413	ng	
92) Benzo(g,h,i)perylene	17.24	276	645468	14.182	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120718\
Data File : BF111219.D
Acq On : 8 Dec 2018 8:17
Operator : JU/SJ
Sample : J6260-03DL 2X
Misc :
ALS Vial : 41 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Manual Integrations
APPROVED

Sohil
12/11/2018 8:44:58 AM

Quant Time: Dec 10 01:47:13 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 05 17:30:55 2018
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.80 min Scan# 802
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Lower	Upper
152	100		
150	152.2	124.6	187.0
115	61.3	52.2	78.2

Manual Integrations
APPROVED

Sohil
12/11/2018 8:44:58 AM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol
Concen: 53.910 ng
RT: 5.41 min Scan# 565
Delta R.T. -0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Lower	Upper
112	100		
64	63.7	50.2	75.4
63	32.3	26.6	39.8








Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 53.992 ng

RT: 6.43 min Scan# 738

Delta R.T. -0.01 min

Lab File: BF111219.D

Acq: 8 Dec 2018 8:17

Tot Ion: 99 Resp: 2344684

Ion Ratio Lower Upper

99 100

42 21.4 19.3 28.9

71 31.4 27.4 41.2

Instrument :

BNA_F





ClientSampleId :

SB-01-COMPDL

Manual Integrations

APPROVED

Sohil
12/11/2018 8:44:58 AM



#10

Phenol

Concen: 3.404 ng

RT: 6.44 min Scan# 740

Delta R.T. -0.01 min

Lab File: BF111219.D

Acq: 8 Dec 2018 8:17


Tgt Ion: 94 Resp: 161168


Ion Ratio Lower Upper

94 100

65 26.4 10.5 50.5

66 33.4 19.8 59.8






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.09 min Scan# 1020
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL




Tot Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.0	13.6
54	10.4	8.4	12.6
68	6.8	5.0	7.4

Manual Integrations
APPROVED

Sohil
12/11/2018 8:44:58 AM






Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1




#23
Nitrobenzene-d5
Concen: 41.342 ng
RT: 7.36 min Scan# 897
Delta R.T. -0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Lower	Upper
82	100		
128	48.4	38.1	57.1
54	55.8	50.3	75.5

Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF1





#31

Naphthalene

Concen: 4.985 ng

RT: 8.10 min Scan# 1023

Delta R.T. -0.01 min

Lab File: BF111219.D

Acq: 8 Dec 2018 8:17

Instrument :

BNA_F





ClientSampleId :

SB-01-COMPDL

Tot Ion:	128	Resp:	444271
Ion Ratio	Lower	Upper	
128	100		
129	10.7	10.5	15.7
127	12.3	12.5	18.7#

Manual Integrations
APPROVED

Sohil
12/11/2018 8:44:58 AM



#37

2-Methylnaphthalene

Concen: 2.282 ng




RT: 8.80 min Scan# 1141


Delta R.T. 0.00 min

Lab File: BF111219.D

Acq: 8 Dec 2018 8:17

Tgt Ion:	142	Resp:	140273
Ion Ratio	Lower	Upper	
142	100		
141	86.1	71.4	107.0
115	31.8	29.2	43.8






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.84 min Scan# 1318
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Resp	Lower	Upper
164	100	960657		
162	99.0		80.6	120.8
160	44.5		35.5	53.3

Manual Integrations
APPROVED

Sohil
12/11/2018 8:44:58 AM






Abundance

Ion 164.00 (163.70 to 164.70): E


Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#42
2,4,6-Tribromophenol
Concen: 56.243 ng
RT: 10.63 min Scan# 1452
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	429476		
332	98.4		77.2	115.8
141	39.3		36.1	54.1






Abundance

Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E






#45
2-Fluorobiphenyl
Concen: 34.646 ng
RT: 9.16 min Scan# 1203
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Lower	Upper
172	100		
171	37.1	33.2	49.8
170	25.0	22.9	34.3

Manual Integrations
APPROVED

Sohil
12/11/2018 8:44:58 AM






Abundance

Ion 172.00 (171.70 to 172.70): E


Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#49
Acenaphthylene
Concen: 2.561 ng
RT: 9.70 min Scan# 1294
Delta R.T. -0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.6	18.3	27.5
153	13.6	11.8	17.8






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E






#50
Dimethylphthalate
Concen: 2.388 ng
RT: 9.55 min Scan# 1269
Delta R.T. -0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Resp Lower	Upper
163	100		
194	3.3	3.0	4.4
164	10.0	8.9	13.3

Manual Integrations
APPROVED

Sohil
12/11/2018 8:44:58 AM






Abundance

Ion 163.00 (162.70 to 163.70): E


Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#52
Acenaphthene
Concen: 6.160 ng
RT: 9.87 min Scan# 1324
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Resp Lower	Upper
154	100		
153	111.8	88.7	133.1
152	53.6	43.8	65.8






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#55
Dibenzofuran
Concen: 4.892 ng
RT: 10.05 min Scan# 1353
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Lower	Upper
168	100		
139	34.8	35.1	52.7#
169	14.2	12.2	18.4




Manual Integrations
APPROVED


Sohil
12/11/2018 8:44:58 AM



#58
Fluorene
Concen: 7.140 ng
RT: 10.39 min Scan# 1411
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Lower	Upper
166	100		
165	96.4	78.4	117.6
167	13.5	11.4	17.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.32 min Scan# 1570
Delta R.T. 0.00 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Lower	Upper
188	100		
94	13.1	8.5	12.7#
80	13.9	9.0	13.6#




Manual Integrations
APPROVED


Sohil
12/11/2018 8:44:58 AM



#71
Phenanthrene
Concen: 51.486 ng
RT: 11.35 min Scan# 1575
Delta R.T. 0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Lower	Upper
178	100		
176	22.7	17.7	26.5
179	18.1	14.0	21.0









#72
 Anthracene
 Concen: 15.601 ng
 RT: 11.40 min Scan# 1583
 Delta R.T. 0.00 min
 Lab File: BF111219.D
 Acq: 8 Dec 2018 8:17

Instrument :
 BNA_F
 ClientSampleId :
 SB-01-COMPDL

Tot Ion:178 Resp: 1145575
 Ion Ratio Lower Upper
 178 100
 176 18.8 17.3 25.9
 179 15.9 13.8 20.8




Manual Integrations
 APPROVED


Sohil
 12/11/2018 8:44:58 AM



#73
 Carbazole
 Concen: 6.739 ng
 RT: 11.55 min Scan# 1608
 Delta R.T. 0.00 min
 Lab File: BF111219.D
 Acq: 8 Dec 2018 8:17

Tgt Ion:167 Resp: 514934
 Ion Ratio Lower Upper
 167 100
 166 20.8 19.4 29.0
 139 13.0 12.6 18.8





#75

Fluoranthene

Concen: 58.654 ng

RT: 12.54 min Scan# 1777

Delta R.T. 0.01 min

Lab File: BF111219.D

Acq: 8 Dec 2018 8:17

Instrument :

BNA_F

ClientSampleId :

SB-01-COMPDL

Tot Ion:202 Resp: 4181104

Ion Ratio Lower Upper

202 100

101 17.0 0.0 35.8





203 20.5 0.0 38.6

Manual Integrations

APPROVED

Sohil

12/11/2018 8:44:58 AM



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.96 min Scan# 2019

Delta R.T. 0.01 min

Lab File: BF111219.D

Acq: 8 Dec 2018 8:17




Tgt Ion:240 Resp: 1087105


Ion Ratio Lower Upper

240 100

120 15.8 12.8 19.2

236 26.1 21.1 31.7









#78
Pvrene
Concen: 46.226 ng
RT: 12.77 min Scan# 1816
Delta R.T. 0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion:202 Resp: 3915066
Ion Ratio Lower Upper
202 100
200 24.0 19.3 28.9
203 19.9 15.4 23.2




Manual Integrations
APPROVED


Sohil
12/11/2018 8:44:58 AM



#79
Terphenyl-d14
Concen: 31.224 ng
RT: 12.91 min Scan# 1840
Delta R.T. 0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion:244 Resp: 1657422
Ion Ratio Lower Upper
244 100
212 6.5 6.6 9.8#
122 13.9 14.2 21.2#









#81
Benzo(a)anthracene
Concen: 34.441 ng
RT: 13.95 min Scan# 2017
Delta R.T. 0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion:228 Resp: 2193665
Ion Ratio Lower Upper
228 100
226 31.5 22.6 34.0
229 21.5 16.7 25.1




Manual Integrations
APPROVED


Sohil
12/11/2018 8:44:58 AM



#83
Chrysene
Concen: 30.946 ng
RT: 13.99 min Scan# 2024
Delta R.T. 0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion:228 Resp: 1956737
Ion Ratio Lower Upper
228 100
226 30.8 26.2 39.2
229 22.5 16.9 25.3









#86
Indeno(1,2,3-cd)pyrene
Concen: 14.674 ng
RT: 16.81 min Scan# 2503
Delta R.T. 0.02 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Lower	Upper
276	100		
138	30.4	20.2	30.4#
277	23.9	15.8	23.6#




Manual Integrations
APPROVED


Sohil
12/11/2018 8:44:58 AM



#87
Perylene-d12
Concen: 20.000 ng
RT: 15.40 min Scan# 2264
Delta R.T. 0.02 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.0	19.4	29.0
265	23.5	17.4	26.0









#88
Benzo(b)fluoranthene
Concen: 39.412 ng m
RT: 14.99 min Scan# 2194
Delta R.T. 0.02 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.2	18.5	27.7
125	15.2	9.9	14.9




Manual Integrations
APPROVED


Sohil
12/11/2018 8:44:58 AM



#89
Benzo(k)fluoranthene
Concen: 15.093 ng m
RT: 15.01 min Scan# 2197
Delta R.T. 0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.5	18.2	27.4
125	15.4	9.9	14.9






#90
Benzo(a)pyrene
Concen: 30.238 ng
RT: 15.34 min Scan# 2253
Delta R.T. 0.02 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Instrument :
BNA_F
ClientSampleId :
SB-01-COMPDL

Tot Ion	Ratio	Lower	Upper
252	100		
253	24.1	17.5	26.3
125	15.9	12.4	18.6

Manual Integrations
APPROVED

Sohil
12/11/2018 8:44:58 AM






Abundance

Ion 252.00 (251.70 to 252.70): E


Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91
Dibenzo(a,h)anthracene
Concen: 4.413 ng m
RT: 16.82 min Scan# 2505
Delta R.T. 0.01 min
Lab File: BF111219.D
Acq: 8 Dec 2018 8:17

Tgt Ion	Ratio	Lower	Upper
278	100		
139	24.6	16.6	25.0
279	28.6	18.8	28.2#






Abundance

Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#92
 Benzo(a,h,i)perylene
 Concen: 14.182 ng
 RT: 17.24 min Scan# 2576
 Delta R.T. 0.02 min
 Lab File: BF111219.D
 Acq: 8 Dec 2018 8:17

Instrument :
 BNA_F
 ClientSampleId :
 SB-01-COMPDL

Tot Ion	Ratio	Lower	Upper
276	100		
277	23.3	19.3	28.9
138	26.7	22.3	33.5

Manual Integrations
 APPROVED

Sohil
 12/11/2018 8:44:58 AM

