

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120719\
 Data File : BF118140.D
 Acq On : 7 Dec 2019 20:21
 Operator : JU
 Sample : K6147-10
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P001-WC010

Manual Integrations
 APPROVED

mohammad
 12/10/2019 11:49:48 AM

Quant Time: Dec 09 07:34:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 06 17:34:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	137095	20.00	ng	0.00
21) Naphthalene-d8	8.15	136	536931	20.00	ng	0.00
39) Acenaphthene-d10	9.92	164	273655	20.00	ng	0.00
64) Phenanthrene-d10	11.40	188	438895	20.00	ng	0.00
76) Chrysene-d12	14.06	240	271943	20.00	ng	-0.01
87) Perylene-d12	15.56	264	268691	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	770144	98.39	ng	0.00
7) Phenol-d6	6.50	99	964282	101.85	ng	0.00
23) Nitrobenzene-d5	7.43	82	583894	61.18	ng	0.00
42) 2,4,6-Tribromophenol	10.71	330	277943	83.61	ng	0.00
45) 2-Fluorobiphenyl	9.23	172	1161054	64.56	ng	0.00
79) Terphenyl-d14	13.00	244	951403	60.16	ng	0.00
Target Compounds						
10) Phenol	6.52	94	51524	4.772	ng	98
50) Dimethylphthalate	9.63	163	151299	7.485	ng	99
71) Phenanthrene	11.43	178	58566	2.510	ng	99
75) Fluoranthene	12.62	202	49842	2.089	ng	93
78) Pyrene	12.86	202	60092	2.804	ng	99
83) Chrysene	14.08	228	49163	2.737	ng	# 86
84) Bis(2-ethylhexyl)phthalate	14.03	149	299210	23.380	ng	# 99
88) Benzo(b)fluoranthene	15.12	252	43989m	2.475	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120719\
Data File : BF118140.D
Acq On : 7 Dec 2019 20:21
Operator : JU
Sample : K6147-10
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P001-WC010

Manual Integrations
APPROVED

mohammad
12/10/2019 11:49:48 AM

Quant Time: Dec 09 07:34:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 06 17:34:43 2019
Response via : Initial Calibration

