

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120720\
 Data File : BF122576.D
 Acq On : 7 Dec 2020 17:23
 Operator : JU/CG
 Sample : L4960-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-6

Quant Time: Dec 08 00:23:23 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.840	152	166837	20.00	ng	0.00
21) Naphthalene-d8	8.122	136	657402	20.00	ng	0.00
39) Acenaphthene-d10	9.875	164	353619	20.00	ng	0.00
64) Phenanthrene-d10	11.363	188	660103	20.00	ng	0.00
76) Chrysene-d12	13.998	240	553451	20.00	ng	-0.01
86) Perylene-d12	15.463	264	435605	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	492180	46.70	ng	0.00
7) Phenol-d6	6.463	99	350834	25.10	ng	-0.02
23) Nitrobenzene-d5	7.404	82	1097904	83.67	ng	0.00
42) 2,4,6-Tribromophenol	10.663	330	400143	86.45	ng	-0.01
45) 2-Fluorobiphenyl	9.198	172	1987867	76.03	ng	0.00
79) Terphenyl-d14	12.951	244	1907546	60.33	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120720\
Data File : BF122576.D
Acq On : 7 Dec 2020 17:23
Operator : JU/CG
Sample : L4960-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-6

Quant Time: Dec 08 00:23:23 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 12:08:49 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.840 min Scan# 808
Delta R.T. -0.000 min
Lab File: BF122576.D
Acq: 7 Dec 2020 17:23




Instrument :
BNA_F
ClientSampleId :
MW-6

Tgt Ion:152 Resp: 166837
Ion Ratio Lower Upper
152 100
150 156.7 125.8 188.6
115 58.9 49.3 73.9



#5
2-Fluorophenol
Concen: 46.70 ng
RT: 5.451 min Scan# 572
Delta R.T. -0.006 min
Lab File: BF122576.D
Acq: 7 Dec 2020 17:23

Tgt Ion:112 Resp: 492180
Ion Ratio Lower Upper
112 100
64 57.2 46.9 70.3
63 29.2 24.5 36.7









#7
Phenol-d6
Concen: 25.10 ng
RT: 6.463 min Scan# 744
Delta R.T. -0.018 min
Lab File: BF122576.D
Acq: 7 Dec 2020 17:23




Instrument :
BNA_F
ClientSampleId :
MW-6


Tgt Ion: 99 Resp: 350834
Ion Ratio Lower Upper
99 100
42 15.8 13.8 20.6
71 31.9 26.9 40.3



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.122 min Scan# 1026
Delta R.T. -0.006 min
Lab File: BF122576.D
Acq: 7 Dec 2020 17:23

Tgt Ion: 136 Resp: 657402
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 7.3 5.5 8.3
68 5.7 4.6 6.8





#23

Nitrobenzene-d5

Concen: 83.67 ng

RT: 7.404 min Scan# 904





Delta R.T. -0.006 min

Lab File: BF122576.D

Acq: 7 Dec 2020 17:23

Tgt Ion: 82 Resp: 1097904

Ion	Ratio	Lower	Upper
82	100		
128	47.4	37.0	55.4
54	46.9	38.4	57.6



#39

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.875 min Scan# 1324




Delta R.T. -0.006 min

Lab File: BF122576.D

Acq: 7 Dec 2020 17:23

Tgt Ion: 164 Resp: 353619

Ion	Ratio	Lower	Upper
164	100		
162	101.7	80.8	121.2
160	44.2	35.4	53.0




Instrument :

BNA_F

ClientSampleId :





MW-6



#42
2,4,6-Tribromophenol
Concen: 86.45 ng
RT: 10.663 min Scan# 1458
Delta R.T. -0.012 min
Lab File: BF122576.D
Acq: 7 Dec 2020 17:23




Instrument :
BNA_F
ClientSampleId :
MW-6


Tgt Ion: 330 Resp: 400143
Ion Ratio Lower Upper
330 100
332 97.2 77.5 116.3
141 29.3 23.8 35.8



#45
2-Fluorobiphenyl
Concen: 76.03 ng
RT: 9.198 min Scan# 1209
Delta R.T. -0.006 min
Lab File: BF122576.D
Acq: 7 Dec 2020 17:23

Tgt Ion: 172 Resp: 1987867
Ion Ratio Lower Upper
172 100
171 36.7 29.6 44.4
170 24.3 19.5 29.3









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.363 min Scan# 1577
Delta R.T. -0.006 min
Lab File: BF122576.D
Acq: 7 Dec 2020 17:23




Instrument :
BNA_F
ClientSampleId :
MW-6


Tgt Ion:188 Resp: 660103
Ion Ratio Lower Upper
188 100
94 8.7 7.4 11.2
80 9.3 8.2 12.4



#76
Chrysene-d12
Concen: 20.00 ng
RT: 13.998 min Scan# 2025
Delta R.T. -0.012 min
Lab File: BF122576.D
Acq: 7 Dec 2020 17:23

Tgt Ion:240 Resp: 553451
Ion Ratio Lower Upper
240 100
120 11.7 8.6 12.8
236 25.7 20.3 30.5





#79

Terphenyl-d14

Concen: 60.33 ng

RT: 12.951 min Scan# 1847





Delta R.T. -0.006 min

Lab File: BF122576.D

Acq: 7 Dec 2020 17:23

Tgt Ion:244 Resp: 1907546

Ion	Ratio	Lower	Upper
244	100		
212	7.4	5.7	8.5
122	11.0	8.3	12.5



#86

Perylene-d12

Concen: 20.00 ng

RT: 15.463 min Scan# 2274

Delta R.T. -0.006 min

Lab File: BF122576.D

Acq: 7 Dec 2020 17:23

Tgt Ion:264 Resp: 435605

Ion	Ratio	Lower	Upper
264	100		
260	23.3	18.6	28.0
265	21.7	17.0	25.6

