

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120722\
 Data File : BF131548.D
 Acq On : 07 Dec 2022 21:45
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF120722

Quant Time: Dec 08 04:32:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 08 04:25:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	40.325	-0.8	101	-0.01
3	Pyridine	40.000	41.555	-3.9	101	-0.01
4	n-Nitrosodimethylamine	40.000	42.853	-7.1	104	0.00
5 S	2-Fluorophenol	80.000	81.669	-2.1	103	0.00
6	Aniline	40.000	41.272	-3.2	105	0.00
7 S	Phenol-d6	80.000	81.321	-1.7	104	0.00
8	2-Chlorophenol	40.000	40.366	-0.9	103	0.00
9	Benzaldehyde	40.000	36.889	7.8	108	0.00
10 C	Phenol	40.000	40.797	-2.0	101	0.00
11	bis(2-Chloroethyl)ether	40.000	40.503	-1.3	103	0.00
12	1,3-Dichlorobenzene	40.000	39.499	1.3	103	0.00
13 C	1,4-Dichlorobenzene	40.000	40.113	-0.3	103	0.00
14	1,2-Dichlorobenzene	40.000	41.423	-3.6	104	0.00
15	Benzyl Alcohol	40.000	42.447	-6.1	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.485	-3.7	106	0.00
17	2-Methylphenol	40.000	40.678	-1.7	103	0.00
18	Hexachloroethane	40.000	41.477	-3.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.309	-3.3	105	0.00
20	3+4-Methylphenols	40.000	41.117	-2.8	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	40.399	-1.0	105	0.00
23 S	Nitrobenzene-d5	80.000	82.823	-3.5	107	0.00
24	Nitrobenzene	40.000	41.129	-2.8	106	0.00
25	Isophorone	40.000	41.532	-3.8	108	0.00
26 C	2-Nitrophenol	40.000	43.636	-9.1	108	0.00
27	2,4-Dimethylphenol	40.000	40.598	-1.5	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.979	0.1	104	0.00
29 C	2,4-Dichlorophenol	40.000	40.362	-0.9	104	0.00
30	1,2,4-Trichlorobenzene	40.000	39.880	0.3	105	0.00
31	Naphthalene	40.000	39.938	0.2	105	0.00
32	Benzoic acid	40.000	39.632	0.9	106	0.00
33	4-Chloroaniline	40.000	41.437	-3.6	107	0.00
34 C	Hexachlorobutadiene	40.000	40.227	-0.6	103	0.00
35	Caprolactam	40.000	44.350	-10.9	117	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.232	-5.6	109	0.00
37	2-Methylnaphthalene	40.000	40.676	-1.7	108	0.00
38	1-Methylnaphthalene	40.000	40.207	-0.5	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.344	-0.9	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.620	-1.5	103	0.00
42 S	2,4,6-Tribromophenol	80.000	87.930	-9.9	114	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.124	-5.3	108	0.00
44	2,4,5-Trichlorophenol	40.000	42.205	-5.5	107	0.00
45 S	2-Fluorobiphenyl	80.000	82.938	-3.7	108	0.00
46	1,1'-Biphenyl	40.000	41.351	-3.4	108	0.00
47	2-Chloronaphthalene	40.000	40.717	-1.8	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120722\
 Data File : BF131548.D
 Acq On : 07 Dec 2022 21:45
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF120722

Quant Time: Dec 08 04:32:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 08 04:25:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.988	-10.0	109	0.00
49	Acenaphthylene	40.000	42.076	-5.2	110	0.00
50	Dimethylphthalate	40.000	43.250	-8.1	113	0.00
51	2,6-Dinitrotoluene	40.000	43.144	-7.9	111	0.00
52 C	Acenaphthene	40.000	42.835	-7.1	111	0.00
53	3-Nitroaniline	40.000	43.910	-9.8	115	0.00
54 P	2,4-Dinitrophenol	40.000	44.015	-10.0	121	0.00
55	Dibenzofuran	40.000	42.097	-5.2	111	0.00
56 P	4-Nitrophenol	40.000	45.456	-13.6	116	0.00
57	2,4-Dinitrotoluene	40.000	44.314	-10.8	114	0.00
58	Fluorene	40.000	42.547	-6.4	113	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.893	-7.2	110	0.00
60	Diethylphthalate	40.000	43.656	-9.1	115	0.00
61	4-Chlorophenyl-phenylether	40.000	41.752	-4.4	109	0.00
62	4-Nitroaniline	40.000	43.427	-8.6	113	0.00
63	Azobenzene	40.000	43.013	-7.5	112	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	115	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.591	-6.5	121	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.933	2.7	113	0.00
67	4-Bromophenyl-phenylether	40.000	39.331	1.7	112	0.00
68	Hexachlorobenzene	40.000	38.831	2.9	110	0.00
69	Atrazine	40.000	42.806	-7.0	117	0.00
70 C	Pentachlorophenol	40.000	42.080	-5.2	116	0.00
71	Phenanthrene	40.000	39.983	0.0	115	0.00
72	Anthracene	40.000	40.174	-0.4	115	0.00
73	Carbazole	40.000	40.320	-0.8	116	0.00
74	Di-n-butylphthalate	40.000	42.552	-6.4	118	0.00
75 C	Fluoranthene	40.000	41.393	-3.5	119	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	132	0.00
77	Benzydine	40.000	35.901	10.2	116	0.00
78	Pyrene	40.000	38.865	2.8	120	0.00
79 S	Terphenyl-d14	80.000	80.106	-0.1	123	0.00
80	Butylbenzylphthalate	40.000	45.667	-14.2	136	0.00
81	Benzo(a)anthracene	40.000	42.772	-6.9	135	0.00
82	3,3'-Dichlorobenzidine	40.000	41.902	-4.8	129	0.00
83	Chrysene	40.000	42.138	-5.3	131	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.778	-9.4	146	0.00
85 c	Di-n-octyl phthalate	40.000	43.063	-7.7	131	0.00
86 I	Perylene-d12	20.000	20.000	0.0	119	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.205	9.5	101	0.00
88	Benzo(b)fluoranthene	40.000	41.149	-2.9	122	0.00
89	Benzo(k)fluoranthene	40.000	42.718	-6.8	125	0.00
90 C	Benzo(a)pyrene	40.000	40.698	-1.7	119	0.00
91	Dibenzo(a,h)anthracene	40.000	36.371	9.1	102	0.00
92	Benzo(g,h,i)perylene	40.000	36.575	8.6	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120722\
Data File : BF131548.D
Acq On : 07 Dec 2022 21:45
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF120722

Quant Time: Dec 08 04:32:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 08 04:25:30 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0