

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120816\
 Data File : BF091504.D
 Acq On : 8 Dec 2016 16:02
 Operator : UM/SJ
 Sample : H5867-06
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 N002-WD-A5-03-01

Quant Time: Dec 09 00:20:09 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 09 00:14:41 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	161218	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	635505	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	274407	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	632914	20.00	ng	0.00
75) Chrysene-d12	13.97	240	403531	20.00	ng	0.00
86) Perylene-d12	15.37	264	28759	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	1040576	105.44	ng	0.02
7) Phenol-d6	6.46	99	1042165	85.79	ng	0.00
23) Nitrobenzene-d5	7.38	82	839919	74.07	ng	0.00
41) 2,4,6-Tribromophenol	10.64	330	441442	169.93	ng	0.00
44) 2-Fluorobiphenyl	9.18	172	1575316	103.95	ng	-0.01
78) Terphenyl-d14	12.93	244	1797723	96.83	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120816\
Data File : BF091504.D
Acq On : 8 Dec 2016 16:02
Operator : UM/SJ
Sample : H5867-06
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
N002-WD-A5-03-01

Quant Time: Dec 09 00:20:09 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 09 00:14:41 2016
Response via : Initial Calibration

