

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120816\
 Data File : BF091515.D
 Acq On : 8 Dec 2016 21:07
 Operator : UM/SJ
 Sample : H5886-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 URSMW-07

Quant Time: Dec 09 00:42:23 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 09 00:14:41 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	162328	20.00	ng	-0.01
21) Naphthalene-d8	8.10	136	723597	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	366593	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	665115	20.00	ng	0.00
75) Chrysene-d12	13.97	240	422770	20.00	ng	0.00
86) Perylene-d12	15.39	264	385602	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	605526	60.94	ng	0.01
7) Phenol-d6	6.45	99	472431	38.62	ng	-0.01
23) Nitrobenzene-d5	7.38	82	723171	56.01	ng	0.00
41) 2,4,6-Tribromophenol	10.64	330	352098	101.45	ng	0.00
44) 2-Fluorobiphenyl	9.18	172	1265049	62.48	ng	-0.01
78) Terphenyl-d14	12.93	244	1250473	64.29	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF120816\
Data File : BF091515.D
Acq On : 8 Dec 2016 21:07
Operator : UM/SJ
Sample : H5886-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
URSMW-07

Quant Time: Dec 09 00:42:23 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 09 00:14:41 2016
Response via : Initial Calibration

