

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120915\
 Data File : BF083681.D
 Acq On : 10 Dec 2015 12:43
 Operator : UM/IZ
 Sample : G4579-13DL2 1000X
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SHED-SOUTH-WALLDL2

Manual Integrations
 APPROVED

MMdadoda
 12/10/2015 5:50:58 PM

Quant Time: Dec 10 13:12:17 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 09 00:23:35 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.68	152	131643	20.00	ng	-0.02
21) Naphthalene-d8	7.57	136	506164	20.00	ng	-0.03
38) Acenaphthene-d10	10.36	164	221467	20.00	ng	-0.02
63) Phenanthrene-d10	12.76	188	333211	20.00	ng	-0.02
75) Chrysene-d12	16.98	240	299695	20.00	ng	-0.01
86) Perylene-d12	18.63	264	232802	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	0.00	99	0	0.00	ng	
23) Nitrobenzene-d5	0.00	82	0	0.00	ng	
41) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
44) 2-Fluorobiphenyl	0.00	172	0	0.00	ng	
78) Terphenyl-d14	15.40	244	292	0.02	ng	-0.01
Target Compounds						
51) Acenaphthene	10.42	154	47298	3.45	ng	Qvalue 93
54) Dibenzofuran	10.72	168	44565	2.34	ng	98
57) Fluorene	11.26	166	57931	3.49	ng	97
70) Phenanthrene	12.80	178	470787	24.59	ng	99
71) Anthracene	12.90	178	45650m	2.30	ng	
74) Fluoranthene	14.73	202	270524	13.81	ng	99
77) Pyrene	15.08	202	170870	7.92	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120915\
Data File : BF083681.D
Acq On : 10 Dec 2015 12:43
Operator : UM/IZ
Sample : G4579-13DL2 1000X
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SHED-SOUTH-WALLDL2

Manual Integrations
APPROVED

MMdadoda
12/10/2015 5:50:58 PM

Quant Time: Dec 10 13:12:17 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 09 00:23:35 2015
Response via : Initial Calibration

