

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120916\
 Data File : BF091525.D
 Acq On : 9 Dec 2016 13:54
 Operator : UM/SJ
 Sample : SSTDICC02.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Manual Integrations
 APPROVED

sohil
 12/12/2016 6:49:22 PM

Quant Time: Dec 09 17:45:06 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 09 15:51:40 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	226139	20.00	ng	-0.01
21) Naphthalene-d8	8.10	136	981003	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	470878	20.00	ng	-0.01
63) Phenanthrene-d10	11.33	188	817537	20.00	ng	0.00
75) Chrysene-d12	13.96	240	691019	20.00	ng	0.00
86) Perylene-d12	15.38	264	563255	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.40	112	80085	5.79	ng	-0.01
7) Phenol-d6	6.44	99	98121	5.76	ng	-0.01
23) Nitrobenzene-d5	7.37	82	88121	5.03	ng	-0.01
41) 2,4,6-Tribromophenol	10.64	330	20403	4.58	ng	0.00
44) 2-Fluorobiphenyl	0.00	172	0d	0.00	ng	
78) Terphenyl-d14	12.91	244	166518	5.24	ng	-0.01

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.40	88	20178	3.22	ng	85
3) Pyridine	3.15	79	42091	2.38	ng	89
4) n-Nitrosodimethylamine	3.05	42	20152	2.17	ng	# 79
6) Aniline	6.46	93	64139	2.83	ng	# 72
8) 2-Chlorophenol	6.59	128	44470	2.93	ng	93
9) Benzaldehyde	6.35	77	33207	3.03	ng	89
11) bis(2-Chloroethyl)ether	6.54	93	43371	3.03	ng	# 82
12) 1,3-Dichlorobenzene	6.75	146	49999	2.96	ng	98
15) Benzyl Alcohol	6.96	79	35382	2.59	ng	93
16) 2,2'-oxybis(1-Chloropropan	7.09	45	62634	2.03	ng	80
17) 2-Methylphenol	7.07	107	36852	3.07	ng	# 75
18) Hexachloroethane	7.32	117	17259	2.89	ng	# 78
19) n-Nitroso-di-n-propylamine	7.22	70	30875	2.88	ng	# 83
20) 3+4-Methylphenols	7.22	107	46087m	3.15	ng	
22) Acetophenone	7.22	105	70747	3.04	ng	# 99
24) Nitrobenzene	7.39	77	48381	2.70	ng	# 86
25) Isophorone	7.63	82	79247	2.56	ng	97
26) 2-Nitrophenol	7.71	139	17432	2.13	ng	# 86
27) 2,4-Dimethylphenol	7.76	122	35922	2.35	ng	99
28) bis(2-Chloroethoxy)methane	7.85	93	50610	2.71	ng	95
29) 2,4-Dichlorophenol	7.95	162	30868	2.30	ng	92
30) 1,2,4-Trichlorobenzene	8.04	180	40692	2.70	ng	97
31) Naphthalene	8.12	128	130878	2.81	ng	98
33) 4-Chloroaniline	8.17	127	53780	2.70	ng	99
34) Hexachlorobutadiene	8.25	225	20684	2.23	ng	98
36) 4-Chloro-3-methylphenol	8.65	107	37641	2.59	ng	98
37) 2-Methylnaphthalene	8.81	142	84422	2.67	ng	97
39) 1,2,4,5-Tetrachlorobenzene	8.98	216	35681	2.69	ng	99
42) 2,4,6-Trichlorophenol	9.09	196	19553m	2.27	ng	
43) 2,4,5-Trichlorophenol	9.13	196	22733	2.63	ng	94
46) 2-Chloronaphthalene	9.30	162	75797	3.01	ng	93
47) 2-Nitroaniline	9.39	65	20509	2.26	ng	91
49) Dimethylphthalate	9.57	163	82564	2.90	ng	100
50) 2,6-Dinitrotoluene	9.63	165	17453	2.74	ng	# 82

Data Path : Z:\HPCHEM1\BNA F\DATA\BF120916\
 Data File : BF091525.D
 Acq On : 9 Dec 2016 13:54
 Operator : UM/SJ
 Sample : SSTDICC02.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Manual Integrations
 APPROVED

sohil
 12/12/2016 6:49:22 PM

Quant Time: Dec 09 17:45:06 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 09 15:51:40 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
51) Acenaphthene	9.88	154	82853	3.09	nq	98
52) 3-Nitroaniline	9.80	138	18406	2.50	nq	# 85
55) 4-Nitrophenol	9.96	139	12148	2.28	nq	# 85
56) 2,4-Dinitrotoluene	10.03	165	19323	2.34	nq	# 71
58) 2,3,4,6-Tetrachlorophenol	10.17	232	17096	2.32	nq	92
61) 4-Nitroaniline	10.40	138	18149	2.62	nq	80
62) Azobenzene	10.54	77	89545	3.12	nq	87
65) n-Nitrosodiphenylamine	10.50	169	63228	2.55	nq	98
66) 4-Bromophenyl-phenylether	10.88	248	22813	2.60	nq	95
67) Hexachlorobenzene	10.94	284	25342	2.67	nq	97
68) Atrazine	11.04	200	20517	2.56	nq	95
71) Anthracene	11.40	178	135931	3.22	nq	96
72) Carbazole	11.56	167	109528	2.86	nq	95
73) Di-n-butylphthalate	11.90	149	111256	2.57	nq	98
74) Fluoranthene	12.53	202	122821	3.05	nq	96
76) Benzidine	12.66	184	57752	2.85	nq	95
77) Pyrene	12.76	202	134228	2.56	nq	96
79) Butylbenzylphthalate	13.39	149	40620	2.07	nq	93
80) Benzo(a)anthracene	13.95	228	103006	2.55	nq	96
81) 3,3'-Dichlorobenzidine	13.92	252	34399	2.42	nq	98
82) Chrysene	13.99	228	113038	2.83	nq	97
83) Bis(2-ethylhexyl)phthalate	13.95	149	55140	2.22	nq	99
84) Di-n-octyl phthalate	14.57	149	89623	2.38	nq	95
85) Indeno(1,2,3-cd)pyrene	16.73	276	79995	2.18	nq	98
87) Benzo(b)fluoranthene	14.97	252	88461m	2.53	nq	
88) Benzo(k)fluoranthene	14.99	252	87876m	2.98	nq	
89) Benzo(a)pyrene	15.31	252	81116	2.58	nq	98
90) Dibenzo(a,h)anthracene	16.75	278	70435	2.42	nq	97
91) Benzo(a,h,i)perylene	17.14	276	69112	2.31	nq	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA F\DATA\BF120916\
Data File : BF091525.D
Acq On : 9 Dec 2016 13:54
Operator : UM/SJ
Sample : SSTDICC02.5
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM

Quant Time: Dec 09 17:45:06 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 09 15:51:40 2016
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 6.81 min Scan# 413
 Delta R.T. -0.01 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
152	100		
150	158.4	122.5	183.7
115	70.8	51.8	77.8




Manual Integrations
 APPROVED


sohil
 12/12/2016 6:49:22 PM



#2
 1,4-Dioxane
 Concen: 3.22 ng
 RT: 2.40 min Scan# 27
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
88	100		
58	72.7	70.3	105.5
43	31.7	31.3	46.9





#3

Pvridine

Concen: 2.38 ng

RT: 3.15 min Scan# 93

Delta R.T. 0.05 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDICC02.5

Tot Ion:	79	Resp:	42091
Ion Ratio	Lower	Upper	
79	100		
52	73.9	71.0	106.4
51	44.9	37.0	55.4

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#4

n-Nitrosodimethylamine

Concen: 2.17 ng




RT: 3.05 min Scan# 84


Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	42	Resp:	20152
Ion Ratio	Lower	Upper	
42	100		
74	120.9	78.9	118.3#
44	8.5	5.8	8.6





#5

2-Fluorophenol

Concen: 5.79 ng

RT: 5.40 min Scan# 290

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDICC02.5

Tot Ion:	112	Resp:	80085
Ion Ratio	Lower	Upper	
112	100		
64	72.7	61.5	92.3
63	37.9	33.3	49.9

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#6

Aniline

Concen: 2.83 ng




RT: 6.46 min Scan# 383


Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	93	Resp:	64139
Ion Ratio	Lower	Upper	
93	100		
66	45.5	55.8	83.6#
65	22.9	29.9	44.9#





#7

Phenol-d6

Concen: 5.76 ng

RT: 6.44 min Scan# 381

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDICC02.5

Tot Ion:	99	Resp:	98121
Ion Ratio	Lower	Upper	
99	100		
42	17.0	20.6	31.0#
71	33.3	29.9	44.9

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#8

2-Chlorophenol

Concen: 2.93 ng




RT: 6.59 min Scan# 394


Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	128	Resp:	44470
Ion Ratio	Lower	Upper	
128	100		
130	33.8	13.1	53.1
64	51.0	24.1	64.1





#9

Benzaldehyde

Concen: 3.03 ng

RT: 6.35 min Scan# 373

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDICC02.5

Tot Ion:	77	Resp:	33207
Ion	Ratio	Lower	Upper
77	100		
105	95.3	66.4	106.4
106	92.8	60.9	100.9

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#11

bis(2-Chloroethyl)ether

Concen: 3.03 ng




RT: 6.54 min Scan# 390


Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	93	Resp:	43371
Ion	Ratio	Lower	Upper
93	100		
63	71.1	74.8	114.8#
95	32.0	12.4	52.4









#12
 1,3-Dichlorobenzene
 Concen: 2.96 ng
 RT: 6.75 min Scan# 408
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
146	100		
148	65.1	51.4	77.0
75	31.3	23.6	35.4




Manual Integrations
 APPROVED


sohil
 12/12/2016 6:49:22 PM



#15
 Benzyl Alcohol
 Concen: 2.59 ng
 RT: 6.96 min Scan# 426
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
79	100		
108	83.8	62.7	94.1
77	70.9	51.7	77.5









#16
2,2'-oxvbis(1-Chloropropane)
Concen: 2.03 ng
RT: 7.09 min Scan# 438
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
45	100		
77	14.2	3.8	43.8
79	10.9	0.2	40.2




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#17
2-Methylphenol
Concen: 3.07 ng
RT: 7.07 min Scan# 436
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
107	100		
108	115.2	67.7	101.5#
77	42.1	40.7	61.1
79	41.3	23.4	35.2#





#18

Hexachloroethane

Concen: 2.89 ng

RT: 7.32 min Scan# 458

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F

ClientSampleId :

SSTDIC02.5

Tot Ion:117 Resp: 17259

Ion Ratio Lower Upper

117 100

119 93.7 78.6 118.0


201 68.2 86.7 130.1#

Manual Integrations

APPROVED

sohil




12/12/2016 6:49:22 PM



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E



#19

n-Nitroso-di-n-propylamine

Concen: 2.88 ng

RT: 7.22 min Scan# 449

Delta R.T. -0.02 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion: 70 Resp: 30875


Ion Ratio Lower Upper

70 100

42 63.8 64.8 97.2#

101 9.0 6.2 9.4

130 18.4 11.7 17.5#






Abundance Ion 70.00 (69.70 to 70.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E





#20

3+4-Methylphenols

Concen: 3.15 ng/ml

RT: 7.22 min Scan# 449

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampled :

SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
107	100		
108	83.9	64.6	104.6
77	196.9	30.9	70.9#
79	27.4	9.3	49.3

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#21

Naphthalene-d8

Concen: 20.00 ng




RT: 8.10 min Scan# 526


Delta R.T. -0.00 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.1	13.7
54	7.2	9.1	13.7#
68	5.0	5.9	8.9#









#22
Acetophenone
Concen: 3.04 ng
RT: 7.22 min Scan# 449
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5

Tot Ion	105	Resp:	70747
Ion Ratio	Lower	Upper	
105	100		
71	7.3	7.9	11.9#
51	30.7	24.5	36.7
120	20.5	16.6	24.8




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#23
Nitrobenzene-d5
Concen: 5.03 ng
RT: 7.37 min Scan# 462
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	82	Resp:	88121
Ion Ratio	Lower	Upper	
82	100		
128	34.9	31.8	47.6
54	54.1	49.1	73.7





#24

Nitrobenzene

Concen: 2.70 ng

RT: 7.39 min Scan# 464

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDIC02.5

Tot Ion:	77	Resp:	48381
Ion Ratio	Lower	Upper	
77	100		
123	46.6	28.5	42.7#
65	14.6	12.5	18.7

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#25

Isophorone

Concen: 2.56 ng




RT: 7.63 min Scan# 485


Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	82	Resp:	79247
Ion Ratio	Lower	Upper	
82	100		
95	7.4	5.7	8.5
138	15.3	13.4	20.0









#26
2-Nitrophenol
Concen: 2.13 ng
RT: 7.71 min Scan# 492
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDIC02.5

Tot Ion	Ratio	Lower	Upper
139	100		
109	32.9	25.3	37.9
65	53.6	55.4	83.2




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#27
2,4-Dimethylphenol
Concen: 2.35 ng
RT: 7.76 min Scan# 496
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
122	100		
107	122.4	96.8	145.2
121	58.6	47.7	71.5





#28

bis(2-Chloroethoxy)methane

Concen: 2.71 ng

RT: 7.85 min Scan# 504

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDICC02.5

Tot Ion:	93	Resp:	50610
Ion Ratio	Lower	Upper	
93	100		
95	29.2	26.0	39.0
123	9.9	8.6	12.8

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#29

2,4-Dichlorophenol

Concen: 2.30 ng




RT: 7.95 min Scan# 513


Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	162	Resp:	30868
Ion Ratio	Lower	Upper	
162	100		
164	60.8	45.0	85.0
98	48.7	21.3	61.3






#30
 1,2,4-Trichlorobenzene
 Concen: 2.70 ng
 RT: 8.04 min Scan# 521
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5




Tot Ion	Ratio	Lower	Upper
180	100		
182	96.9	75.8	113.8
145	29.7	27.0	40.4

Manual Integrations
 APPROVED

sohil
 12/12/2016 6:49:22 PM




Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E






#31
 Naphthalene
 Concen: 2.81 ng
 RT: 8.12 min Scan# 528
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.1	13.7
127	12.6	11.0	16.4



Abundance Ion 128.00 (127.70 to 128.70): E
 Ion 129.00 (128.70 to 129.70): E
 Ion 127.00 (126.70 to 127.70): E









#33
4-Chloroaniline
Concen: 2.70 ng
RT: 8.17 min Scan# 532
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDIC02.5

Tot Ion	Ratio	Lower	Upper
127	100		
129	33.1	26.2	39.4
65	32.2	26.6	39.8
92	18.8	15.4	23.0




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#34
Hexachlorobutadiene
Concen: 2.23 ng
RT: 8.25 min Scan# 539
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
225	100		
223	61.1	49.8	74.6
227	68.4	52.6	79.0









#36
 4-Chloro-3-methylphenol
 Concen: 2.59 ng
 RT: 8.65 min Scan# 574
 Delta R.T. -0.01 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
107	100		
144	22.8	18.9	28.3
142	74.4	58.5	87.7




Manual Integrations
 APPROVED


sohil
 12/12/2016 6:49:22 PM



#37
 2-Methylnaphthalene
 Concen: 2.67 ng
 RT: 8.81 min Scan# 588
 Delta R.T. -0.01 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.7	71.7	107.5
115	37.0	25.0	37.6









#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.85 min Scan# 679
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
164	100		
162	103.4	82.6	124.0
160	47.4	36.7	55.1




Manual Integrations
APPROVED




sohil
12/12/2016 6:49:22 PM



#39
1,2,4,5-Tetrachlorobenzene
Concen: 2.69 ng
RT: 8.98 min Scan# 603
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
216	100		
214	79.1	63.8	95.8
179	23.1	18.2	27.2
108	20.9	17.2	25.8





#41

2,4,6-Tribromophenol

Concen: 4.58 ng

RT: 10.64 min Scan# 748

Delta R.T. -0.00 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tot Ion:	330	Resp:	20403
Ion Ratio	Lower	Upper	
330	100		
332	102.1	76.2	114.4
141	39.2	33.6	50.4

Instrument :





BNA_F

ClientSampleId :

SSTDIC02.5

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#42

2,4,6-Trichlorophenol

Concen: 2.27 ng m


RT: 9.09 min Scan# 613


Delta R.T. -0.00 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	196	Resp:	19553
Ion Ratio	Lower	Upper	
196	100		
198	105.8	77.4	116.0
200	40.3	25.2	37.8#






#43
 2,4,5-Trichlorophenol
 Concen: 2.63 ng
 RT: 9.13 min Scan# 616
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion:	196	Resp:	22733
Ion Ratio	Lower	Upper	
196	100		
198	100.6	77.0	115.6
97	46.7	33.5	50.3
132	30.3	20.5	30.7

Manual Integrations
 APPROVED

sohil
 12/12/2016 6:49:22 PM






Abundance

Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E


Ion 97.00 (96.70 to 97.70): BF0


Ion 132.00 (131.70 to 132.70): E



#44
 2-Fluorobiphenyl
 Concen: 0.00 ng
 Expected RT: 9.18 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion:	172
Sig	Exp Ratio
172	100
171	36.7
170	24.6









#46
 2-Chloronaphthalene
 Concen: 3.01 ng
 RT: 9.30 min Scan# 631
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
162	100		
127	35.7	34.3	51.5
164	32.1	26.6	40.0




Manual Integrations
 APPROVED


sohil
 12/12/2016 6:49:22 PM



#47
 2-Nitroaniline
 Concen: 2.26 ng
 RT: 9.39 min Scan# 639
 Delta R.T. -0.01 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
65	100		
92	65.7	53.6	80.4
138	99.4	91.0	136.6









#49
Dimethylphthalate
Concen: 2.90 ng
RT: 9.57 min Scan# 655
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDIC02.5

Tot Ion:	163	Resp:	82564
Ion Ratio	Lower	Upper	
163	100		
194	3.6	3.0	4.4
164	9.9	7.8	11.8




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#50
2,6-Dinitrotoluene
Concen: 2.74 ng
RT: 9.63 min Scan# 660
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion:	165	Resp:	17453
Ion Ratio	Lower	Upper	
165	100		
63	51.2	59.4	89.0#
89	58.6	50.8	76.2






#51
Acenaphthene
Concen: 3.09 ng
RT: 9.88 min Scan# 682
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5




Tot Ion	Ratio	Lower	Upper
154	100		
153	114.9	91.0	136.6
152	51.9	44.2	66.2

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM




Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E






#52
3-Nitroaniline
Concen: 2.50 ng
RT: 9.80 min Scan# 675
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
138	100		
108	13.3	8.1	12.1#
92	98.4	92.5	138.7



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF0









#55
4-Nitrophenol
Concen: 2.28 ng
RT: 9.96 min Scan# 689
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDIC02.5

Tot Ion:139 Resp: 12148
Ion Ratio Lower Upper
139 100
109 80.4 54.9 94.9
65 99.8 102.2 142.2#




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#56
2,4-Dinitrotoluene
Concen: 2.34 ng
RT: 10.03 min Scan# 695
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion:165 Resp: 19323
Ion Ratio Lower Upper
165 100
63 46.4 29.0 43.4#
89 81.5 43.1 64.7#
182 2.5 1.9 2.9









#58
 2,3,4,6-Tetrachlorophenol
 Concen: 2.32 ng
 RT: 10.17 min Scan# 707
 Delta R.T. -0.01 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion:	232	Resp:	17096
Ion Ratio	Lower	Upper	
232	100		
131	52.9	38.4	57.6
130	1.9	1.8	2.8
166	34.5	23.9	35.9




Manual Integrations
 APPROVED


sohil
 12/12/2016 6:49:22 PM



#61
 4-Nitroaniline
 Concen: 2.62 ng
 RT: 10.40 min Scan# 727
 Delta R.T. -0.02 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion:	138	Resp:	18149
Ion Ratio	Lower	Upper	
138	100		
92	62.1	26.9	66.9
108	83.6	49.1	89.1









#62
Azobenzene
Concen: 3.12 ng
RT: 10.54 min Scan# 740
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
77	100		
182	19.2	12.5	52.5
105	20.6	3.7	43.7
51	28.3	11.9	51.9




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.33 min Scan# 809
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
188	100		
94	7.0	9.2	13.8#
80	7.1	10.5	15.7#









#65
n-Nitrosodiphenylamine
Concen: 2.55 ng
RT: 10.50 min Scan# 736
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
169	100		
168	66.3	53.8	80.8
167	35.5	29.8	44.8




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#66
4-Bromophenyl-phenylether
Concen: 2.60 ng
RT: 10.88 min Scan# 769
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
248	100		
250	97.9	78.2	117.4
141	99.5	71.9	107.9





#67

Hexachlorobenzene

Concen: 2.67 ng

RT: 10.94 min Scan# 775

Delta R.T. -0.00 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F


ClientSampleId :

SSTDIC02.5




Tot Ion	Ratio	Resp	Lower	Upper
284	100	25342		
142	21.9	17.9	26.9	
249	27.4	23.6	35.4	

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 2.56 ng


RT: 11.04 min Scan# 783

Delta R.T. -0.01 min



Lab File: BF091525.D


Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Resp	Lower	Upper
200	100	20517		
173	22.1	7.8	47.8	
215	43.8	24.3	64.3	



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#71
 Anthracene
 Concen: 3.22 ng
 RT: 11.40 min Scan# 815
 Delta R.T. -0.01 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
178	100		
176	17.5	15.8	23.6
179	14.4	12.7	19.1




Manual Integrations
 APPROVED


sohil
 12/12/2016 6:49:22 PM



#72
 Carbazole
 Concen: 2.86 ng
 RT: 11.56 min Scan# 829
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
167	100		
166	20.6	18.1	27.1
139	12.3	11.5	17.3









#73
Di-n-butylphthalate
Concen: 2.57 ng
RT: 11.90 min Scan# 859
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDIC02.5

Tot Ion	Ratio	Resp	Lower	Upper
149	100			
150	8.6	7.5	11.3	
104	4.7	4.0	6.0	




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#74
Fluoranthene
Concen: 3.05 ng
RT: 12.53 min Scan# 914
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Resp	Lower	Upper
202	100			
101	12.7	0.0	29.5	
203	17.0	0.0	37.8	





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.96 min Scan# 1039

Delta R.T. -0.00 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F

ClientSampleId :

SSTDIC02.5

Tot Ion:240 Resp: 691019

Ion Ratio Lower Upper


240 100

120 8.8 12.1 18.1#

236 25.6 21.0 31.6

Manual Integrations
APPROVED




sohil
12/12/2016 6:49:22 PM



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 2.85 ng

RT: 12.66 min Scan# 925

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54


Tgt Ion:184 Resp: 57752

Ion Ratio Lower Upper

184 100

185 13.7 13.1 19.7



183 10.1 8.9 13.3




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
Pvrene
Concen: 2.56 ng
RT: 12.76 min Scan# 934
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5

Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	20.1	17.6	26.4
203	15.9	14.0	21.0




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#78
Terphenyl-d14
Concen: 5.24 ng
RT: 12.91 min Scan# 947
Delta R.T. -0.01 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	7.3	6.5	9.7
122	11.0	9.5	14.3









#79
Butylbenzylphthalate
Concen: 2.07 ng
RT: 13.39 min Scan# 989
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDIC02.5

Tot Ion	Ratio	Lower	Upper
149	100		
91	78.1	57.4	86.2
206	16.8	15.4	23.2




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#80
Benzo(a)anthracene
Concen: 2.55 ng
RT: 13.95 min Scan# 1038
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.8	22.5	33.7
229	19.2	16.4	24.6









#81
 3,3'-Dichlorobenzidine
 Concen: 2.42 ng
 RT: 13.92 min Scan# 1035
 Delta R.T. -0.00 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
252	100		
254	66.3	52.1	78.1
126	10.9	7.5	11.3




Manual Integrations
 APPROVED


sohil
 12/12/2016 6:49:22 PM



#82
 Chrysene
 Concen: 2.83 ng
 RT: 13.99 min Scan# 1041
 Delta R.T. -0.01 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.7	25.0	37.4
229	19.7	16.0	24.0





#83

Bis(2-ethylhexyl)phthalate

Concen: 2.22 ng

RT: 13.95 min Scan# 1038

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDIC02.5

Tot Ion:	149	Resp:	55140
Ion Ratio	Lower	Upper	
149	100		
167	23.9	19.6	29.4
279	2.1	1.9	2.9

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#84

Di-n-octyl phthalate

Concen: 2.38 ng




RT: 14.57 min Scan# 1092


Delta R.T. -0.00 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	149	Resp:	89623
Ion Ratio	Lower	Upper	
149	100		
167	1.3	1.3	1.9
43	12.5	11.9	17.9









#85
Indeno(1,2,3-cd)pyrene
Concen: 2.18 ng
RT: 16.73 min Scan# 1281
Delta R.T. -0.02 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Instrument :
BNA_F
ClientSampleId :
SSTDICC02.5

Tot Ion	Ratio	Lower	Upper
276	100		
138	26.7	21.0	31.4
277	24.0	20.2	30.2




Manual Integrations
APPROVED


sohil
12/12/2016 6:49:22 PM



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.38 min Scan# 1163
Delta R.T. -0.00 min
Lab File: BF091525.D
Acq: 9 Dec 2016 13:54

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.4	18.8	28.2
265	21.3	17.0	25.6





#87

Benzo(b)fluoranthene

Concen: 2.53 ng m

RT: 14.97 min Scan# 1127

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDICC02.5

Tot Ion:	252	Resp:	88461
Ion Ratio	Lower	Upper	
252	100		
253	21.6	18.2	27.2
125	10.0	8.6	13.0

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#88

Benzo(k)fluoranthene

Concen: 2.98 ng m




RT: 14.99 min Scan# 1129


Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	252	Resp:	87876
Ion Ratio	Lower	Upper	
252	100		
253	20.8	17.8	26.8
125	12.0	9.6	14.4





#89

Benzo(a)pyrene

Concen: 2.58 ng

RT: 15.31 min Scan# 1157

Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Instrument :

BNA_F





ClientSampleId :

SSTDICC02.5

Tot Ion:	252	Resp:	81116
Ion Ratio	Lower	Upper	
252	100		
253	22.1	18.2	27.4
125	11.8	10.4	15.6

Manual Integrations
APPROVED

sohil
12/12/2016 6:49:22 PM



#90

Dibenzo(a,h)anthracene

Concen: 2.42 ng




RT: 16.75 min Scan# 1283


Delta R.T. -0.01 min

Lab File: BF091525.D

Acq: 9 Dec 2016 13:54

Tgt Ion:	278	Resp:	70435
Ion Ratio	Lower	Upper	
278	100		
139	16.5	15.1	22.7
279	22.8	19.0	28.6





#91
 Benzo(a,h,i)perylene
 Concen: 2.31 ng
 RT: 17.14 min Scan# 1317
 Delta R.T. -0.02 min
 Lab File: BF091525.D
 Acq: 9 Dec 2016 13:54

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC02.5

Tot Ion:	276	Resp:	69112
Ion Ratio	Lower	Upper	
276	100		
277	23.7	19.0	28.4
138	22.6	15.4	23.0

Manual Integrations
 APPROVED

sohil
 12/12/2016 6:49:22 PM

