

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120920\
 Data File : BF122612.D
 Acq On : 9 Dec 2020 14:37
 Operator : JU/CG
 Sample : L4969-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NORTH-BRANFORD-03

Manual Integrations
 APPROVED

mohammad
 12/10/2020 2:42:50 PM

Quant Time: Dec 09 17:13:05 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	159961	20.00	ng	0.00
21) Naphthalene-d8	8.122	136	624239	20.00	ng	0.00
39) Acenaphthene-d10	9.875	164	329891	20.00	ng	0.00
64) Phenanthrene-d10	11.363	188	580268	20.00	ng	0.00
76) Chrysene-d12	14.004	240	402766	20.00	ng	0.00
86) Perylene-d12	15.468	264	411751	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.463	112	1065891	105.49	ng	0.00
7) Phenol-d6	6.469	99	1377179	102.77	ng	-0.01
23) Nitrobenzene-d5	7.404	82	865049	69.43	ng	0.00
42) 2,4,6-Tribromophenol	10.669	330	304524	70.52	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	1606032	65.85	ng	0.00
79) Terphenyl-d14	12.951	244	1546280	67.20	ng	0.00
Target Compounds						
						Qvalue
10) Phenol	6.481	94	38297	2.76	ng	94
50) Dimethylphthalate	9.592	163	101953	3.87	ng	100
75) Fluoranthene	12.574	202	101572	2.81	ng	99
78) Pyrene	12.804	202	117325	3.77	ng	97
88) Benzo(b)fluoranthene	15.039	252	64066m	2.22	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120920\
Data File : BF122612.D
Acq On : 9 Dec 2020 14:37
Operator : JU/CG
Sample : L4969-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NORTH-BRANFORD-03

Manual Integrations
APPROVED

mohammad
12/10/2020 2:42:50 PM

Quant Time: Dec 09 17:13:05 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 12:08:49 2020
Response via : Initial Calibration

