

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120920\
 Data File : BF122621.D
 Acq On : 9 Dec 2020 20:52
 Operator : JU/CG
 Sample : L4984-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TR3-A-D-COMP

Quant Time: Dec 10 02:51:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.840	152	191226	20.00	ng	0.00
21) Naphthalene-d8	8.122	136	762078	20.00	ng	0.00
39) Acenaphthene-d10	9.875	164	410502	20.00	ng	0.00
64) Phenanthrene-d10	11.363	188	770977	20.00	ng	0.00
76) Chrysene-d12	14.004	240	712885	20.00	ng	0.00
86) Perylene-d12	15.463	264	615108	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	1251693	103.63	ng	0.02
7) Phenol-d6	6.475	99	1408646	87.93	ng	0.00
23) Nitrobenzene-d5	7.404	82	1117414	73.46	ng	0.00
42) 2,4,6-Tribromophenol	10.669	330	598152	111.32	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	2007941	66.16	ng	0.00
79) Terphenyl-d14	12.957	244	2338948	57.43	ng	0.00
Target Compounds						
32) Benzoic acid	7.863	122	81444	9.45	ng	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120920\
Data File : BF122621.D
Acq On : 9 Dec 2020 20:52
Operator : JU/CG
Sample : L4984-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TR3-A-D-COMP

Quant Time: Dec 10 02:51:15 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 12:08:49 2020
Response via : Initial Calibration

