

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120920\
 Data File : BF122629.D
 Acq On : 10 Dec 2020 1:03
 Operator : JU/CG
 Sample : L4983-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TR1-A-D-COMP

Quant Time: Dec 10 03:08:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.840	152	199789	20.00	ng	0.00
21) Naphthalene-d8	8.122	136	782420	20.00	ng	0.00
39) Acenaphthene-d10	9.875	164	424666	20.00	ng	0.00
64) Phenanthrene-d10	11.363	188	789731	20.00	ng	0.00
76) Chrysene-d12	14.004	240	689069	20.00	ng	0.00
86) Perylene-d12	15.462	264	569054	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.463	112	995630	78.90	ng	0.00
7) Phenol-d6	6.469	99	1257301	75.12	ng	-0.01
23) Nitrobenzene-d5	7.398	82	818610	52.42	ng	-0.01
42) 2,4,6-Tribromophenol	10.663	330	177784	31.98	ng	-0.01
45) 2-Fluorobiphenyl	9.198	172	1483633	47.25	ng	0.00
79) Terphenyl-d14	12.951	244	1654192	42.02	ng	0.00
Target Compounds						
10) Phenol	6.481	94	51205	2.95	ng	86
50) Dimethylphthalate	9.592	163	150123	4.43	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120920\
Data File : BF122629.D
Acq On : 10 Dec 2020 1:03
Operator : JU/CG
Sample : L4983-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TR1-A-D-COMP

Quant Time: Dec 10 03:08:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 12:08:49 2020
Response via : Initial Calibration

