



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120920\
 Data File : BF122604.D
 Acq On : 9 Dec 2020 8:40
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 07 12:08:49 2020



AutoFind: Scans 1620, 1621, 1622; Background Corrected with Scan 1615

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.0	40982	PASS
68	69	0.00	2	1.6	707	PASS
69	198	0.00	100	32.5	43033	PASS
70	69	0.00	2	0.6	240	PASS
127	198	10	80	45.0	59605	PASS
197	198	0.00	2	0.3	345	PASS
198	198	100	100	100.0	132320	PASS
199	198	5	9	6.7	8875	PASS
275	198	10	60	28.7	37915	PASS
365	198	1	100	3.4	4546	PASS
441	198	0.01	100	19.0	25113	PASS
442	442	100	100	100.0	161016	PASS
443	442	15	24	19.0	30600	PASS