




Data Path : Z:\HPCHEM1\BNA_G\Data\BG121115\
Data File : BG020070.D
Acq On : 10 Dec 2015 10:11
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 08:30:43 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG120215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 02 18:50:01 2015
Response via : Initial Calibration



TIC: BG020070.D

(69) Pentachlorophenol (C)

17.123min (-0.027) 0.00ng

response 86743

Ion	Exp%	Act%
265.70	100	100
268.00	59.50	62.69
264.00	60.10	65.40
0.00	0.00	0.00