

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121020\
 Data File : BF122637.D
 Acq On : 10 Dec 2020 8:58
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 10 11:55:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.594	0.541	8.9	89	-0.02
3	Pyridine	1.569	1.445	7.9	87	-0.02
4	n-Nitrosodimethylamine	0.629	0.564	10.3	85	-0.02
5 S	2-Fluorophenol	1.263	1.242	1.7	96	0.00
6	Aniline	1.972	1.961	0.6	95	0.00
7 S	Phenol-d6	1.675	1.551	7.4	90	0.00
8	2-Chlorophenol	1.392	1.298	6.8	90	0.00
9	Benzaldehyde	1.014	0.962	5.1	105	0.00
10 C	Phenol	1.736	1.618	6.8	92	0.00
11	bis(2-Chloroethyl)ether	1.326	1.233	7.0	90	0.00
12	1,3-Dichlorobenzene	1.648	1.482	10.1	88	0.00
13 C	1,4-Dichlorobenzene	1.661	1.497	9.9	88	0.00
14	1,2-Dichlorobenzene	1.604	1.371	14.5	87	0.00
15	Benzyl Alcohol	1.154	1.048	9.2	86	0.00
16	2,2'-oxybis(1-Chloropropane	1.891	1.580	16.4	81	0.00
17	2-Methylphenol	1.107	1.066	3.7	91	0.00
18	Hexachloroethane	0.592	0.525	11.3	86	0.00
19 P	n-Nitroso-di-n-propylamine	0.960	0.834	13.1	84	0.00
20	3+4-Methylphenols	1.398	1.217	12.9	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
22	Acetophenone	0.497	0.469	5.6	88	0.00
23 S	Nitrobenzene-d5	0.399	0.358	10.3	82	0.00
24	Nitrobenzene	0.397	0.374	5.8	87	0.00
25	Isophorone	0.688	0.652	5.2	87	0.00
26 C	2-Nitrophenol	0.194	0.194	0.0	89	0.00
27	2,4-Dimethylphenol	0.284	0.312	-9.9	100	0.00
28	bis(2-Chloroethoxy)methane	0.471	0.421	10.6	82	0.00
29 C	2,4-Dichlorophenol	0.332	0.309	6.9	85	0.00
30	1,2,4-Trichlorobenzene	0.376	0.338	10.1	84	0.00
31	Naphthalene	1.104	1.030	6.7	87	0.00
32	Benzoic acid	0.226	0.172	23.9	65	0.00
33	4-Chloroaniline	0.461	0.434	5.9	86	0.00
34 C	Hexachlorobutadiene	0.231	0.203	12.1	81	0.00
35	Caprolactam	0.100	0.088	12.0	80	0.00
36 C	4-Chloro-3-methylphenol	0.327	0.306	6.4	85	0.00
37	2-Methylnaphthalene	0.730	0.672	7.9	85	0.00
38	1-Methylnaphthalene	0.691	0.618	10.6	83	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	0.662	0.652	1.5	85	0.00
41 P	Hexachlorocyclopentadiene	0.293	0.315	-7.5	87	0.00
42 S	2,4,6-Tribromophenol	0.262	0.248	5.3	80	0.00
43 C	2,4,6-Trichlorophenol	0.458	0.417	9.0	77	0.00
44	2,4,5-Trichlorophenol	0.473	0.451	4.7	82	0.00
45 S	2-Fluorobiphenyl	1.479	1.276	13.7	80	0.00
46	1,1'-Biphenyl	1.660	1.603	3.4	84	0.00
47	2-Chloronaphthalene	1.375	1.267	7.9	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121020\
 Data File : BF122637.D
 Acq On : 10 Dec 2020 8:58
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 10 11:55:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.401	0.352	12.2	73	0.00
49	Acenaphthylene	2.031	1.922	5.4	82	0.00
50	Dimethylphthalate	1.597	1.421	11.0	77	0.00
51	2,6-Dinitrotoluene	0.337	0.323	4.2	80	0.00
52 C	Acenaphthene	1.314	1.214	7.6	79	0.00
53	3-Nitroaniline	0.390	0.341	12.6	73	0.00
54 P	2,4-Dinitrophenol	0.165	0.139	15.8	68	0.00
55	Dibenzofuran	1.849	1.709	7.6	81	0.00
56 P	4-Nitrophenol	0.278	0.235	15.5	69	0.00
57	2,4-Dinitrotoluene	0.449	0.419	6.7	78	0.00
58	Fluorene	1.470	1.300	11.6	81	0.00
59	2,3,4,6-Tetrachlorophenol	0.415	0.363	12.5	74	0.00
60	Diethylphthalate	1.554	1.336	14.0	74	0.00
61	4-Chlorophenyl-phenylether	0.724	0.640	11.6	79	0.00
62	4-Nitroaniline	0.396	0.329	16.9	70	0.00
63	Azobenzene	1.428	1.278	10.5	77	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
65	4,6-Dinitro-2-methylphenol	0.122	0.122	0.0	76	0.00
66 c	n-Nitrosodiphenylamine	0.707	0.691	2.3	79	0.00
67	4-Bromophenyl-phenylether	0.271	0.253	6.6	75	0.00
68	Hexachlorobenzene	0.300	0.278	7.3	75	0.00
69	Atrazine	0.229	0.203	11.4	72	0.00
70 C	Pentachlorophenol	0.159	0.148	6.9	69	0.00
71	Phenanthrene	1.189	1.110	6.6	77	0.00
72	Anthracene	1.179	1.141	3.2	79	0.00
73	Carbazole	1.069	1.001	6.4	76	0.00
74	Di-n-butylphthalate	1.343	1.171	12.8	72	0.00
75 C	Fluoranthene	1.246	1.105	11.3	72	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
77	Benzydine	0.433	0.449	-3.7	71	0.00
78	Pyrene	1.546	1.469	5.0	73	0.00
79 S	Terphenyl-d14	1.143	1.015	11.2	72	0.00
80	Butylbenzylphthalate	0.697	0.597	14.3	65	0.00
81	Benzo(a)anthracene	1.337	1.245	6.9	73	0.00
82	3,3'-Dichlorobenzidine	0.479	0.490	-2.3	78	0.00
83	Chrysene	1.330	1.235	7.1	72	0.00
84	Bis(2-ethylhexyl)phthalate	0.954	0.820	14.0	67	0.00
85 c	Di-n-octyl phthalate	1.582	1.354	14.4	67	0.00
86 I	Perylene-d12	1.000	1.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	1.313	1.403	-6.9	95	0.00
88	Benzo(b)fluoranthene	1.403	1.225	12.7	73	0.00
89	Benzo(k)fluoranthene	1.283	1.205	6.1	83	0.00
90 C	Benzo(a)pyrene	1.228	1.116	9.1	79	0.00
91	Dibenzo(a,h)anthracene	1.074	1.154	-7.4	96	0.00
92	Benzo(g,h,i)perylene	1.040	1.185	-13.9	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121020\
Data File : BF122637.D
Acq On : 10 Dec 2020 8:58
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 10 11:55:49 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 12:08:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0